

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92.

I. General Information.

Establishment:

- Address: Siemens Medical Systems, Inc.
186 Wood Avenue South
Iselin, N.J. 08830

Registration Number: 2240869

Contact Person: Ms. Kathleen Rutherford
Manager, Regulatory Submissions
(908) 321-4779
(908) 321-4841

Date of Summary Preparation: 4/18/97

Device Name:

- Trade Name: MAGNETOM Project 047
- Classification Name: Magnetic Resonance Diagnostic Device,
CFR § 892.1000
- Classification: Class II
- Performance Standards: None established under Section 514 of
the Food, Drug, and Cosmetic Act.

II. Safety and Effectiveness Information Supporting Substantial Equivalence.

Device Description:

The SIEMENS MAGNETOM Project 047 is a 1.5T whole body magnetic resonance imaging (MRI) system. The MAGNETOM Project 047 is the second member of a family of magnetic resonance imaging (MRI) systems.

SIEMENS

Intended Use:

Whole body magnetic resonance imaging.

Technological Characteristics:

The MAGNETOM Project 047 system is a high field strength system. The new MAGNETOM system has a new compact design and a redesigned patient table and rf coil system.


General Safety and Effectiveness Concerns:

The safety and performance characteristics for this system are within the scope of parameters cleared for market for the MAGNETOM Vision and MAGNETOM Impact systems. No additional safety concerns have been raised.

The device labeling for the Project 047 system will contain instructions for use, indications for use, precautions, cautions, contraindications, and warnings to the user.

Substantial Equivalence:

The MAGNETOM Project 047 system is substantially equivalent to the SIEMENS MAGNETOM Vision and MAGNETOM Impact magnetic resonance imaging systems.



Kathleen M. Rutherford
Manager, Regulatory Submissions
Imaging Systems Group
Siemens Medical Systems, Inc.

5/6/97

Date



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Ms. Kathleen Rutherford
Manager, Regulatory Submission
Siemens Medical Systems, Inc.
186 Wood Avenue South
Iselin, N.J. 08830

Re: K971684
MAGNETOM Project 047
Dated: May 6, 1997
Received: May 7, 1997
Regulatory class: II
21 CFR 892.1000/Procode: 90 LNH

AUG - 5 1997

Dear Ms. Rutherford:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4613. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsmamain.html>.

Sincerely yours,

Lillian Yin, Ph.D.
Director, Division of Reproductive,
Abdominal, Ear, Nose and Throat,
and Radiological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known) K97 1684

Device Name: MAGNETOM Project 047


Indications for Use:

The SIEMENS MAGNETOM Project 047 is a 1.5 Tesla whole body magnetic resonance imaging (MRI) system intended for general diagnostic use. This MRI system will present images which reflect the spatial distribution and/or magnetic resonance spectra which are a function of the frequency and distribution of nuclei exhibiting nuclear magnetic resonance. Other physical parameters derived from the images and/or spectra may also be produced. The device includes hydrogen-1 (proton) imaging.

(please do not write below this line- continue on another page if needed)

Concurrence of CDRH, Office of Device Evaluation

Prescription Use OR Over-The-Counter Use



(Division Sign-Off)
Division of Reproductive, Abdominal, ENT,
and Radiological Devices
510(k) Number 971684