



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

MAR - 4 1998

Mr. Francis X. Casey
Invivo Research, Inc.
12601 Research Parkway
Orlando, FL 32826

Re: K971840
Omni-Trak™ 3150 Series Patient Monitor
Regulatory Class: II (two)
Product Code: 74 MHX
Dated: February 16, 1998
Received: February 17, 1998

Dear Mr. Casey

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic (QS) inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Mr. Francis X. Casey

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4648. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,

A handwritten signature in black ink, reading "Thomas J. Callahan". The signature is fluid and cursive, with the first name "Thomas" being the most prominent part.

Thomas J. Callahan, Ph.D.
Director
Division of Cardiovascular,
Respiratory, and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

ATTACHMENT 1

INDICATIONS FOR USE

This device is intended for use as a patient vital signs monitor during magnetic resonance imaging (MRI) scans. It is intended to be used by or on the order of a physician or similarly qualified health care professional. This device is intended for use in the hospital environment (MRI, Radiology Clinic, or similar settings). This device is not intended for home use.

This device includes the capability for monitoring the following patient parameters:

- ECG
- Heart Rate
- Pulse Oximetry
- Non-Invasive Blood Pressure
- Invasive Blood Pressure
- End-Tidal Carbon Dioxide / Nitrous Oxide
- Respiration effort



(Division Sign-Off)
Division of Cardiovascular, Respiratory,
and Neurological Devices
510(k) Number K971840