



zimmer

P.O. Box 708
Warsaw, IN 46581-0708
219 267-6131

MAR - 2 1998

K973357

**Summary of Safety and Effectiveness
Coonrad/Morrey Total Elbow, New Hinge Pin**

• **Submitted by:**

Zimmer, Inc.
P.O. Box 708
Warsaw, IN 46581-0708

• **Prepared by:**

Charlene Brumbaugh
Specialist
Global Regulatory Affairs
Telephone: 219-372-4962
Telefax: 219-372-4605

• **Date:**

September 4, 1997

• **Trade Name:**

Coonrad/Morrey Total Elbow

• **Common Name:**

Elbow Prosthesis

• **Classification Name:**

Prosthesis, Elbow, constrained, Cemented

• **Predicate Devices:**

Coonrad III Total Elbow, marketed by Zimmer



**Summary of Safety and Effectiveness
Coonrad/Morrey Total Elbow, New Hinge Pin
(Continued)**

- **Device Description**

The Coonrad/Morrey Total Elbow is closely based on the Coonrad III Total Elbow (K883665) cleared by FDA on February 3, 1989, with several exceptions.

- **Intended Use**

The Coonrad/Morrey Total Elbow is indicated for:

- Post-traumatic lesions or bone loss contributing to elbow instability
- Ankylosed joints, especially in cases of bilateral ankylosis from causes other than sepsis
- Advanced rheumatoid or degenerative arthritis with incapacitating pain
- Instability or loss of motion when the degree of joint damage precludes less radical procedures

The candidate for total elbow arthroplasty should exhibit joint destruction which significantly compromises the activities of daily living. Patients with single-joint involvement (generally those with traumatic or degenerative arthritis) or significant lower extremity disability which require walking aids are less amenable to treatment than patients with advanced and predominantly upper extremity involvement. If possible, elbow replacement should be done after hip or knee surgery to avoid excessive stress to the prosthesis required by crutch walking during total hip or knee rehabilitation.

- **Performance Data**

Performance testing was conducted to determine force required to unlock the hinge pin assembly. Results indicate the product is safe and effective.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Ms. Charlene Brumbaugh
Specialist
Global Regulatory Affairs
Zimmer
P.O. Box 708
Warsaw, Indiana 46581-0708

MAR - 2 1998

Re: K973357
Trade Name: Coonrad/Morrey Total Elbow, New Hinge Pin
Regulatory Class: III
Product Code: JDC
Dated: September 4, 1997
Received: September 8, 1997

Dear Ms. Brumbaugh:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Pre-market Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic (QS) inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your pre-market notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Ms. Charlene Brumbaugh

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,


fr Celia M. Witten, Ph.D., M.D.
Director
Division of General and
Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K973357

Device Name: Coonrad/Morrey Total Elbow, New Hinge Pin


Indications For Use:

- Post-traumatic lesions or bone loss contributing to elbow instability
- Ankylosed joints, especially in cases of bilateral ankylosis from causes other than sepsis
- Advanced rheumatoid or degenerative arthritis with incapacitating pain
- Instability or loss of motion when the degree of joint damage precludes less radical procedures

The candidate for total elbow arthroplasty should exhibit joint destruction which significantly compromises the activities of daily living. Patients with single joint involvement (generally those with traumatic or degenerative arthritis) or significant lower extremity disability which require walking aids are less amenable to treatment than patients with advanced and predominantly upper extremity involvement. If possible, elbow replacement should be done after hip or knee surgery to avoid excessive stress to the prosthesis required by crutch walking during total hip or knee rehabilitation.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


 (Division Sign-Off)
 Division of General Restorative Devices
 510(k) Number K973357

Prescription Use X
 (Per 21 CFR 801.109)

OR

Over-The-Counter Use _____

(Optional Format 1-2-96)

RA08702K.510