



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUN 5 1998

Mr. Hiroji Sekiguchi
• International Division II Manager
Official Correspondent
Nakanishi, Incorporated
240 Kamihinata, Kanuma-Shi, Tochigi-Ken 322
JAPAN

Re: K980162
Trade Name: Mach-Sigma Handpiece, Ultrapush
Regulatory Class: I
Product Code: EFB
Dated: May 26, 1998
Received: May 29, 1998

Dear Mr. Sekiguchi:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531

Page 2 - Mr. Sekiguchi


through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4692. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address

"<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,



Timothy A. Ulatowski
Director
Division of Dental, Infection Control,
and General Hospital Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K980162

Device Name: Air Turbine Handpiece, MACH-Sigma (Σ) Ultra-Push

Indications For Use:

The device is intended for use in general dental applications, where high speed cutting is required; such as cutting a tooth for crown preparation, cavity preparation, removing metal crown or fillings.

This device uses a USA-patented, long proven check valve, a Non-Retraction Valve, in the water line of the handpiece to prevent the oral fluids, which could be contaminated, from being drawn into the handpiece spray port(s) of the handpiece. The contaminants could reach beyond the handpiece, but this check valve, as being installed in the handpiece body, prevents them from reaching beyond the handpiece.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Susan Pruner

(Division Sign-Off)
Division of Dental, Infection Control,
and General Hospital Devices

510(k) Number K980162

Prescription Use ☒
(Per 21 CFR 801.109)

OR

Over-The-Counter Use ☐

(Optional Format 1-2-96)