

DEC 4 1998

K9826/4

5/10

RÜSCH.
INTERNATIONAL
Group Regulatory Affairs
A Subsidiary of Teleflex Incorporated (USA)

Tall Pines Park
Jaffrey, NH 03452
(603) 532-7706
FAX (603) 532-8211 or 6108

510(k) Summary

1. Submitter Name, Address, and Date of Submission.

Miss Karenann J. Brozowski
Group Regulatory Affairs Director
Rüsch International
Tall Pines Park
Jaffrey, New Hampshire 03452

Telephone: (603) 532-7706
Facsimile: (603) 532-8211

Contact: Same as above

2. Name of the Device, Common, Proprietary (if Known), and Classification.

Classification Name: Tracheal Prosthesis

Common Name: Tracheal Bronchial Stent

Proprietary Name: Rüsch Polyflex Stent Kit

3. Identification of the legally marketed device to which the submitter claims equivalence.

The Rüsch Polyflex Stent is substantially equivalent to the Schneider Wall Stent, the Cook Gianturco-Wallace Tracheobronchial "Z" Stent, and Bryan Dumon Tracheal Bronchial Stent-, Hood Tracheal Bronchial Stent, and Hood Westaby Stent

4. Description of the Device.

The Rüsç Polyflex Stent Kit consists of a medical silicone supported by a polyester core along with the components required to insert the device into the human body. These components consist of the Introducer Sleeve, the Stent Loader, Soft Positioner, and Blue Stopper.

5. Intended Use of the Device.

The indications for use for the Polyflex Stent Kit are as follows:

Airway complications such as anastomosis and stenosis, Trachea-esophageal fistula, Stenosis of the central airways (such as the trachea and main bronchus), and Compression or strictures due to tumors (Trachea and main bronchus).

6. Summary of Technological Characteristics.

The following technological characteristics are the same as or equivalent to the predicate device, the Schneider Wallstent. The primary material of the Rüsç Polyflex Stent is medical grade silicone surrounding the braided polyester woven reinforcement. This is equivalent to the proprietary polymer surrounding the superalloy braided monofilament metal reinforcement of the Schneider Stent. Both Stents are supplied with Stent insertion accessories, which allow for placement of the Stent into the body.



NOV - 3 2006

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Boston Scientific Corporation
% Ms. Angela Byland
Manager, Regulatory Affairs
Cardiovascular
Two Scimed Place
Maple Grove, Minnesota 55311

Re: K982614
Trade/Device Name: Rusch Polyflex Stent
Regulation Number: 21 CFR 878.3720
Regulation Name: Tracheal prosthesis
Regulatory Class: II
Product Code: NYT
Dated: October 22, 1998
Received: October 26, 1998

Dear Ms. Byland:

This letter corrects our substantially equivalent letter of December 4, 1998.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other


Page 2 - Ms. Angela Byland

Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to continue marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0115. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



Mark N. Melkerson

Director

Division of General, Restorative
and Neurological Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

510(k) Number (if known): K982614

Device Name: Rusch Polyflex Stent Kit

Indications for Use:

- Airway complications such as anastomosis and stenosis
- Compression or stricture due to tumors (trachea and main bronchus)
- Tracheo-esophageal fistula
- Stenosis of the central airways (such as trachea and main bronchus)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use X
(Per 21 CFR 801.109)

OR

Over-The-Counter Use _____

(Optional Format 1-2-96)

[Signature]
(Division Sign-Off)

Division of General Restorative Devices

510(k) Number K982614