



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Mr. Donald L. Baker
Director of Customer Relations
and Technical Support
Medical Division
Radiometer America Inc.
811 Sharon Drive
Westlake, Ohio 44145-1598

JAN 6 1999

Food and Drug Administration
2098 Gaither Road
Rockville MD 20850

Re: K982928
Trade Name: NPT7 Series Analyzer with DRI-TEK Cartridge
Regulatory Class: II Product Code: CHL, JHP, CGA, CGZ, CEM,
JGS, & GKF
Regulatory Class: I Product Code: KHP
Dated: August 12, 1998
Received: August 20, 1998

Dear Mr. Baker:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4588. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597, or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,

A handwritten signature in dark ink, reading "Steven Gutman". The signature is written in a cursive, slightly slanted style.

Steven I. Gutman, M.D., M.B.A.
Director
Division of Clinical
Laboratory Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

RADIOMETER AMERICA INC.
MEDICAL DIVISION

EXHIBIT V

510(k) NUMBER (IF KNOWN): K982928

DEVICE NAME: NPT7 Series

INDICATIONS FOR USE:

NPT7 Series is a stand-alone blood gas analyzer that measures pH, pCO_2 , pO_2 , and Co-oximetry parameters on human arterial/venous and capillary whole blood. The NPT7 Near Patient Testing System in its various configurations is intended for use in laboratory and point-of-care testing of pH, blood gases and co-oximetry parameters on whole blood samples by minimally trained personnel. The device is designed for stationary use as well as use in moving vehicles and air crafts.

Glen Cooper
(Division Sign-Off)
Division of Clinical Laboratory Devices
510(k) Number K982928

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED.)

(Concurrence of CDRH, Office of Device Evaluation (ODE))

Prescription Use ✓
(Per 21 CFR 801-109)

OR

Over-The-Counter-Use _____
(Optional Format 1-2-96)