



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

OCT 15 1999

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

DAIWA PRODUCTS, INC.
c/o Mr. E.J. Smith
Smith Associates
P.O. Box 4341
Crofton, MD 21114

Re: K983616
DAIWA MODEL 180
Regulatory Class: II (TWO)
Product Code: 74 DXN
Dated: July 27, 1999
Received: July 28, 1999

Dear Mr. Smith:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Mr. E.J. Smith

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4648. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,

A handwritten signature in black ink, reading "Thomas J. Callahan". The signature is fluid and cursive, with the first letters of the first and last names being capitalized and prominent.

Thomas J. Callahan, Ph.D.

Director

Division of Cardiovascular, Respiratory
and Neurological Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure


510(k) Number (if known): K983614

Device Name: Daiwa Model 180 Automatic Inflation Wrist Digital Blood Pressure Monitors

Classification Panel:

Indications for Use:

The Daiwa Model 180 Automatic Inflation Wrist Digital Blood Pressure Monitor is designed to provide signals from which systolic and diastolic pressures can be derived through the use of the oscillometric method. The device also measures the heart rate. The device is for adult use only and measures wrist circumferences from 5 1/4" to 7 3/4". The end user should not have common arrhythmias, such as atrial or ventricular premature beats or atrial fibrillation



(Division Sign-Off)
Division of Cardiovascular, Respiratory,
and Neurological Devices

510(k) Number K983614

(PLEASE DO NOT WRITE BELOW THIS LINE- CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use _____

or

Over-the-Counter Use ✓