



FEB 11 2000

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. Donald R. Ellis
Eastman Kodak Company
901 Elmgrove Road
Rochester, NY 14653

Re: K983928
Kodak LifeView Telemonitoring System (Model 110 Patient Station
and Model 1000 Nurse Station)
Regulatory Class: II (two)
Product Code: 74 MWI, DRG, and DXN
Dated: December 22, 1999
Received: December 23, 1999

Dear Mr. Ellis:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Pre-market Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 - Mr. Donald R. Ellis

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4648. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597, or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,



Celia M. Witten, Ph.D., M.D.
Acting Director
Division of Cardiovascular,
Respiratory, and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

“INDICATIONS FOR USE”

510(K) Number: K983928

Device Name: KODAK LifeView

Indication of use: The Home Health Monitoring System is intended for use as a monitoring device, whereby a health care professional can view and communicate with the patient between offices visits to collect readings of patient blood pressure, ~~blood pressure~~, pulse, temperature, and listen to sounds emanating from the bowel, heart and lungs.

The system is contraindicated for two populations. Patient's and/or caregivers that are physically unable to use this device. Patients' and/or caregivers who's cognitive and or language skills are impaired to such an extent as to preclude communication with medical personnel or effectively operate the device

This device should not be used for continuous monitoring.

Concurrence of CDRH, Office of Device Evaluation

Prescription Use

OR

Over-The-Counter

*2/14/00
P. Casey*

Jeffrey A. Steadman

Division Sign-Off
Division of Cardiovascular, Respiratory,
and Neurological Devices

510(k) Number _____