



Xomed, Inc.
Martin Sargent
Senior Regulatory Affairs Specialist
6743 Southpoint Dr. North
Jacksonville, Florida 32216-0980

June 8, 2021

Re: K992282
Trade/Device Name: Micro-France Liposuction Cannulas
Regulation Number: 21 CFR 878.5040
Regulation Name: Suction lipoplasty system
Regulatory Class: Class II
Product Code: QPB

Dear Martin Sargent:

The Food and Drug Administration (FDA) is sending this letter to notify you of an administrative change related to your previous substantial equivalence (SE) determination letter dated September 23, 1999. Specifically, FDA is updating this SE Letter because FDA has created a new product code to better categorize your device technology.

Please note that the 510(k) submission was not re-reviewed. For questions regarding this letter please contact Cindy Chowdhury, OHT4: Office of Surgical and Infection Control Devices, 240-402-6647, Cindy.Chowdhury@fda.hhs.gov.

Sincerely,

Cindy Chowdhury -S

Cindy Chowdhury, Ph.D., M.B.A.
Assistant Director
DHT4B: Division of Infection Control
and Plastic Surgery Devices
OHT4: Office of Surgical
and Infection Control Devices
Office of Product Evaluation and Quality
Center for Devices and Radiological Health



SEP 23 1999

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. Martin D. Sargent
Senior Regulatory Affairs Specialist
Xomed, Inc.
6743 Southpoint Drive North
Jacksonville, Florida 32216

Re: K992282
Trade Name: Micro-France Liposuction Cannulas
Regulatory Class: II
Product Code: MUU
Dated: July 6, 1999
Received: July 7, 1999

Dear Mr. Sargent:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (Pre-market Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic (QS) inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

Page 2 – Mr. Martin D. Sargent

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4595. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,

A handwritten signature in black ink, appearing to read 'Celia M. Witten', is written over the typed name.


Celia M. Witten, Ph.D., M.D.
Director
Division of General and
Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K992282
Device Name: Micro-France Liposuction Cannulas
Indications for Use:

Liposuction cannulas are indicated for aesthetic body contouring and general tissue aspiration.

(Please do not write below this line - continue on another page if needed)
Concurrence of CDRH, Office of Device Evaluation (ODE)


(Division Sign-Off)
Division of General Restorative Devices K992282
510(k) Number _____

Prescription Use X
(Per 21 CFR 801.109)

Or

Over-the-Counter Use _____

(Optional Format 1-2-96)

510(k) Summary

K992282

1.0 Date Prepared

July 6, 1999

2.0 Submitter (Contact)

Martin D. Sargent
Xomed Surgical Products
Jacksonville, FL
(904) 279-7586

3.0 Device Name

Proprietary Name: Micro-France liposuction cannulas
Common Name: Liposuction cannulas
Classification Name: Suction lipoplasty accessories

5.0 Device Classification

Suction lipoplasty accessories:
ProcCode MUU Class II 21 CFR 878.5040

6.0 Device Description

Micro-France liposuction cannulas consist of a hollow stainless steel tube with one or more distal openings and a handle with a suction line connector.

7.0 Intended Use

Liposuction cannulas are indicated for aesthetic body contouring and general tissue aspiration.

8.0 Substantial Equivalence

The Micro-France liposuction cannulas are substantially equivalent to devices marketed by Byron Medical (K981172) and Wells Johnson Co. (K832520) in their intended use, materials, and overall design.