



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

OCT 13 2000

Mr. Earl Smart
HDC Corporation
2109 O'Toole Avenue
San Jose, CA 95131

Re: K994059
CLA Kit (Nerve Block Infusion Kit)
Regulatory Class: II (two)
Product Code: 73 CAZ
Dated: September 6, 2000
Received: September 7, 2000

Dear Mr. Smart:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. However, you are responsible to determine that the medical devices you use as components in the kit have either been determined as substantially equivalent under the premarket notification process (Section 510(k) of the act), or were legally on the market prior to May 28, 1976, the enactment date of the Medical Device Amendments. Please note: If you purchase your device components in bulk (i.e., unfinished) and further process (e.g., sterilize) you must submit a new 510(k) before including these components in your kit. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, and labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval) it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Current Good Manufacturing Practice requirements, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic QS


Page 2 - Mr. Earl Smart

inspections, FDA will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, the Food and Drug Administration (FDA) may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal Laws or Regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on the labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4648. Additionally, for questions on the promotion and advertising of your device, please contact the Office of ~~Compliance~~ at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll free number (800) 638-2041 or (301) 443-6597, or at its internet address "<http://www.fda.gov/cdrh/dsma/dsmamain.html>".

Sincerely yours,



James E. Dillard III
Director
Division of Cardiovascular and
Respiratory Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510(k) Number

K_994059


Device Name

CLA Kit (Nerve Block Infusion Kit)

Indications for Use

The CLA Kit (Nerve Block Infusion Kit) is intended for the injection of local anesthetic to achieve peripheral nerve conduction blocks, by use of an insulated hollow needle for electrical stimulation to locate peripheral nerves, and the injection of local anesthetic either directly through the needle, or for up to 72 hours via a catheter inserted through the needle.

Concurrence of CDRH, Office of Device Evaluation (ODE)

for 
Division of Cardiovascular & Respiratory Devices
510(k) Number K994059

Prescription Use
(Per 21 CFR 801.109)

OR

Over-The-Counter Use