



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

April 10, 2014

Barbara Chiponis
Sr. Pr. Regulatory Affairs Specialist
Medtronic, Inc.
Cardiac Rhythm Disease Management
8200 Coral Sea Street NE
Mounds View, MN 55112

Re: P010015/S205
Consulta[®] CRT-P Model C4TR01
Syncra[®] CRT-P Model C2TR01
Procure: NKE

P010031/S381
Consulta[®] CRT-D Model D224TRK
Consulta[®] CRT-D Model D204TRM
Maximo[®] II CRT-D Model D284TRK
Maximo[®] II CRT-D Model D264TRM
Concerto[®] II CRT- D Model D274TRK
Protecta[®] CRT-D Model D334TRG
Protecta[®] CRT-D Model D334TRM
Protecta[®] XT CRT-D Model D314TRG
Protecta[®] XT CRT-D Model D314TRM
Viva[™] XT CRT-D Model DTBA1D1
Viva[™] XT CRT-D Model DTBA1D4
Viva[™] S CRT-D Model DTBB1D1
Viva[™] S CRT-D Model DTBB1D4
Brava[™] CRT-D Model DTBC1D1
Brava[™] CRT-D Model DTBC1D4
Procure: NIK

Filed: May 29, 2013

Dear Ms. Chiponis:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your premarket approval application (PMA) supplement for the expansion of the indications for use of the CRT-P and CRT-D devices listed above to include NYHA Functional Class I, II, or III patients who have a left ventricular ejection fraction (LVEF) $\leq 50\%$, are on stable, optimal heart failure medical therapy if indicated, and have atrioventricular

block (AV block) that is expected to require a high percentage of ventricular pacing that cannot be managed with algorithms to minimize right ventricular pacing. These devices are indicated for the following:

The **Consulta CRT-P system** is indicated for:

- NYHA Functional Class III and IV patients who remain symptomatic despite stable, optimal heart failure medical therapy and have a LVEF $\leq 35\%$ and a prolonged QRS duration.
- NYHA Functional Class I, II, or III patients who have a LVEF $\leq 50\%$, are on stable, optimal heart failure medical therapy if indicated and have atrioventricular block (AV block) that are expected to require a high percentage of ventricular pacing that cannot be managed with algorithms to minimize right ventricular pacing. Optimization of heart failure medical therapy that is limited due to AV block or the urgent need for pacing should be done post implant.

Rate adaptive pacing is provided for those patients developing a bradycardia indication who might benefit from increased pacing rates concurrent with increases in activity.

Dual chamber and atrial tracking modes are indicated for patients who may benefit from maintenance of AV synchrony.

Antitachycardia pacing (ATP) is indicated for termination of atrial tachyarrhythmias in patients with one or more of the above pacing indications.

The **Syncra CRT-P system** is indicated for:

- NYHA Functional Class III and IV patients who remain symptomatic despite stable, optimal heart failure medical therapy and have a LVEF $\leq 35\%$ and a prolonged QRS duration.
- NYHA Functional Class I, II, or III patients who have a LVEF $\leq 50\%$, are on stable, optimal heart failure medical therapy if indicated and have atrioventricular block (AV block) that are expected to require a high percentage of ventricular pacing that cannot be managed with algorithms to minimize right ventricular pacing. Optimization of heart failure medical therapy that is limited due to AV block or the urgent need for pacing should be done post implant.

Rate adaptive pacing is provided for those patients developing a bradycardia indication who might benefit from increased pacing rates concurrent with increases in activity.

Dual chamber and atrial tracking modes are indicated for patients who may benefit from maintenance of AV synchrony.

For the Consulta CRT-D Model D224TRK, Concerto II CRT-D Model D274TRK, Consulta CRT-D Model D204TRM, Protecta XT CRT-D Model D314TRM, Protecta CRT-D Model D334TRM, Protecta XT CRT-D Model D314TRG, Protecta CRT-D Model D334TRG, Viva XT CRT-D Model DTBA1D4, Viva XT CRT-D Model DTBA1D1, Viva S CRT-D Model DTBB1D4, and Viva S CRT-D Model DTBB1D1:

The [name of device] CRT-D system is indicated for patients who require ventricular antitachycardia pacing and ventricular defibrillation for automated treatment of life threatening ventricular arrhythmias, for use in patients with atrial tachyarrhythmias, or those patients who are at significant risk for developing atrial tachyarrhythmias and for providing cardiac resynchronization therapy in heart failure patients on stable, optimal heart failure medical therapy if indicated, and meet any of the following classifications:

- New York Heart Association (NYHA) Functional Class III or IV and who have a left ventricular ejection fraction $\leq 35\%$ and a prolonged QRS duration.
- Left bundle branch block (LBBB) with a QRS duration ≥ 130 ms, left ventricular ejection fraction $\leq 30\%$, and NYHA Functional Class II.
- NYHA Functional Class I, II, or III and who have left ventricular ejection fraction $\leq 50\%$ and atrioventricular block (AV block) that are expected to require a high percentage of ventricular pacing that cannot be managed with algorithms to minimize right ventricular pacing. Optimization of heart failure medical therapy that is limited due to AV block or the urgent need for pacing should be done post implant.

For the Maximo II CRT-D Model D284TRK, Maximo II CRT-D Model D264TRM, Brava CRT-D Model DTBC1D1 and Brava CRT-D Model DTBC1D4:

The [name of device] CRT-D system is indicated for ventricular antitachycardia pacing and ventricular defibrillation for automated treatment of life threatening ventricular arrhythmias and for providing cardiac resynchronization therapy in heart failure patients on stable, optimal heart failure medical therapy if indicated, and meet any of the following classifications:

- New York Heart Association (NYHA) Functional Class III or IV and who have a left ventricular ejection fraction $\leq 35\%$ and a prolonged QRS duration.
- Left bundle branch block (LBBB) with a QRS duration ≥ 130 ms, left ventricular ejection fraction $\leq 30\%$, and NYHA Functional Class II.
- NYHA Functional Class I, II, or III and who have left ventricular ejection fraction $\leq 50\%$ and atrioventricular block (AV block) that are expected to require a high

percentage of ventricular pacing that cannot be managed with algorithms to minimize right ventricular pacing. Optimization of heart failure medical therapy that is limited due to AV block or the urgent need for pacing should be done post implant.

We are pleased to inform you that the PMA supplement is approved. You may begin commercial distribution of the device as modified in accordance with the conditions of approval described below.

The sale and distribution of this device are restricted to prescription use in accordance with 21 CFR 801.109 and under section 515(d)(1)(B)(ii) of the Federal Food, Drug, and Cosmetic Act (the act). The device is further restricted under section 515(d)(1)(B)(ii) of the act insofar as the labeling must specify the specific training or experience practitioners need in order to use the device. FDA has determined that these restrictions on sale and distribution are necessary to provide reasonable assurance of the safety and effectiveness of the device. Your device is therefore a restricted device subject to the requirements in sections 502(q) and (r) of the act, in addition to the many other FDA requirements governing the manufacture, distribution, and marketing of devices.

Continued approval of this PMA is contingent upon the submission of periodic reports, required under 21 CFR 814.84, at intervals of one year (unless otherwise specified) from the date of approval of the original PMA. Two copies of this report, identified as "Annual Report" and bearing the applicable PMA reference number, should be submitted to the address below. The Annual Report should indicate the beginning and ending date of the period covered by the report and should include the information required by 21 CFR 814.84.

In addition to the above, and in order to provide continued reasonable assurance of the safety and effectiveness of the device, the Annual Report must include, separately for each model number (if applicable), the number of devices sold and distributed during the reporting period, including those distributed to distributors. The distribution data will serve as a denominator and provide necessary context for FDA to ascertain the frequency and prevalence of adverse events, as FDA evaluates the continued safety and effectiveness of the device.

In addition, because your device is a pacemaker, implantable cardioverter-defibrillator (ICD), or system lead, FDA has determined that the following additional information is necessary to provide continued reasonable assurance of the safety and effectiveness of the device. In the Annual Report, provide the following information known by or reported to the applicant:

1. The number of pulse generators domestically implanted and the number of reported explants and deaths.
2. A breakdown of the reported deaths into pulse generators related and non-pulse generator related.

3. A breakdown of the reported explants into the number reported that were:
 - a. For pacemakers and pulse generators: at end of battery life, the number that had complications not resolvable by programming, and, as applicable, the numbers that experienced other safety and effectiveness complications as ascertained by the user, applicant, or otherwise, or
 - b. For leads: associated with mechanical failure, associated with clinical complications, and as applicable, the numbers that experienced other safety and effectiveness complications as ascertained by the user, applicant, or otherwise.
4. The number of pulse generators returned to the applicant for cause from domestic sources, with a breakdown into:
 - a. For pacemakers and pulse generators: the number currently in analysis, the number operating properly, and the number at normal battery depletion and failed (with the failure mechanisms described).
 - b. For leads: the number currently in analysis, the number operating properly, the number failed (with failure mechanisms described); broken down into groupings for full leads and partial leads.
5. A cumulative survival table for the pulse generators.

Before making any change affecting the safety or effectiveness of the device, you must submit a PMA supplement or an alternate submission (30-day notice) in accordance with 21 CFR 814.39. All PMA supplements and alternate submissions (30-day notice) must comply with the applicable requirements in 21 CFR 814.39. For more information, please refer to the FDA guidance document entitled, "Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process"

(www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm089274.htm).

You are reminded that many FDA requirements govern the manufacture, distribution, and marketing of devices. For example, in accordance with the Medical Device Reporting (MDR) regulation, 21 CFR 803.50 and 21 CFR 803.52, you are required to report adverse events for this device. Manufacturers of medical devices, including in vitro diagnostic devices, are required to report to FDA no later than 30 calendar days after the day they receive or otherwise becomes aware of information, from any source, that reasonably suggests that one of their marketed devices:

1. May have caused or contributed to a death or serious injury; or
2. Has malfunctioned and such device or similar device marketed by the manufacturer would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.

Additional information on MDR, including how, when, and where to report, is available at www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm.

In accordance with the recall requirements specified in 21 CFR 806.10, you are required to submit a written report to FDA of any correction or removal of this device initiated by you to: (1) reduce a risk to health posed by the device; or (2) remedy a violation of the act caused by the device which may present a risk to health, with certain exceptions specified in 21 CFR 806.10(a)(2). Additional information on recalls is available at www.fda.gov/Safety/Recalls/IndustryGuidance/default.htm.

CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading. CDRH will notify the public of its decision to approve your PMA by making available, among other information, a summary of the safety and effectiveness data upon which the approval is based. The information can be found on the FDA CDRH Internet HomePage located at www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/DeviceApprovalsandClearances/PMAApprovals/default.htm. Written requests for this information can also be made to the Food and Drug Administration, Dockets Management Branch, (HFA-305), 5630 Fishers Lane, Rm. 1061, Rockville, MD 20852. The written request should include the PMA number or docket number. Within 30 days from the date that this information is placed on the Internet, any interested person may seek review of this decision by submitting a petition for review under section 515(g) of the act and requesting either a hearing or review by an independent advisory committee. FDA may, for good cause, extend this 30-day filing period.

Failure to comply with any post-approval requirement constitutes a ground for withdrawal of approval of a PMA. The introduction or delivery for introduction into interstate commerce of a device that is not in compliance with its conditions of approval is a violation of law.

You are reminded that, as soon as possible and before commercial distribution of your device, you must submit an amendment to this PMA submission with copies of all approved labeling in final printed form. Final printed labeling that is identical to the labeling approved in draft form will not routinely be reviewed by FDA staff when accompanied by a cover letter stating that the final printed labeling is identical to the labeling approved in draft form. If the final printed labeling is not identical, any changes from the final draft labeling should be highlighted and explained in the amendment.

All required documents should be submitted in six copies, unless otherwise specified, to the address below and should reference the above PMA number to facilitate processing.

U.S. Food and Drug Administration
Center for Devices and Radiological Health
PMA Document Mail Center – WO66-G609
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

If you have any questions concerning this approval order, please contact Erin Cutts at 301-796-6307.

Sincerely yours,

Owen P. Faris -S

for

Bram D. Zuckerman, MD
Director
Division of Cardiovascular Devices
Office of Devices Evaluation
Center for Devices and Radiological Health