

MAR - 5 2010

1400328

Date: February 2, 2010

**510(k) Summary**

**3-1. 510(k) owner (submitter)**

- |                           |   |
|---------------------------|---|
| 1) Name                   | KURARAY MEDICAL INC.  |
| 2) Address                | 1621 Sakazu, Kurashiki, Okayama 710-0801, Japan   |
| 3) Contact person         | Michio Takigawa<br>Quality Assurance Department   |
| 4) Contact person in U.S. | Kiyoyuki Arikawa<br>KURARAY AMERICA INC.<br>600 Lexington Avenue, 26th Floor<br>New York, NY 10022<br>Tel: (212)-986-2230 (Ext. 115) or (800)-879-1676<br>Fax: (212)-867-3543 |

**3-2. Name of Device**

- |                             |  |
|-----------------------------|--|
| 1) Trade / Proprietary name | CLEARFIL MAJESTY Posterior   |
| 2) Classification name      | Tooth shade resin material<br>(21 CFR section 872.3690. Product code: EBF) |
| 3) Common name              | Restorative composite resin  |
| 4) Device Listing number    | R001413  |

**3-3. Predicate device**

- |                               |  |
|-------------------------------|--|
| 1) CLEARFIL MAJESTY Posterior | 510(k) Number: K063595                     |
|                               | Classification: Tooth shade resin material |
|                               | Product Code: EBF                          |
|                               | 21 CFR Section: 872.3690                   |
|                               | Applicant: KURARAY MEDICAL INC.            |

### **3-4. Device Description**

CLEARFIL MAJESTY Posterior is a light-cure, radiopaque restorative composite resin which provides accurate color matching, high polish ability and excellent physical properties, making it ideal for both anterior and posterior restorations. It is formulated with optimal viscosity assuring easy handling and placement.

All chemical ingredients of the subject device have been used in the predicate device. And the chemical composition of the subject device is very similar to that of the predicate device. Furthermore, physical and mechanical properties of the subject device were evaluated according to ISO 4049 in comparison with the predicate device, in which the similarity between the subject device and the predicate device was shown in terms of effectiveness/performance. Therefore, it was concluded that the subject device was closely similar to the predicate device.

### **3-5. Similarities and/or differences Discussion**

#### **1) Intended uses**

The intended uses of the subject device are the same as those of the predicate device.

- (1) Direct restorations for anterior and posterior teeth (Class I – V cavities)
- (2) Correction of tooth position and tooth shape (e.g. diastema closure, dwarfed tooth, etc.)
- (3) Intraoral repairs of fractured crowns/bridges

#### **2) Chemical ingredients / Safety**

The chemical composition of the subject device is very similar to that of the predicate device, suggesting that the safety of the subject device is closely similar to the predicate device. Furthermore, the safety of the predicate device has been shown in the original application. Therefore, we have concluded that the subject device was biologically safe.

#### **3) Effectiveness / Performance**

It has been verified that the subject device complies with the requirements of the applicable FDA recognized consensus standard, ISO 4049: 2000 "Dentistry - Polymer-based filling, restorative and luting materials". As to comparison with the predicate device, according to ISO 4049: 2000, both the subject device and the predicate device comply with ISO 4049: 2000, indicating that the subject device is as effective as and performs as well as the predicate device.

**3-6. Biocompatibility**

All chemical ingredients of the subject device are the same as those of the predicate device as described in 7-4. From this fact, it can be said that the safety of the subject device is closely similar to the predicate device. Furthermore, we have already shown the biocompatibility of the predicate device in the original application. Additionally, the predicate device has been sold for 2 years and there were no reported problems or recalls according to the post market adverse event reporting requirement. Therefore, it was concluded that the subject device was biologically safe.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room -WO66-G609  
Silver Spring, MD 20993-0002

MAR - 5 2010

Kuraray Medical, Incorporated  
C/O Mr. Kiyoyuki Arikawa  
Kuraray America, Incorporated  
600 Lexington Avenue, 26<sup>th</sup> Floor  
New York, New York 1002

Re: K100328  
Trade/Device Name: CLEARFIL MAJESTY Posterior  
Regulation Number: 21CFR 872.3690  
Regulation Name: Tooth Shade Resin Material  
Regulatory Class: II  
Product Code: EBF  
Dated: February 2, 2010  
Received: February 8, 2010

Dear Mr. Arikawa:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

### Indications for Use

510(k) Number (if known): 140.0328

Device Name: CLEARFIL MAJESTY Posterior

Indications for Use:

- 1) Direct restorations for anterior and posterior teeth (Class I – V cavities)
- 2) Correction of tooth position and tooth shape (e.g. diastema closure, dwarfed tooth, etc.)
- 3) Intraoral repairs of fractured crowns/bridges

Prescription Use   X    
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use   N/A    
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

*RS Betz DDS for Dr. K.P. Mulry*  
 (Division Sign-Off)  
 Division of Anesthesiology, General Hospital  
 Infection Control, Dental Devices

510(k) Number: K100328