

1c 100378

**510(k) Summary**

The 510(k) Summary is submitted in accordance with 21 CFR Part 807, Section 807.92

APR 28 2010

**Submitter's Name:** The Daavlin Distributing Company

**Registration Number:** 1526255

**Address:** 205 West Bement Street  
Bryan, Ohio 43506

**Telephone:** 419.636.6304

**Contact:** Michele Thiel

**Date Prepared:** February 3, 2010

**Device Trade Name:** 1 Series Phototherapy Unit

**Device Common Name:** 1 Series Partial- Body Phototherapy Unit

**Device Classification:** Class II

**Product Code:** FTC

**Regulation Number:** CFR 878.4630

**Regulation Name:** Ultraviolet lamp for dermatologic/skin disorders

K 100378

**Predicate Device:** Daavlin Distributing Company

Digital DermaPal®

K073587

Daavlin Distributing Company

Flex Controlled Phototherapy Equipment

K050695

National Biological Corporation

Panosol Tru-Blu Phototherapy System

K070934

DUSA Pharmaceuticals, Inc.

BLU-U, MODEL 4170

K031805

K100378

**Device Description:**

The 1 Series Unit is a partial-body phototherapy panel ultraviolet light source controlled by an integrated digital timer or a microprocessor controller. The operator interface consists of two main components: a LCD, and a Membrane with 4 buttons. When the operator sets the treatment time or the desired dose using the operator interface located on the control panel, the treatment begins by illuminating the lamps which emit ultraviolet light or blue light. The spectral output peaks at wavelengths of 305nm,311nm,350nm,365nm, or 420nm depending on lamp. The 1 Series Unit is a therapeutic product designed for individuals who require ultraviolet radiation for diagnosed skin disorders or blue light radiation for acne vulgaris.

**Predicate Device Comparison:**

The 1 Series Unit is constructed in the same design configuration as the predicate device, utilizing identical energy sources (UV or blue light lamps ) and materials of identical composition. The intended use, general and specific indications for use, spectral output, mode of operation, labeling, treatment area, and general operating principals of the 1 Series light emitting medical device are the same or similar to those of the predicate devices.

**Intended Use:**

The 1 Series Unit is a therapeutic product designed for individuals who require ultraviolet or blue light radiation for diagnosed skin disorders.

**Performance Data:**

The 1 Series Phototherapy Unit performance data is the same as or very similar to that of the claimed predicate devices. The UV or blue light lamps and cabinet construction used in the production of the predicate devices and the 1 Series Phototherapy Unit are similar.

**Conclusion:**

On the basis of the information provided in this Summary, the Daavlin Distributing Company believes the 1 Series Phototherapy Unit is substantially equivalent to the legally commercialized predicate devices.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room –WO66-G609  
Silver Spring, MD 20993-0002

The Daavlin Distributing Company  
% Ms. Michele Thiel  
205 West Bement Street  
Bryan, Ohio 43506

APR 28 2010

Re: K100378

Trade/Device Name: 1 Series Phototherapy Unit  
Regulation Number: 21 CFR 878.4630  
Regulation Name: Ultraviolet lamp for dermatologic disorders  
Regulatory Class: Class II  
Product Code: FTC  
Dated: April 22, 2010  
Received: April 22, 2010

Dear Ms. Thiel:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical


Page 2 - Ms. Michele Thiel

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Mark N. Melkerson  
Director  
Division of Surgical, Orthopedic  
And Restorative Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**Indication for Us**

510(k) Number K100378

Device Name 1 Series Phototherapy Unit

**Indications for Use**

The 1 Series Unit is a therapeutic product designed for individuals who require ultraviolet radiation for diagnosed skin disorders. The 1-Series, equipped with PL-L36W /03 blue lamps, is indicated for the treatment of mild to moderate acne vulgaris

Prescription Use X OR Over-the-Counter Use \_\_\_\_  
(per 21 CFR 801.109)

(Please do not write below this line – Continue on another page if needed)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Neil R. Ogden for sign  
(Division Sign-Off)  
Division of Surgical, Orthopedic,  
and Restorative Devices

510(k) Number K100378