

510(k) SUMMARY OF SAFETY AND EFFECTIVENESS

Biogel[®] OrthoPro[®] Indicator[®] Underglove

This 510(k) summary of safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR 807.92(c).

Date Prepared: April 8, 2010

Applicant: Mölnlycke Health Care US, LLC
5550 Peachtree Parkway, Suite 500
Norcross, GA 30092
Registration number: 3004763499
Owner/Operator Number: 9067000

Official Correspondent: Angela L. Bunn, RAC
Director, Regulatory Affairs of the Americas
Tel: 678-250-7930
Fax: 678-250-7981
e-mail: angela.bunn@molnlycke.com

Trade/Proprietary Name: Biogel[®] OrthoPro[®] Indicator[®] Underglove

Common Name: Surgeon's Glove

Classification Name: Surgeon's Glove

Device Class: Class I

Regulation Number: 21 CFR 878.4460

Product Code: KGO

Predicate Device Name(s): Biogel[®] Indicator[®] Underglove
(K071565 and K980942)
Biogel[®] Orthopedic Surgical Glove (K071465)

Description of Device:

The proposed device, the Biogel[®] OrthoPro[®] Indicator[®] Underglove is manufactured of natural rubber latex. The Biogel[®] OrthoPro[®] Indicator[®] Underglove is manufactured of the exact same material and coated with the Biogel[®] Coating which is used on the currently cleared device(s) that have been legally marketed by Mölnlycke Health Care for many years.

The glove former design used in the manufacture of this glove allows the Biogel[®] OrthoPro[®] Indicator[®] Underglove to provide the user with these additional features:

- Slightly Curved Former with Independent/Displaced thumb
- Improved Fit, Feel and Comfort (compared to the current Biogel[®] Orthopaedic Surgical Glove)

Intended Use/Indication for Use:

The Biogel[®] OrthoPro[®] Indicator[®] Underglove is a disposable device made of natural rubber latex that is intended to be worn on the hands, usually in a surgical setting, to provide a barrier against potentially infectious material and other contaminants.

Technological Characteristics:

The Biogel[®] OrthoPro[®] Indicator[®] Underglove is substantially equivalent to the Biogel[®] Indicator[®] Underglove (K071565 and K980942) and Biogel[®] Orthopaedic Surgical Glove (K071465).

All of the assessed devices have similar indications for use, materials, product design, labeling claims and method of operation.

The difference in the proposed device, Biogel[®] OrthoPro[®] Indicator[®] Underglove is the thickness and a slight change to the former.

The Biogel[®] OrthoPro[®] Indicator[®] Underglove characteristics are summarized below as compared to ASTM requirements.

<u>Characteristic</u>	<u>Standard</u>
Dimensions	Meets ASTM D3577
Physical Properties	Meets ASTM D3577
Freedom from Holes	Meets ASTM D3577
Biocompatibility	Meets ISO 10993-1
LAL Test Results	ASTM D7102
Protein Extractables	ASTM D5712

Performance Data:

The performance data are summarized above.



Clinical Testing:

No clinical data was required.

Conclusion:

Based on the performance testing, it can be concluded that the Biogel[®] OrthoPro[®] Indicator[®] Underglove is equivalent to the Biogel[®] Indicator[®] Underglove (K071565 and K980942) and Biogel[®] Orthopedic Surgical Glove (K071465) predicates with respect to intended use, materials, design, and technological characteristics.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

Ms. Angela L. Bunn
Director , Regulatory Affairs
Molnycke Health Care US, LLC
5550 Peachtree Parkway, Suite 500
Norcross, Georgia 30092

NOV 23 2010

Re: K101163

Trade/Device Name: Biogel[®] OrthoPro[®] Indicator[®] Underglove with ≤ 50 ug/g total water extractable protein statement
Regulation Number: 21 CFR 878.4460
Regulation Name: Surgeon's Glove
Regulatory Class: I
Product Code: KGO
Dated: November 11, 2010
Received: November 12, 2010

Dear Ms. Bunn:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to

<http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Anthony D. Watson", with a horizontal line underneath the name.

Anthony D. Watson, B.S., M.S., M.B.A.
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure



INDICATIONS FOR USE

NOV 23 2010

510(k) Number (if known): K101163

Device Name: **Biogel[®] OrthoPro[®] Indicator[®] Underglove with $\leq 50\mu\text{g/g}$ total water extractable protein statement**


Indications For Use:

The Biogel[®] OrthoPro[®] Indicator[®] Underglove is a disposable device made of natural rubber latex that is intended to be worn on the hands, usually in a surgical setting, to provide a barrier against potentially infectious material and other contaminants.

Prescription Use _____ AND/OR Over-The-Counter Use X
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


Division Sign-Off)

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

Page 1 of 1

10(k) Number: K101163