

K101207

## 510(k) Summary

### JJGC Indústria e Comércio de Materiais Dentários SA Neodent Implant System

July 15, 2010

AUG 02 2010

#### ADMINISTRATIVE INFORMATION

Manufacturer Name: JJGC Indústria e Comércio de Materiais Dentários SA  
Av. Juscelino Kubitschek de Oliveira, 3291 - CIC  
Curitiba - PR - CEP, 81270-200, Brazil  
Telephone: +5 41 2169 4071  
Fax: +5 41 2169 4046

Official Contact: Holger Heussinger

Representative/Consultant: Kevin A. Thomas, PhD  
Floyd G. Larson  
PaxMed International, LLC  
11234 El Camino Real, Suite 200  
San Diego, CA 92130 USA  
Telephone: +1 (858) 792-1235  
Fax: +1 (858) 792-1236  
email: kthomas@paxmed.com  
flarson@paxmed.com

#### DEVICE NAME AND CLASSIFICATION

Trade/Proprietary Name: Neodent Implant System  
Common Name: Endosseous dental implant  
Endosseous dental implant abutment  
Classification Regulations: 21 CFR 872.3640  
21 CFR 872.3630  
Product Code: DZE  
NHA  
Classification Panel: Dental Products Panel  
Reviewing Branch: Dental Devices Branch

#### INTENDED USE

The Neodent Implant System is intended to be surgically placed in the bone of the upper or lower jaw to provide support for prosthetic devices such as artificial teeth, to restore chewing function. It may be used with single-stage or two-stage procedures, for single or multiple unit

restorations, and may be loaded immediately when good primary stability is achieved and with appropriate occlusal loading. Multiple tooth applications may be rigidly splinted.

## DEVICE DESCRIPTION

Neodent Implant System implants included in this submission are threaded, self tapping, root form, endosseous dental implants with an external hex abutment interface. They are made of commercially pure titanium, with a surface that is grit blasted and acid etched. Implants are provided in both a straight, double thread design and a tapered, single thread design. In addition, multiple straight abutments are provided for each diameter implant for both cement retained and screw retained prostheses.

## EQUIVALENCE TO MARKETED DEVICE

JJGC Indústria e Comércio de Materiais Dentários SA demonstrated that for the purposes of FDA's regulation of medical devices, the Neodent Implant System is substantially equivalent in indications and design principles to the following predicate devices, each of which has been determined by FDA to be substantially equivalent to pre-amendment devices:

Lifecore Biomedical Dental Implant Systems cleared under K002037,  
Nobel Biocare NOBELSPEEDY™ Implants cleared under K050406,  
Straumann ITI® Dental Implant System cleared under K033922,  
Straumann synOcta® Prosthetics cleared under K041295,  
Straumann UCLA Gold Abutment cleared under K022859,  
Astra Tech Fixture MicroThread™ OsseoSpeed™ cleared under K053384, and  
Thommen Medical SPI® ART Abutment cleared under K073141.

The subject device and the predicate devices have the same intended use and have the same technological characteristics. The subject and predicate implants are all made of commercially pure titanium conforming to ASTM F67 and have similar surface treatments. The subject and predicate abutments are also made of the same materials (titanium alloy, gold alloy, zirconia). The subject and predicate devices encompass the same range of physical dimensions, including diameter and length of the implants, and diameter, height, and gingival height of the abutments. The subject and predicate devices are packaged in similar materials and sterilized using similar methods.

In summary, the Neodent Implant System has the following similarities to the predicate devices:

- has the same intended use,
- uses the same operating principle,
- incorporates the same basic design,
- incorporates the same or very similar materials, and
- has similar packaging and is sterilized using the same materials and processes.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room -WO66-G609  
Silver Spring, MD 20993-0002

JJGC Industria E Comercio De Materiais Dentarios S  
Mr. Kevin A. Thomas  
Regulatory Affairs  
PaxMed International, LLC  
11234 El Camino Real, Suite 200  
San Diego, California 92130

**AUG 02 2010**

Re: K101207  
Trade/Device Name: Neodent Implant System  
Regulation Number: 21 CFR 872.3640  
Regulation Name: Endosseous Dental Implant  
Regulatory Class: II  
Product Code: DZE  
Dated: July 16, 2010  
Received: July 19, 2010

Dear Mr. Thomas:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

Indications for Use

K101207

510(k) Number (if known): \_\_\_\_\_

Device Name: Neodent Implant System

AUG 02 2010

Indications for Use:

The Neodent Implant System is intended to be surgically placed in the bone of the upper or lower jaw to provide support for prosthetic devices such as artificial teeth, to restore chewing function. It may be used with single-stage or two-stage procedures, for single or multiple unit restorations, and may be loaded immediately when good primary stability is achieved and with appropriate occlusal loading. Multiple tooth applications may be rigidly splinted.

Prescription Use  X   
(Part 21 CFR 801 Subpart D)

AND/OR Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

*R. Betz MS for Dr. Susan Rimmer*  
(Division Sign-Off)

Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

Page 1 of  1

510(k) Number:  K10207