

K101455

510(K) SUMMARY OF SAFETY AND EFFECTIVENESS

This summary of safety and effectiveness is provided as part of this Premarket Notification in compliance with 21 CFR, Part 807, Subpart E, Section 807.92.

1. Submitter's Information: 21 CFR 807.92(a)(1)

MEDISON CO., LTD.
1003, Daechi-dong, Gangnam-gu,
Seoul 135-280, Korea

Contact Person:
Mr. Kyung-Am, Shim
Regulatory Affairs Manager

Telephone: 82.2.2194.1381
Facsimile: 82.2.2194.1399
Email: kashim@medison.com

Data Prepared: March 25, 2010

2. Name of the device:Common/Usual Name:

Diagnostic Ultrasound System and Accessories

Proprietary Name:

EKO 7 Diagnostic Ultrasound System

<u>Classification Names:</u>	<u>FR Number</u>	<u>Product Code</u>
Ultrasonic Pulsed Doppler Imaging System	892.1550	IYN
Ultrasound Pulsed Echo Imaging System	892.1560	IYO
Diagnostic Ultrasound Transducer	892.1570	ITX

3. Identification of the predicate or legally marketed device:

K092159, ACCUVIX V20 Diagnostic Ultrasound System
K093849, ACCUVIX V10 Diagnostic Ultrasound System
K020789, GE Vivid 3 Expert/Pro Diagnostic Ultrasound System

4. Device Description:

The EKO 7 is a dedicated cardiac purpose, mobile, software controlled, diagnostic ultrasound system. Its function is to acquire ultrasound data and to display the data as B mode, M mode, color Doppler imaging, power Doppler imaging, Harmonic imaging, PW spectral Doppler mode, CW spectral Doppler mode or as a combination of these modes. The EKO 7 also gives the operator the ability to measure anatomical structures and offers analysis packages that provide information that is used to make a diagnosis by competent health care professionals. The EKO 7 has real time acoustic output display with two basic indices, a mechanical index and a thermal index, which are both automatically displayed.

The EKO 7 has been designed to meet the following product safety standards:

- UL 60601-1, Safety requirements for Medical Equipment
- CSA C22.2 No. 601.1, Safety requirements for Medical Equipment
- IEC60601-2-37, Diagnostic Ultrasound Safety Standards
- EN/IEC60601-1, Safety requirements for Medical Equipment
- EN/IEC60601-1-2, EMC requirements for Medical Equipment
- NEMA UD-2, Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment
- NEMA UD-3, Standard for Real Time Display of Thermal and Mechanical Acoustic Output Indices on Diagnostic Ultrasound Equipment
- IEC 61157, Declaration of the acoustic output
- ISO10993-1, Biocompatibility

5. Intended Uses:

The EKO 7 Diagnostic Ultrasound System and transducers are intended for diagnostic ultrasound imaging and fluid analysis of the human body.

The clinical applications include: Fetal, Abdominal, Pediatric, Small Organs, Adult Cephalic, Trans-esophageal (non-Cardiac, Cardiac), Muscular-Skeletal (conventional, superficial), Cardiac Adult, Cardiac Pediatric and Peripheral-vessel.

6. Technological Characteristics:

The EKO 7 is substantially equivalent to the ACCUVIX V20 Diagnostic Ultrasound System, cleared via K092159, and the ACCUVIX V10 Diagnostic Ultrasound System, cleared via K094623. All systems transmit ultrasonic energy into patients, then perform post processing of received echoes to generate on-screen display of anatomic structures and fluid flow within the body. All system allow for specialized measurements of structures and flow, and calculations.

END of 510(K) Summary



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center - WO66-G609
Silver Spring, MD 20993-0002

JUL 19 2010

Medison Co., Ltd.
% Mr. Mark Job
Responsible Third Party Official
Regulatory Technology Services LLC
1394 25th Street NW
BUFFALO MN 55313

Re: K101455
Trade/Device Name: EKO 7 Diagnostic Ultrasound System
Regulation Number: 21 CFR 892.1550
Regulation Name: Ultrasonic pulsed doppler imaging system
Regulatory Class: II
Product Code: IYN, IYO, and ITX
Dated: June 23, 2010
Received: June 24, 2010

Dear Mr. Job:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the EKO 7 Diagnostic Ultrasound System, as described in your premarket notification:

Transducer Model Number

C1-4EC
C2-6IC
L3-8
L5-13IS
MPT3-7

P2-4BA
P3-8CA
P4-12
CW2.0
CW4.0

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

If you have any questions regarding the content of this letter, please contact Shahram Vaezy at (301) 796-6242.

Sincerely yours,



Donald St. Pierre
Acting Director
Division of Radiological Devices
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure(s)

Indications for Use

510(k) Number (if known): _____

Device Name: EKO 7 Diagnostic Ultrasound System

Indications for Use:

The EKO 7 Diagnostic Ultrasound System and transducers are intended for diagnostic ultrasound imaging and fluid analysis of the human body.

The clinical applications include: Fetal, Abdominal, Pediatric, Small Organs, Neonatal Cephalic, Adult Cephalic, Trans-esophageal (non-Cardiac, Cardiac), Muscular-Skeletal (conventional, superficial), Cardiac Adult, Cardiac Pediatric and Peripheral-vessel.


Prescription Use ✓
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)


(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

Page 1 of 12

510K K101455

DIAGNOSTIC ULTRASOUND INDICATIONS STATEMENT

510(k) No.:

Device Name: EKO 7 Diagnostic Ultrasound System

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)	N	N	N	N	N	Note 1	Notes 4, 7
	Abdominal	N	N	N	N	N	Note 1	Notes 4, 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	N	N	N	N	N	Note 1	Note 5,6,7,9
	Small Organ (See Note 5)	N	N	N		N	Note 1	Note 5,6,7,9
	Neonatal Cephalic	N	N	N	N	N	Note 1	Note 4, 7
	Adult Cephalic	N	N	N	N	N	Note 1	Note 4,7
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)	N	N	N	N	N	Note 1	Note 7
	Musculo-skel. (Convent.)	N	N	N		N	Note 1	Note 5,6,7,9
	Musculo-skel. (Superfic.)	N	N	N		N	Note 1	Note 5,6,7,9
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult	N	N	N	N	N	Note 1	Note 4,7
	Cardiac Pediatric	N	N	N	N	N	Note 1	Note 4,7
	Trans-esophageal (Cardiac)	N	N	N	N	N	Note 1	Note 7
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	N	N	N	N	N	Note 1	Note 5,6,7,9
	Other (spec.)							

N= new indication; P= previously cleared by FDA; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients


Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
 Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Radiological Devices

Office of In Vitro Diagnostic Device Evaluation and Safety

Indications For Use

510K

15101455

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: C1-4EC for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)	P	P	P		P	Note 1	Notes 7
	Abdominal	P	P	P		P	Note 1	Notes 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	P	P	P		P	Note 1	Notes 7
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared under K060087; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
Prescription Use (Per 21 CFR 801.109)


(Division Sign-Off)

Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

Indications For Use

510K

K101455

Section 1.3, Page 3

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: C2-6IC for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)	P	P	P		P	Note 1	Notes 7
	Abdominal	P	P	P		P	Note 1	Notes 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	P	P	P		P	Note 1	Notes 7
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared under K092159; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
 Prescription Use (Per 21 CFR 801.109)


 (Division Sign-Off)

Division of Radiological Devices

Office of In Vitro Diagnostic Device Evaluation and Safety

Indications For Use

Section 1.3, Page 4

510K

K101455

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: L3-8 for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)							
	Abdominal							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	N	N	N		N	Note 1	Note 5, 6, 7, 9
	Small Organ (See Note 5)	N	N	N		N	Note 1	Note 5, 6, 7, 9
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	N	N	N		N	Note 1	Note 5, 6, 7, 9
	Musculo-skel. (Superfic.)	N	N	N		N	Note 1	Note 5, 6, 7, 9
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	N	N	N		N	Note 1	Note 5, 6, 7, 9
	Other (spec.)							

N= new indication; P= previously cleared by FDA; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
 Prescription Use (Per 21 CFR 801.109)


 (Division Sign-Off)

Division of Radiological Devices

Office of In Vitro Diagnostic Device Evaluation and Safety

Indications For Use

Section 1.3, Page 5

510K 15101455

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: L5-13IS for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)							
	Abdominal							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	P	P	P		P	Note 1	Note 5, 6, 7, 9
	Small Organ (See Note 5)	P	P	P		P	Note 1	Note 5, 6, 7, 9
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)	P	P	P		P	Note 1	Note 5, 6, 7, 9
	Musculo-skel. (Superfic.)	P	P	P		P	Note 1	Note 5, 6, 7, 9
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel	P	P	P		P	Note 1	Note 5, 6, 7, 9
	Other (spec.)							

N= new indication; P= previously cleared under K092159; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

Indications For Use

510K

K101455

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: MPT3-7 for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:


Clinical Application		Mode of Operation (*includes simultaneous B-mode)							
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)	
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal (See Note 3)								
	Abdominal								
	Intra-operative (See Note 6)								
	Intra-operative (Neuro.)								
	Laparoscopic								
	Pediatric								
	Small Organ (See Note 5)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (non-Cardiac)		N	N	N		N	Note 1	Note 7
	Musculo-skel. (Convent.)								
	Musculo-skel. (Superfic.)								
Intra-luminal									
Other (spec.)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Trans-esophageal (Cardiac)		N	N	N		N	Note 1	Note 7
Other (spec.)									
Peripheral Vessel	Peripheral vessel								
	Other (spec.)								

N= new indication; P= previously cleared by FDA; E= added under Appendix E

Additional Comments:

- Color Doppler includes Power (Amplitude) Doppler
- Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD
- Note 2: Includes imaging for guidance of biopsy
- Note 3: Includes infertility monitoring of follicle development
- Note 4: Color M-mode
- Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients
- Note 6: Abdominal organs and peripheral vessel
- Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging
- Note 8: 3D imaging
- Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
 Prescription Use (Per 21 CFR 801.109)


 (Division Sign-Off)
 Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: P2-4BA for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)							
	Abdominal	P	P	P	P	P	Note 1	Note 4, 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic	P	P	P	P	P	Note 1	Note 4, 7
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult	P	P	P	P	P	Note 1	Note 4, 7
	Cardiac Pediatric	P	P	P	P	P	Note 1	Note 4, 7
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared under K092159; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
 Prescription Use (Per 21 CFR 801.109)



(Division Sign-Off)

Division of Radiological Devices

Office of In Vitro Diagnostic Device Evaluation and Safety

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: P3-8CA for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)							
	Abdominal	P	P	P	P	P	Note 1	Note 4, 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic	P	P	P	P	P	Note 1	Note 4, 7
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult	P	P	P	P	P	Note 1	Note 4, 7
	Cardiac Pediatric	P	P	P	P	P	Note 1	Note 4, 7
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared under K092159; E= added under Appendix E

Additional Comments:

- Color Doppler includes Power (Amplitude) Doppler
- Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD
- Note 2: Includes imaging for guidance of biopsy
- Note 3: Includes infertility monitoring of follicle development
- Note 4: Color M-mode
- Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients
- Note 6: Abdominal organs and peripheral vessel
- Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging
- Note 8: 3D imaging
- Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
 Prescription Use (Per 21 CFR 801.109)


 (Division Sign-Off)

Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: P4-12 for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)							
	Abdominal	N	N	N	N	N	Note 1	Note 4, 7
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric	N	N	N	N	N	Note 1	Note 4, 7
	Small Organ (See Note 5)							
	Neonatal Cephalic	N	N	N	N	N	Note 1	Note 4, 7
	Adult Cephalic	N	N	N	N	N	Note 1	Note 4, 7
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult	N	N	N	N	N	Note 1	Note 4, 7
	Cardiac Pediatric	N	N	N	N	N	Note 1	Note 4, 7
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel							
	Other (spec.)							

N= new indication; P= previously cleared by FDA; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients


Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
Prescription Use (Per 21 CFR 801.109)


(Division Sign-Off)

Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety

Indications For Use

510K 17101455

Section 1.3, Page 10

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: CW2.0 for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)							
	Abdominal							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric							
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic					P		
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult					P		
	Cardiac Pediatric					P		
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel					P		
	Other (spec.)							

N= new indication; P= previously cleared under K092159; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
Prescription Use (Per 21 CFR 801.109)


(Division Sign-Off)

Division of Radiological Devices

Office of In Vitro Diagnostic Device Evaluation and Safety

Indications For Use

510K

K101455

DIAGNOSTIC ULTRASOUND INDICATIONS FOR USE STATEMENT

510(k) No.:

Device Name: CW4.0 for use with EKO 7

Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application		Mode of Operation (*includes simultaneous B-mode)						
General (Track I only)	Specific (Tracks I & III)	B	M	PWD	CWD	Color Doppler*	Combined* (Spec.)	Other (Spec.)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal (See Note 3)							
	Abdominal							
	Intra-operative (See Note 6)							
	Intra-operative (Neuro.)							
	Laparoscopic							
	Pediatric				P			
	Small Organ (See Note 5)							
	Neonatal Cephalic							
	Adult Cephalic				P			
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (non-Cardiac)							
	Musculo-skel. (Convent.)							
	Musculo-skel. (Superfic.)							
Intra-luminal								
Other (spec.)								
Cardiac	Cardiac Adult				P			
	Cardiac Pediatric				P			
	Trans-esophageal (Cardiac)							
	Other (spec.)							
Peripheral Vessel	Peripheral vessel				P			
	Other (spec.)							

N= new indication; P= previously cleared under K092159; E= added under Appendix E

Additional Comments:

Color Doppler includes Power (Amplitude) Doppler

Note 1: B+C+PW, B+PD+PW, B+C+CW, B+PD+CW, B+C+M, Dual B, Dual B+C, Dual B+PD

Note 2: Includes imaging for guidance of biopsy

Note 3: Includes infertility monitoring of follicle development

Note 4: Color M-mode

Note 5: For example: thyroid, parathyroid, breast, scrotum and penis in adult, pediatric and neonatal patients

Note 6: Abdominal organs and peripheral vessel

Note 7: Tissue Harmonic Imaging (THI) / Pulse Inversion Harmonic Imaging

Note 8: 3D imaging

Note 9: Panoramic imaging

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)
 Prescription Use (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Radiological Devices

Office of In Vitro Diagnostic Device Evaluation and Safety

Indications For Use

Section 1.3, Page 12

510K

K101455