

K 10 1967

510(K) Summary

Page 1 of 3

Date Prepared [21CFR 807.92(a)(1)]

7/12/2010

Submitter's Information [21CFR 807.92(a)(1)]

AUG 12 2010

Mr. Jonathan Achenbach
Sr. Dir. R&D, Clinical & Regulatory Affairs
Ellman International Inc.
333 Royal Ave. Oceanside, NY. 11572
Telephone: 516-267-6744

The establishment registration number for Ellman International is 2428235

Trade Name, Common Name, Classification [21 CFR 807.92(a)(2)]

Trade Name

Non-Ablative Wrinkle Treatment Handpiece

Device common, usual or Classification Names

Device: Electrosurgical, cutting & coagulation & accessories

Regulation Description: Electro surgical cutting and coagulation device and accessories

Class

Classification: Class II
Product Code: GEI

Device Description [21 CFR 807.92(a)(4)]

Accessory handpiece electrode is intended for use with the Ellman Surgitron IEC120 / 4.0 Dual Radio-Frequency generator. The device is considered as an accessory to the RF generator. The generator has been cleared by the FDA under 510(k) K082834 Ellman International Non-Ablative Technique for Surgitron IEC.

The purpose of this submission is to have the subject device cleared for the following change in **Cleaning / Disinfection methods** as per the Instructions for Use:

Method:

The Electrode Tip shall be cleaned and disinfected pre and post contact with the patient with an over the counter disinfectant wipe moistened with a full spectrum disinfectant containing a solution of Alkyl dimethyl benzyl ammonium chlorides and Alkyl dimethyl ethylbenzyl ammonium chlorides. The device is to remain wetted by the disinfectant solution for a minimum of 10 (ten) minutes.

Ellman International has presented a successfully executed test protocol as part of this submission to demonstrate that the methods described above achieve an adequate log reduction of inoculated challenge organisms on the test devices.

Ellman International has shown that the cleaning and disinfection method described above is similar to the cleaning and disinfection method provided with ALMA LASERS ACCENT as cleared by K070004 and ALMA LASERS FAMILY OF ACCENT RF SYSTEMS as cleared by K072699.

Further to this, Ellman International has presented substantial evidence to support the claims made in method above to provide an effective alternate cleaning and disinfection instruction to the existing instructions for use.

The subject device is packaged with the following items:

- IEC 3 Button Reusable Fingerswitch Handpiece
- Pelleve'Reusable Electrode

Intended Use [21 CFR 807.92(a)(5)]

The device has the following "Indications For Use"

- Non-ablative treatment of mild to moderate facial wrinkles and rhytides for skin phototypes I-IV

Technological Characteristics [21 CFR 807.92(a)(6)]

Ellman International maintains that the subject device is substantially equivalent to the predicate devices in that they all are similar in technology, function, intended use and method of cleaning and disinfection.

Labels, Labeling [21 CFR XXXX]

Per section "Product Labels"

K101967

Page 3 of 3

Performance Data [21 CFR 807.92(b)(1)]

As defined in SGS Protocol # 09A0908673 and demonstrated in SGS Report# 09A0908675

Predicate Devices [21 CFR 807.92(a)(3)(1)]

- ALMA LASERS LTD.: ACCENT as cleared by K070004
- ALMA LASERS LTD.: ALMA LASERS FAMILY OF ACCENT RADIOFREQUENCY (RF) SYSTEMS, MODELS: ACCENT, ACCENT XL as cleared by K072699
- ELLMAN INTERNATIONAL INC Surgitron, Non-Ablative Technique as cleared by K082834

The subject device is substantially equivalent (materials, technology) to the above listed devices.

Consideration for Special 510(k) Status

In accordance with the FDA guidance document: The New 510(k) Paradigm - Alternate Approaches to Demonstrating Substantial Equivalence in Premarket Notifications - Final Guidance Attachment 2, Ellman Internal maintains that this change in the cleaning and disinfection method is a "Device Modification" and therefore requests that the FDA considers it as such.

As there are no technological, structural, design or intended use changes for the subject device, Ellman International believes that no further data or clinical study is required to support the position that the device is safe and effective to use with the changes delineated in the instructions for use.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room - WO66-G609
Silver Spring, MD 20993-0002

Ellman International, Inc.
% Mr. Jonathan Achenbach
Sr. Director, R&D, Clinical
and Regulatory Affairs
3333 Royal Avenue
Oceanside, New York 11572

Re: K101967

Trade/Device Name: Non-Ablative Wrinkle Treatment Handpiece
Regulation Number: 21 CFR 878.4400
Regulation Name: Electrosurgical cutting and coagulation device and accessories
Regulatory Class: Class II
Product Code: GEI
Dated: July 12, 2010
Received: July 13, 2010

Dear Mr. Achenbach:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act

or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

For Peter D. Rummen
Mark N. Melkerson
Director
Division of Surgical, Orthopedic
And Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

*Dr. Rummen
ms
mpv
DEP D.R.*

Enclosure

"INDICATIONS FOR USE"

Statement

510(k) Number (if known): K 10 1967

AUG 12 2010

Device Name: Non-Ablative Wrinkle Treatment Handpiece

The Device has the following "Indications for Use":

Non-ablative treatment of mild to moderate facial wrinkles and rhytides for skin phototypes I-IV

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)


Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use
(per 21 CFR 801.109)

OR

Over-The Counter Use

(Optional Format 1-2-96)



(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K-101967