

K102209

DEC 3 2010

510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR § 807.92

**Submitter:** BIOMET 3i  
4555 Riverside Drive  
Palm Beach Gardens, FL 33410

**Establishment Reg. Number:** 1038806

**Contact Person:** Simeon Simone, Regulatory Affairs Specialists

**Contact Phone/Fax/Email:** Tel. 561-776-6831  
Fax. 561-383-3407  
Email: [simeon.simone@biomet.com](mailto:simeon.simone@biomet.com)

**Date Prepared:** August 2<sup>nd</sup>, 2010

**Trade/Proprietary Name:** Encode® Patient Specific Abutments

**Common/Usual Name:** Dental Abutment

**Classification Name:** Endosseous Dental Implant Abutment

**Device Classification/Code:** 872.3630

**Predicate Device(s) :** BIOMET 3i – Encode® Patient Specific Dental Abutments K032263 & K052648

**Device Description:**

Encode® Abutments are designed specifically for a patient in a CAD/CAM system. The abutments are designed from a three-dimensional intra-oral optical scan or resin model scan and then machined/milled according to the parameters created in a digital file which is derived from the scan. The abutments are manufactured from titanium alloy or ceramic.

**Indications for Use:**

BIOMET 3i Patient Specific Dental Abutments (Encode®) Designed Using Cadent iTero Scanner and inLab Software v3.5 are intended for use as an accessory to endosseous dental implants to support a prosthetic device in a partially or edentulous patient. These are intended for use to support single and multiple tooth prostheses, in the mandible or maxilla. The prostheses can be screw or cement retained to the abutment.

**Performance Data:**

Validation performed on scanning equipment and software to ensure accuracy of scanning models to produce the intended design as cleared in **K032263 & K052648**.

**Equivalence Data:**

BIOMET 3i Patient-Specific Dental Abutments (Encode®) have the same intended use and indications, principles of operation, and technological characteristics as previously cleared BIOMET 3i Patient-Specific Dental Abutments. The differences in scanning and CAD/CAM processes do not raise any new questions of safety or effectiveness. Validation data demonstrates that the modified process results in a finished device that is as safe and effective as BIOMET 3i's Patient-Specific Patient-Specific Dental Abutments that are currently cleared with previous scanner systems. Thus, the BIOMET 3i Patient-Specific Dental Abutments are substantially equivalent to its predicate devices.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room -WO66-G609  
Silver Spring, MD 20993-0002

Mr. Simeon Simone  
Regulatory Affairs Specialists  
BIOMET 3I, Incorporated  
4555 Riverside Drive  
Palm Beach Gardens, Florida 33410

DEC - 3 2010

Re: K102209

Trade/Device Name: BIOMET 3i Patient Specific Dental Abutments (Encode®)  
Designed Using Cadent iTero Scanner and inLab Software v.3.5  
Regulation Number: 21 CFR 872.3630  
Regulation Name: Endosseous Dental Implant Abutment  
Regulatory Class: II  
Product Code: NHA  
Dated: November 19, 2010  
Received: November 23, 2010

Dear Mr. Simone:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Page 2- Mr. Simone

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

K102209

Indications for Use

DEC 3 - 2010

510(k) Number (if known): \_\_\_\_\_

Device Name: BIOMET 3i Patient Specific Dental Abutments (Encode®) Designed Using Cadent iTero Scanner and inLab Software v.3.5

Indications for Use:

BIOMET 3i Patient Specific Dental Abutments (Encode®) Designed Using Cadent iTero Scanner and inLab Software v3.5 are intended for use as an accessory to endosseous dental implants to support a prosthetic device in a partially or edentulous patient. These are intended for use to support single and multiple tooth prostheses, in the mandible or maxilla. The prostheses can be screw or cement retained to the abutment.

Prescription Use  X   
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

*Susan Payne*  
Division Sign-Off, General Hospital  
Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

510(k) Number: K102209