

K102253



<b>Echo Blaster</b> 510(k) Premarket Submission		
Date: DEC 20 2010	Rev.: I	Pg. 1,2-3

### 510k Summary

MAR 14 2011

#### 1. Identifying information

Manufacturer           TELEMED  
 Address                 Dariaus ir Gireno str. 42 Vilnius LT-02189 Lithuania  
 Telephone             +370-5 2106272 +370-5 2106273  
 Fax                     +370-5 2306733  
 E-mail                 [info@telemed.lt](mailto:info@telemed.lt)  
 Contact                Yury Sokolov / Engineering Manager  
 Name of Device        Echo Blaster 64 / Echo Blaster 128

#### 2. Class and Predicate Information

Classification	Ultrasonic Pulsed Echo Imaging System	21 CFR 892.1560, IYO
Name	Diagnostic Ultrasonic Transducers	21 CFR 892.1570, ITX
Common Name	Ultrasound imaging system	
Proprietary Name	Echo Blaster	
Class	Regulatory Class II	
Predicate Devices	Ardent Sound, Inc.; Voyager	K050551
	Ardent Sound, Inc.; Seeker/Spark	K060800
	Medison America, Inc; MYSONO 201	K003121

#### 3. Performance Standards

Performance Standards   None

The Echo Blaster family has been designed to meet the following:

Safety and EMC Requirements for Medical Equipment:

- IEC 60601-1: 2000, Part 1: General requirements for safety.
- IEC 60601-1-2: 2001, Part 1: General requirements for basic safety and essential performance, 2.Collateral standard: Electromagnetic compatibility - Requirements and tests
- IEC 60601-1-2: 2001, Part 1: General requirements for basic safety and essential performance, 2.Collateral standard: Electromagnetic compatibility - Requirements and tests
- IEC 60601-1-2: 2001, Part 1: General requirements for basic safety and essential performance, 2.Collateral standard: Electromagnetic compatibility - Requirements and tests
- IEC 60601-1-2: 2001, Part 1: General requirements for basic safety and essential performance, 2.Collateral standard: Electromagnetic compatibility - Requirements and tests
- IEC 60601-1-4: 1996, Part 1: General Requirements for Safety, 4.Collateral Standard: Programmable Electrical Medical Systems
- IEC 60601-2-37: 2007-08 Particular requirements for the basic safety and essential performance of ultrasonic medical diagnostic and monitoring equipment
- NEMA UD 2-2004: 2003, Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment
- NEMA UD 3-2004: 2004, Standard for Real Time Display of Thermal and Mechanical



- Acoustic Output Indices on Diagnostic Ultrasound Equipment
- AIUM MUS: 2002, Medical Ultrasound Safety
- ISO-10993-1, Biological Evaluation of Medical Devices Part 1: Evaluation and Testing, 2003
- ISO-10993-5, Biological Evaluation of Medical Devices Part 5: Tests for in vitro cytotoxicity, 1999
- ISO-10993-10, Biological Evaluation of Medical Devices Part 10: Tests for irritation and delayed-type hypersensitivity, 2002
- ISO-10993-11, Biological Evaluation of Medical Devices Part 11: Tests for systemic toxicity, 2006
- IEC 62304: 2006 Medical device software -- Software life cycle processes
- ISO 14971:2007 Medical devices -- Application of risk management to medical devices

Essential Requirements of Council directive 93/42/EEC (Medical Device Directive)

The system's acoustic output is in accordance with ALARA principle (as low as reasonably achievable)

**4. Indication for Use**

**Echo Blaster 128 / Echo Blaster 64** ultrasound imaging systems are intended to be used for applications in fetal, abdominal, pediatric, small organ (breast, thyroid and testicles), cephalic (neonatal and adult), musculoskeletal (conventional), musculoskeletal (superficial), cardiac adult and peripheral vessel (including imaging for needle guidance, see ch.1.3). It is possible to provide diagnostic information (B, B+M and M-mode imaging) outside of an imaging lab, including at the bedside systems, for navigated medical application, in operating rooms/critical care units.

<b>System</b> Echo Blaster 64		Fetal*, Abdominal*, Pediatric*, Small Organ* (Breast, Thyroid, Testicles), Neonatal Cephalic, Adult Cephalic, Musculo-skeletal* (Conventional), Musculo-skeletal* (Superficial), Cardiac Adult*, Peripheral vessel*
<b>Transducers</b>	C3.5/60/64	Fetal*, Abdominal*, Pediatric*
	HL7.5/40/64	Pediatric*, Small Organ* (Breast, Thyroid, Testicles), Neonatal Cephalic, Adult Cephalic, Musculo-skeletal* (Conventional), Musculo-skeletal* (Superficial), Peripheral vessel*
	PV6.5/10/64	Small Organ* (Breast, Thyroid, Testicles), Cardiac Adult*, Peripheral vessel*
<b>System</b> Echo Blaster 128		Fetal*, Abdominal*, Pediatric*, Small Organ* (Breast, Thyroid, Testicles), Neonatal Cephalic, Adult Cephalic, Musculo-skeletal* (Conventional), Musculo-skeletal* (Superficial), Cardiac Adult*, Peripheral vessel*
<b>Transducers</b>	C3.5/60/128Z	Fetal*, Abdominal*, Pediatric*
	HL9.0/40/128Z	Pediatric*, Small Organ* (Breast, Thyroid, Testicles), Neonatal Cephalic, Adult Cephalic, Musculo-skeletal* (Conventional), Musculo-skeletal* (Superficial), Peripheral vessel*
	PV6.5/10/128Z	Small Organ* (Breast, Thyroid, Testicles), Cardiac Adult*, Peripheral vessel*
	L5.0/80/128Z	Musculo-skeletal*(Conventional), Musculo-skeletal* (Superficial)

\*Including Imaging for needle guidance

K102253



### 5. Device Description

**Echo Blaster 128 / Echo Blaster 64** systems are intended for the multipurpose ultrasound examinations, based on electronic linear and convex scanning.

**Echo Blaster 128 / Echo Blaster 64** system is a combination of proprietary hardware and software that has been designed for real-time imaging and is intended to be a basic diagnostic tool. The System is based on a modular and flexible architecture allowing for both mobile and stationary (installed) configurations. The system is designed for imaging with transducer ranges of 2 to 10 MHz.

The devices referenced in this submission represent a transportable, software-controlled, diagnostic ultrasound system with accessories. This submission does not include technology or control feature changes nor deviations from indications for use different from those demonstrated in previously cleared devices operating in ultrasound B-Mode or M-Mode, inclusive of the predicate devices so claimed.

The **Echo Blaster 128 / Echo Blaster 64** only contains the hardware and firmware, everything else (e.g. ultrasound software, database) is located on a standard PC that is connected to the **Echo Blaster 128 / Echo Blaster 64** via USB 2.0. Minimum requirements are given for the PC. The probes are connected to the **Echo Blaster 128 / Echo Blaster 64**. All Sonograms are saved on the PC and can there be evaluated, printed and archived. The **Echo Wave** software was especially designed for the **Echo Blaster 128 / Echo Blaster 64**. Software able to reside in a Windows-based PC.

The **Echo Blaster 128 / Echo Blaster 64** can be used together with the appropriate probes for the entire ultrasound diagnostic (2MHz to 10MHz probes). Two probes can work simultaneously for **Echo Blaster 128 2Z** modifications.

The devices included in this submission are as follows:

**Echo Blaster 128 EXT-1Z** ultrasound system utilizing as hardware and firmware an ultrasound engine contained in a small stand alone enclosure for connection to a host PC via a USB port;

**Echo Blaster 128 INT-1Z / INT-2Z** ultrasound systems utilizing as hardware and firmware an ultrasound engine contained in a small enclosure for insertion to a host PC (to a drive bays).

A probe, 128 element convex array, at a central ultrasonic frequency of approximately 3.5MHz, model **C3.5/60/128Z**.

A probe, 128 element linear array at a central ultrasonic frequency of approximately 9 MHz, model **HL9.0/40/128Z**.

A probe, 128 element convex array at a central ultrasonic frequency of approximately 6.5 MHz, model **PV6.5/10/128Z**.

A probe, 128 element convex array at a central ultrasonic frequency of approximately 6.5 MHz, model **L5/80/128Z**.

**Echo Blaster 64 EXT-1T** ultrasound system utilizing as hardware and firmware an ultrasound engine contained in a small stand alone enclosure for connection to a host PC via a USB port;

A probe, 64 element convex array, at a central ultrasonic frequency of approximately 3.5 MHz,

K102253



<b>Echo Blaster</b> 510(k) Premarket Submission		
Date: DEC 20 2010	Rev.: I	Pg. 1,2-6

model C3.5/60/64.

A probe, 64 element linear array at a central ultrasonic frequency of approximately 7.5 MHz, model HL7.5/40/64.

A probe, 64 element convex array at a central ultrasonic frequency of approximately 6.5 MHz, model PV6.5/10/64.

### 6. General Safety and Effectiveness

The **Echo Blaster 128 / Echo Blaster 64** Ultrasound Systems are similar to currently distributed ultrasonic pulsed echo imaging systems.

There are no technological characteristics or features or indications for use in this Submission that are not previously evaluated and approved in the predicate devices, nor are there such technologies, features and indications for use not commonly used in the practice of diagnostic ultrasound.

The **Echo Blaster** Ultrasound Systems and its accessories are designed for compliance to all applicable medical devices safety standards, as referenced in DECLARATION OF CONFORMITY (Appendix 06). Prior release for manufacturing, all such devices, so designed, are tested and determined to be in full compliance with acoustic output, biocompatibility, cleaning and disinfection effectiveness. No additional clinical testing is required, as the indications for use are not a novel indication as shown by the predicate devices in Section 1.5 Predicate Device Comparison.

**Maximum acoustic output level is under by the FDA recommended limit and power level is displayed all the time.**

### 7. Patient Contact Materials

The materials of probes, coming in contact with patient are:

- Silicone Rubber
- Acrylonitrile Butadien Styrene (ABS)
- Polyphenylsulfone (PPSU)


Standard for the biological evaluation:

- ISO-10993-1, Biological Evaluation of Medical Devices Part 1: Evaluation and Testing, 2003
- ISO-10993-5, Biological Evaluation of Medical Devices Part 5: Tests for in vitro cytotoxicity , 1999
- ISO-10993-10, Biological Evaluation of Medical Devices Part 10: Tests for irritation and delayed-type hypersensitivity, 2002
- ISO-10993-11, Biological Evaluation of Medical Devices Part 11: Tests for systemic toxicity, 2006

### 8. Software

The **Echo Blaster 128 / Echo Blaster 64** systems contain the hardware and software which collect and pro-processes "rough" data and send it via USB 2.0 connection to a Windows® based PC. The main application software is **Echo Wave** software running on the PC, it is receiving data, processing and showing image/data on the screen. The main user interface shows an ultrasound image, controls and drop-out menus. The ultrasound images and calculated/measured data can be stored

K102253

	<b>Echo Blaster</b> <b>510(k) Premarket Submission</b>	
	Date: DEC 20 2010	Rev.: I

in memory.

**9. Conclusion**

In accordance with the FDA and based on the information provided in this Premarket notification, TELEMED concludes that the **Echo Blaster 128 / Echo Blaster 64** are safe and effective and substantially equivalent to predicate devices described herein.



Food and Drug Administration  
10903 New Hampshire Avenue  
Silver Spring, MD 20993

TELEMED  
% Mr. Norbert Stuibler  
Responsible Third Party Official  
TÜV SÜD America, Inc.  
1775 Old Highway 8 NW, Ste 104  
NEW BRIGHTON MN 55112-1891

MAR 14 2011

Re: K102253  
Trade/Device Name: 'Echo Blaster 64' and 'Echo Blaster 128'  
Regulation Number: 21 CFR 892.1560  
Regulation Name: Ultrasonic pulsed echo imaging system  
Regulatory Class: II  
Product Code: IYO and ITX  
Dated: February 4, 2011  
Received: March 2, 2011

Dear Mr. Stuibler:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the 'Echo Blaster 64' and 'Echo Blaster 128', as described in your premarket notification:

Transducer Model Number

'Echo Blaster 64'

C3.5/60/64 Convex Array  
HL7.5/40/64 Linear Array  
PV6.5/10/64 Convex Array

'Echo Blaster 128'

C3.5/60/128Z Convex Array  
HL9.0/40/128Z Linear Array  
PV6.5/10/128Z Convex Array  
L5/80/128Z Linear Array

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

If you have any questions regarding the content of this letter, please contact Shahram Vaezy at (301) 796-6242.

Sincerely Yours,



Mary S. Pastel, Sc. D.  
Director  
Division of Radiological Devices  
Office of In Vitro Diagnostic Device  
Evaluation and Safety  
Center for Devices and Radiological Health

Enclosure(s)



### 1.3 Indications for Use

All indications for use of subject devices and probes are identified in the table forms:

<b>System</b> Echo Blaster 64		Fetal*, Abdominal*, Pediatric*, Small Organ* (Breast, Thyroid, Testicles), Neonatal Cephalic, Adult Cephalic, Musculo-skeletal* (Conventional), Musculo-skeletal* (Superficial), Cardiac Adult*, Peripheral vessel*	Table 1.3-1
<b>Transducers</b>	C3.5/60/64	Fetal*, Abdominal*, Pediatric*	Table 1.3-2
	HL7.5/40/64	Pediatric*, Small Organ* (Breast, Thyroid, Testicles), Neonatal Cephalic, Adult Cephalic, Musculo-skeletal* (Conventional), Musculo-skeletal* (Superficial), Peripheral vessel*	Table 1.3-3
	PV6.5/10/64	Small Organ* (Breast, Thyroid, Testicles), Cardiac Adult*, Peripheral vessel*	Table 1.3-4
<b>System</b> Echo Blaster 128		Fetal*, Abdominal*, Pediatric*, Small Organ* (Breast, Thyroid, Testicles), Neonatal Cephalic, Adult Cephalic, Musculo-skeletal* (Conventional), Musculo-skeletal* (Superficial), Cardiac Adult*, Peripheral vessel*	Table 1.3-5
<b>Transducers</b>	C3.5/60/128Z	Fetal*, Abdominal*, Pediatric*	Table 1.3-6
	HL9.0/40/128Z	Pediatric*, Small Organ* (Breast, Thyroid, Testicles), Neonatal Cephalic, Adult Cephalic, Musculo-skeletal* (Conventional), Musculo-skeletal* (Superficial), Peripheral vessel*	Table 1.3-7
	PV6.5/10/128Z	Small Organ* (Breast, Thyroid, Testicles), Cardiac Adult*, Peripheral vessel*	Table 1.3-8
	L5.0/80/128Z	Musculo-skeletal*(Conventional), Musculo-skeletal* (Superficial)	Table 1.3-9

\*Including Imaging for needle guidance

\_\_\_\_\_  
(Division Sign-Off)

Division of Radiological Devices  
Office of In Vitro Diagnostic Device Evaluation and Safety

510K           K102253





**Diagnostic Ultrasound Indications for Use Form**

**System:** Echo Blaster 64

**Intended Use:** Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:


**Table 1.3-1**

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (specify)	Other * (specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal <sup>2</sup>	N	N				N B+M	
	Abdominal <sup>2</sup>	N	N				N B+M	
	Intraoperative (specify)							
	Intraoperative (Neuro)							
	Laparoscopic							
	Pediatric <sup>2</sup>	N	N				N B+M	
	Small Organ (specify) <sup>2</sup>	N <sup>1</sup>	N <sup>1</sup>				N B+M	
	Neonatal Cephalic	N	N				N B+M	
	Adult Cephalic	N	N				N B+M	
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (Non-Card)							
	Musculo-skeletal (Conventional) <sup>2</sup>	N	N				N B+M	
	Musculo-skeletal (Superficial) <sup>2</sup>	N	N				N B+M	
	Intravascular							
Other (specify)								
Cardiac	Cardiac Adult <sup>2</sup>	N	N				N B+M	
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (specify)							
Peripheral Vessel	Peripheral vessel <sup>2</sup>	N	N				N B+M	
	Other (specify)							

N= new indication; P= previously cleared by FDA; E= added under this appendix

Note <sup>1</sup> - Small Organs (specifically Breast, Thyroid, Testicles)

Note <sup>2</sup> - Including imaging for needle guidance

  
 (Division Sign-Off)  
 Division of Radiological Devices  
 Office of In Vitro Diagnostic Device Evaluation and Safety

510K           K102253





**Diagnostic Ultrasound Indications for Use Form**

**System:** Echo Blaster 64

**Transducer:** Linear array HL7.5/40/64

**Intended Use:** Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:


**Table 1.3-3**

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (specify)	Other * (specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intraoperative (specify)							
	Intraoperative (Neuro)							
	Laparoscopic							
	Pediatric <sup>2</sup>	N	N				N B+M	
	Small Organ (specify) <sup>2</sup>	N <sup>1</sup>	N <sup>1</sup>				N B+M	
	Neonatal Cephalic	N	N				N B+M	
	Adult Cephalic	N	N				N B+M	
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (Non-Card)							
	Musculo-skeletal (Conventional) <sup>2</sup>	N	N				N B+M	
	Musculo-skeletal (Superficial) <sup>2</sup>	N	N				N B+M	
	Intravascular							
Other (specify)								
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (specify)							
Peripheral Vessel	Peripheral vessel <sup>2</sup>	N	N				N B+M	
	Other (specify)							

N= new indication; P= previously cleared by FDA; E= added under this appendix

Note <sup>1</sup> - Small Organs (specifically Breast, Thyroid, Testicles)

Note <sup>2</sup> - Including imaging for needle guidance

  
 (Division Sign-Off)  
 Division of Radiological Devices  
 Office of In Vitro Diagnostic Device Evaluation and Safety

510K         K102253



**Diagnostic Ultrasound Indications for Use Form**

**System:** Echo Blaster 64

**Transducer:** Convex array PV6.5/10/64

**Intended Use:** Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

**Table 1.3-4**

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (specify)	Other * (specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intraoperative (specify)							
	Intraoperative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (specify) <sup>2</sup>	N <sup>1</sup>	N <sup>1</sup>					N B+M
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (Non-Card)							
	Musculo-skeletal (Conventional) <sup>2</sup>							
	Musculo-skeletal (Superficial) <sup>2</sup>							
	Intravascular							
Other (specify)								
Cardiac	Cardiac Adult <sup>2</sup>	N	N					N B+M
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (specify)							
Peripheral Vessel	Peripheral vessel <sup>2</sup>	N	N					N B+M
	Other (specify)							

N= new indication; P= previously cleared by FDA; E= added under this appendix

Note <sup>1</sup> - Small Organs (specifically Breast, Thyroid, Testicles)

Note <sup>2</sup> - Including imaging for needle guidance

*Mary Spital*  
\_\_\_\_\_  
(Division Sign-Off)

Division of Radiological Devices  
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K102253



**Echo Blaster**  
510(k) Premarket Submission

Date: JUL 16 2010      Rev.: C      Pg. 1.3-6

**Diagnostic Ultrasound Indications for Use Form**

**System:** Echo Blaster 128

**Intended Use:** Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

**Table 1.3-5**

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (specify)	Other * (specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal <sup>2</sup>	N	N				N B+M	
	Abdominal <sup>2</sup>	N	N				N B+M	
	Intraoperative (specify)							
	Intraoperative (Neuro)							
	Laparoscopic							
	Pediatric <sup>2</sup>	N	N				N B+M	
	Small Organ (specify) <sup>2</sup>	N <sup>1</sup>	N <sup>1</sup>				N B+M	
	Neonatal Cephalic	N	N				N B+M	
	Adult Cephalic	N	N				N B+M	
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (Non-Card)							
	Musculo-skeletal (Conventional) <sup>2</sup>	N	N				N B+M	
	Musculo-skeletal (Superficial) <sup>2</sup>	N	N				N B+M	
Intravascular								
Other (specify)								
Cardiac	Cardiac Adult <sup>2</sup>	N	N				N B+M	
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
Other (specify)								
Peripheral Vessel	Peripheral vessel <sup>2</sup>	N	N				N B+M	
	Other (specify)							

N= new indication; P= previously cleared by FDA; E= added under this appendix

Note <sup>1</sup> -Small Organs (specifically Breast, Thyroid, Testicles)

Note <sup>2</sup> - Including imaging for needle guidance

(Division Sign-Off)

Division of Radiological Devices  
Office of In Vitro Diagnostic Device Evaluation and Safety

510K 12102253



**Diagnostic Ultrasound Indications for Use Form**

**System:** Echo Blaster 128                      **Transducer:** Convex array C3.5/60/128Z

**Intended Use:** Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:


**Table 1.3-6**

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (specify)	Other * (specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal <sup>2</sup>	N	N				N B+M	
	Abdominal <sup>2</sup>	N	N				N B+M	
	Intraoperative (specify)							
	Intraoperative (Neuro)							
	Laparoscopic							
	Pediatric <sup>2</sup>	N	N				N B+M	
	Small Organ (specify)							
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (Non-Card)							
	Musculo-skeletal (Conventional)							
	Musculo-skeletal (Superficial)							
	Intravascular							
Other (specify)								
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (specify)							
Peripheral Vessel	Peripheral vessel							
	Other (specify)							

N= new indication; P= previously cleared by FDA; E= added under this appendix

Note <sup>1</sup> -Small Organs (specifically Breast, Thyroid, Testicles)

Note <sup>2</sup> - Including imaging for needle guidance

  
 (Division Sign-Off)  
 Division of Radiological Devices  
 Office of In Vitro Diagnostic Device Evaluation and Safety  
 510K K102253



<b>Echo Blaster</b> 510(k) Premarket Submission		
Date: JUL 16 2010	Rev.: C	Pg. 1.3-8

**Diagnostic Ultrasound Indications for Use Form**

**System:** Echo Blaster 128

**Transducer:** Linear array HL9.0/40/128Z

**Intended Use:** Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

**Table 1.3-7**

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (specify)	Other * (specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intraoperative (specify)							
	Intraoperative (Neuro)							
	Laparoscopic							
	Pediatric <sup>2</sup>		N	N				N B+M
	Small Organ (specify) <sup>2</sup>		N <sup>1</sup>	N <sup>1</sup>				N B+M
	Neonatal Cephalic		N	N				N B+M
	Adult Cephalic		N	N				N B+M
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (Non-Card)							
	Musculo-skeletal (Conventional) <sup>2</sup>		N	N				N B+M
	Musculo-skeletal (Superficial) <sup>2</sup>		N	N				N B+M
	Intravascular							
Other (specify)								
Cardiac	Cardiac Adult							
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (specify)							
Peripheral Vessel	Peripheral vessel <sup>2</sup>	N	N				N B+M	
	Other (specify)							

N= new indication; P= previously cleared by FDA; E= added under this appendix

Note <sup>1</sup> - Small Organs (specifically Breast, Thyroid, Testicles)

Note <sup>2</sup> - Including imaging for needle guidance

*Mary Spital*  
\_\_\_\_\_  
(Division Sign-Off)

Division of Radiological Devices  
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K102253



**Diagnostic Ultrasound Indications for Use Form**

**System:** Echo Blaster 128

**Transducer:** Convex array PV6.5/10/128Z

**Intended Use:** Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

**Table 1.3-8**

Clinical Application		Mode of Operation						
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (specify)	Other * (specify)
Ophthalmic	Ophthalmic							
Fetal Imaging & Other	Fetal							
	Abdominal							
	Intraoperative (specify)							
	Intraoperative (Neuro)							
	Laparoscopic							
	Pediatric							
	Small Organ (specify) <sup>2</sup>	N <sup>1</sup>	N <sup>1</sup>					N B+M
	Neonatal Cephalic							
	Adult Cephalic							
	Trans-rectal							
	Trans-vaginal							
	Trans-urethral							
	Trans-esoph. (Non-Card)							
	Musculo-skeletal (Conventional) <sup>2</sup>							
	Musculo-skeletal (Superficial) <sup>2</sup>							
	Intravascular							
Other (specify)								
Cardiac	Cardiac Adult <sup>2</sup>	N	N					N B+M
	Cardiac Pediatric							
	Intravascular (Cardiac)							
	Trans-esoph. (Cardiac)							
	Intra-cardiac							
	Other (specify)							
Peripheral Vessel	Peripheral vessel <sup>2</sup>	N	N					N B+M
	Other (specify)							

N= new indication; P= previously cleared by FDA; E= added under this appendix

Note <sup>1</sup> - Small Organs (specifically Breast, Thyroid, Testicles)

Note <sup>2</sup> - Including imaging for needle guidance

*Mary Spotal*  
(Division Sign-Off)

Division of Radiological Devices  
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K102253





**Diagnostic Ultrasound Indications for Use Form**

**System:** Echo Blaster 128

**Transducer:** Linear array L5/80/128Z

**Intended Use:** Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

**Table 4.3-1**

Clinical Application		Mode of Operation							
General (Track 1 Only)	Specific (Track 1 & 3)	B	M	PWD	CWD	Color Doppler	Combined (specify)	Other * (specify)	
Ophthalmic	Ophthalmic								
Fetal Imaging & Other	Fetal								
	Abdominal								
	Intraoperative (specify)								
	Intraoperative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ (specify)								
	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph. (Non-Card)								
	Musculo-skeletal (Conventional) <sup>2</sup>		N						
	Musculo-skeletal (Superficial) <sup>2</sup>		N						
	Intravascular								
Other (specify)									
Cardiac	Cardiac Adult								
	Cardiac Pediatric								
	Intravascular (Cardiac)								
	Trans-esoph. (Cardiac)								
	Intra-cardiac								
	Other (specify)								
Peripheral Vessel	Peripheral vessel								
	Other (specify)								

N= new indication; P= previously cleared by FDA; E= added under this appendix

Note <sup>1</sup> - Small Organs (specifically Breast, Thyroid, Testicles)

Note <sup>2</sup> - Including imaging for needle guidance

(Division Sign-Off)

Division of Radiological Devices  
Office of In Vitro Diagnostic Device Evaluation and Safety

510K K102253