



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center – WO66-0609
Silver Spring, MD 20993-0002

Immunostics, Inc.
c/o Mr. Richard M. Peoples
QA/RA Manager
3505 Sunset Avenue
Ocean, New Jersey 07712

JAN 28 2011

Re: k102664
Trade/Device Name: Hema Screen™ ER
Regulation Number: 21 CFR 864.6550
Regulation Name: Occult blood test
Regulatory Class: Class II
Product Code: KHE
Dated: December 22, 2010
Received: January 12, 2011

Dear Mr. Peoples:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); and good manufacturing practice


Page 2 – Mr. Richard Peoples

requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Maria M. Chan, Ph.D
Director
Division of Immunology and Hematology Devices
Office of *In Vitro* Diagnostic Device Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indications for Use Form

510(k) Number (if known): K102664

Device Name: hema-screen™ ER FOBt


Indications for Use:

hema-screen™ ER is a rapid, convenient, and non-offensive qualitative method for detecting occult blood in the stool. It is intended for professional use as an aid in the diagnosis of asymptomatic gastrointestinal conditions that may manifest themselves by the presence of occult blood in the stool. This test is recommended for use in routine hospital testing, mass screening programs for colorectal cancer, and in testing of postoperative patients and newborn infants.

Prescription Use X AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)


Division Sign-Off
Office of In Vitro Diagnostic Device
Evaluation and Safety

510(k) K102664