

**510(k) Summary**  
2.6F Vascu-PICC®  
Summary of Safety and Effectiveness  
Prepared October 1, 2010

DEC 21 2010

K102966

---

**General Information**

Submitter: MEDCOMP®  
1499 Delp Drive  
Harleysville, PA 19438  
Phone: (215) 256-4201  
Fax: (215) 256-9191

Contact: Jean Callow  
Regulatory Specialist

Device Trade Name: 2.6F Vascu-PICC®  
Common Name: Peripherally Inserted Central Catheter (PICC)  
Classification Name: LJS - Catheter, Intravascular, Therapeutic, Long-Term  
Greater than 30 Days  
CFR Reference: 21 CFG 880.5970, Class II  
Classification Panel: General Hospital

---

**Predicate Devices:**

Device Trade Name: 1.9F Vascu-PICC®  
Common Name: Peripherally Inserted Central Catheter (PICC)  
Classification Name: LJS -Catheter, Intravascular, Therapeutic, Long-  
Term Greater than 30 Days  
CFR Reference: 21 CFR 880.5970, Class II  
Classification Panel: General Hospital  
Premarket Notification: K091466

---

**Indications for Use:** The 2.6F Vascu-PICC® catheters are indicated for short or long term access to the central venous system via peripherally insertion in neonates, infants, and children. It may be used for administration of fluids, medication, and nutritional therapy.

---

**Device Description:**

- Designed for peripheral vein catherization
- Comprised of a polyurethane material
- The lumen is connected to the extensions by a hub with a suture wing for placement.
- Clamps are provided on the extension tube to prevent air/fluid communication.
- Female luer connectors provide the connection for intravenous administration.

---

**Safety and Performance Tests**

Biocompatibility requirements of *ISO 10993 Biological Evaluation of Medical Devices Part 1: Evaluation and Testing* for externally communicating, blood contacting, long-term devices were met.

Performance testing of the 2.6F Vascu-PICC® was conducted in accordance with the following international standards:

- *ISO 10555-1: 1997, Sterile Single Use-Intravascular Catheters, General Requirements*
- *ISO 10555-3: 1997, Sterile Single Use-Intravascular Catheters, Central Venous Catheters*
- *ISO 594-2: Conical Fittings with a 6% (Luer) Taper for Syringes, Needles, and Certain Other Medical Equipment – Part 2: Lock Fittings*

Subject product testing has yielded acceptable safety and performance outcomes.

Risk Management of the subject device was conducted in accordance with an internal protocol based on *ISO 14971: 2000, Medical Devices – Risk Management for Medical Devices*. The analysis did not identify any new types or safety or efficacy questions for the proposed device.

The results of these tests, in conjunction with the substantial equivalence claims effectively demonstrate that the 2.6F Vascu-PICC® is substantially equivalent to the cited predicate device.

---

#### **Summary of Substantial Equivalence**

Based on the indications for use and safety and performance testing, the 2.6F Vascu-PICC® meets the requirements that are considered for its intended use and is substantially equivalent in design materials, sterilization, and indications for use to the predicate device.

---



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room -WO66-G609  
Silver Spring, MD 20993-0002

Ms. Jean Callow  
Regulatory Specialist  
Medcomp  
1499 Delp Drive  
Harleysville Pennsylvania 19438

DEC 21 2010

Re: K102966  
Trade/Device Name: 2.6F Vascu-PICC®  
Regulation Number: 21 CFR 880.5970  
Regulation Name: Percutaneous, Implanted Long-Term Intravascular Catheter  
Regulatory Class: II  
Product Code: LJS  
Dated: December 1, 2010  
Received: December 2, 2010

Dear Ms. Callow:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

DEC 21 2010

510(k) Number (if known): K102966

Device Name: 2.6F Vascu-PICC®

### Indications for Use:

The 2.6F Vascu-PICC® catheters are indicated for short or long term access to the central venous system via peripheral insertion in neonates, infants, and children. It may be used for administration of fluids, medication, and nutritional therapy.

Prescription Use X AND/OR Over-The-Counter Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER  
PAGE OF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Richard C. Chapman 12/20/10  
(Division Sign-Off)

Page 1 of 1

Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

10(k) Number: K102966