

NOV - 9 2010

**PANGAO®**

Premarket Notification  
 Traditional Sec. 510(k) Submission  
 PG-800B Electronic Blood Pressure Monitor  
 Sec. III 510(k) Summary

**Sec. III 510(k) Summary**

This 510(k) Summary is prepared per the request of 21 CFR 807.92. The Assigned 510(k) Number is: \_\_\_\_\_

<b>Date</b>	20 JUL 2010
<b>Sponsor</b>	<p><b>Shenzhen Pango Electronic Co., Ltd</b>          No.25, 1st Industrial Park, Fenghuang Road, Xikeng Village, Henggang Town,          Longgang District, Shenzhen, Guangdong, 518115, China          Contact Person: Ms. Xiaoyun Yang, Vice General Manager          T: +86-755-33825988   F: +86-755-33825989   E: sales@pan-go.com</p>
<b>Submission Correspondent</b>	<p>MS. Diana Hong / MR. Lee Fu          Shanghai Mid-Link Consulting Co., Ltd          P.O.BOX 237-023, Shanghai, 200237, China</p>
<b>Proposed Device Classification</b>	<b>Electronic Blood Pressure Monitor, PG-800B</b>
<b>Intended Use</b>	System, Measurement, Blood-pressure, Non-invasive   DXN   21 CFR 870.1130   Class II
<b>Device Description</b>	<p>PG-800B Electronic Blood Pressure Monitor is intended to measure the systolic and diastolic blood pressure as well as the pulse rate of adult person via non-invasive oscillometric technique in which an inflatable cuff is wrapped around the upper arm. It can be used at medical facilities or at home. The intended arm circumference is 22-32 cm.</p> <p>The proposed device, PG-800B Electronic Blood Pressure Monitor, is a battery driven automatic non-invasive blood pressure monitor. It can measure systolic and diastolic blood pressure as well as the pulse rate of adult person at arm within its claimed range and accuracy via the oscillometric technique. The device has the data storage function. It has an bar indicating function, which can indicate the WHO (World Health Organization) Blood Pressure Classification of the measured blood pressure by referencing Diastolic Blood Pressure issued at Journal of Hypertension 1999. Vol 17, No.2.</p>
<b>Testing</b>	<p>Electric Safety per IEC 60601-1:1988+A1:1991+A2:1995 / EMC per IEC 60601-1-2:2001+A1:2004          Performance and Clinical Verification Test per ANSI/AAMI SP10:2002+A1:2003+A2:2006          Biocompatibility per ISO 10993-5:2006 / ISO 10993-10:2002+AMD:2006</p>
<b>Predicate Device</b>	KD-595 Blood Pressure Monitor as cleared in K070828.
<b>SE Conclusion</b>	<p>The proposed device, PG-800A Electronic Blood Pressure Monitor, measures the blood pressure via same principle, oscillometric, at the same site, upper arm, as the predicate device. Both of them are driven by batteries. The proposed device, PG-800B Electronic Blood Pressure Monitor, is claimed to be Substantially Equivalent (SE) to the predicate device, KD595 Blood Pressure Monitor as cleared in K070828, in aspect of safety and effectiveness.</p>



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room –WO66-G609  
Silver Spring, MD 20993-0002

Shenzhen Pango Electronic Co., Ltd.  
c/o Mr. Ned Devine  
Third Party reviewer  
Underwriters Laboratories, Inc.  
333 Pfingsten Rd.  
Northbrook, IL 60062

NOV - 9 2010

Re: K102988  
Trade/Device Name: Electronic Blood Pressure Monitor, PG-800B  
Regulation Number: 21 CFR 870.1130  
Regulation Name: Non invasive Blood Pressure Measurement System  
Regulatory Class: Class II (two)  
Product Code: DXN  
Dated: October 22, 2010  
Received: October 28, 2010

Dear Mr. Devine:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800).638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



 Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

K102988 1/1

**PANGAO<sup>®</sup>**

Premarket Notification  
Traditional Sec. 510(k) Submission  
PG-800B Electronic Blood Pressure Monitor  
Sec. II Indication for Use Statement

**Sec. II Indication for Use Statement**

510(k) Number: K102988

Device Name: **PG-800B Electronic Blood Pressure Monitor**

NOV - 9 2010

**Indications for Use:**

PG-800B Electronic Blood Pressure Monitor is intended to measure the systolic and diastolic blood pressure as well as the pulse rate of adult person via non-invasive oscillometric technique in which an inflatable cuff is wrapped around the upper arm.

It can be used at medical facilities or at home.

The intended arm circumference is 22-32 cm.

Prescription Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use   X    
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

*Danna B. Kachner*  
(Division Sign-Off)  
Division of Cardiovascular Devices

Page   1   of   1  

510(k) Number K102988