

510(k) Summary

MAR - 3 2011

Date Prepared: October 20, 2010  
 Submitter: Nanosonics Ltd  
 Unit 24 566, Gardeners Rd  
 Alexandria, NSW, 2015  
 Australia  
 Ph: +61 2 8063 1600  
 Fax: +61 2 9317 5010

Contact Information: Mr. Ron Weinberger  
 General Manager – Innovation and Technology  
 Nanosonics Ltd  
 Unit 24, 566 Gardeners Road  
 Alexandria, NSW 2015  
 Ph: +61 2 8063 1600  
 Fax: +61 2 9317 5010

Device Trade Name: Trophon Chemical Indicator  
 Common or Usual Name: Chemical Indicator  
 Device Classification Name: Chemical Indicator for Liquid Chemical Germicide  
 Physical/Chemical Sterilization Process Indicator  
 Product Code: JOJ  
 Class: II  
 Regulation Number: 21 CFR 880.2800  
 Substantial Equivalence: The Trophon Chemical Indicator is substantially  
 equivalent to Serim D-CIDE GTA 1.5% Test Strip.  
 (K092346)

Device Description: The device is a qualitative, single use, disc (26 mm in  
 diameter) that has a hydrogen peroxide sensitive  
 chemical indicating ink applied to a Tyvek substrate.  
 The hydrogen peroxide sensitive chemical indicating  
 ink has been designed to transition from an initial red  
 color to a signal yellow color when subjected to  
 sources of hydrogen peroxide.

Intended Use: The Trophon Chemical Indicator is intended by  
 Nanosonics for use by health care providers for  
 confirming that the disinfectant delivered into the  
 Chamber of the Nanosonics Trophon EPR disinfectant  
 device is above the minimum effective concentration  
 (MEC) required to achieve the stated performance of  
 the Trophon EPR.

The color of the Tropon Chemical Indicator changes from red to yellow when exposed to hydrogen peroxide, the active ingredient in Tropon EPR-C40 disinfectant. This color change occurs above the minimum effective concentration (MEC) established for this solution.

K103126  
p2/2

**Conclusion:**

The Tropon Chemical Indicator and the predicate device are both single use indicators used to monitor the active ingredient/critical process parameters in a specific solution/process. There are no significant technological risks and no new risks are presented by the design of the Tropon Chemical Indicator.

Product Comparison		
Technological Characteristics		
	Predicate Device Serim D-CIDE GTA 1.5% Test Strip. (K092346)	Subject Device Tropon Chemical Indicator
<b>Description</b>	Chemical indicator based on proprietary ink technology which changes color on exposure to sterilant	Chemical indicator based on proprietary ink technology which changes color on exposure to sterilant
<b>Usage</b>	As indicator to confirm that glutaraldehyde exceeds the validated minimum effective concentration.	As indicator to confirm that hydrogen peroxide exceeds the validated minimum effective concentration.
<b>Indicator Technology</b>	Reacting Chemicals and background dye	Reacting Chemicals and background dye
<b>Nature of Indication</b>	Qualitative – color change	Qualitative – color change
<b>Single Use</b>	Single Use disposable	Single Use disposable
<b>Technical Characteristics</b>	Glutaraldehyde reacts with sodium sulfite to form a colorless addition product and a base. The base then reacts with a pH indicator producing a purple color. The indicator pad also contains sodium bisulfite, which reacts with both glutaraldehyde and with the base.	Hydrogen peroxide reacts with components within the proprietary ink formulation to cause a pH shift, resulting in a color change from an initial red color to a yellow signal color.
<b>User</b>	Trained Health Care Professionals	Trained Health Care Professionals



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room - WO66-G609  
Silver Spring, MD 20993-0002

Mr. Ron Weinberger  
General Manager-Innovation and Technology  
Nanosonics, Limited  
Unit 24 566, Gardeners Road  
Alexandria, NSW, 2015  
Australia

MAR - 3 2011

Re: K103126  
Trade/Device Name: Trophon Chemical Indicator  
Regulation Number: 21 CFR 880.2800  
Regulation Name: Sterilization Process Indicator  
Regulatory Class: II  
Product Code: JOJ  
Dated: February 16, 2011  
Received: February 18, 2011

Dear Mr. Weinberger:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2- Mr. Weinberger


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to

<http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.  
Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

K103126

510(k) Number (if known):

Device Name: Trophon Chemical Indicator

Indications for Use:

The Trophon Chemical Indicator is used exclusively for monitoring the High Level Disinfection process when placed within the Trophon EPR chamber.

The color of the Trophon Chemical Indicator changes from red to yellow when exposed to hydrogen peroxide, the active ingredient in Trophon EPR-C40 disinfectant. This occurs above the minimum effective concentration (MEC) established for this solution.


Prescription Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use  X   
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

  
\_\_\_\_\_  
(Division Sign-Off)  
Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices  
510(k) Number:  K103126