

510(k) SUMMARY

MAR 25 2011

Lanx Spinal Fixation System**Submitter Information**

Lanx, Inc.
390 Interlocken Crescent, Suite 890
Broomfield, CO 80021
(303) 443-7500
Contact Person: Damon Belloni
Date Prepared: March 4, 2011

Device Identification

Proprietary Name: Lanx Spinal Fixation System
Common Name: Lanx SFS
Classification: Pedicle Screw Spinal System (MNH, MNI, NKB), 21 CFR
888.3070 and/or Spinal Interlaminar Fixation Orthosis (KWP),
21 CFR 888.3050
Device Class: Class II and Class III

Predicate Device Information

Lanx Spinal Fixation System K043484, K071633, K071877, K083581, K0905252, K092536,
K093285, K100935

Intended Use / Indications for Use

The Lanx Spinal Fixation System (SFS) is intended to be used to help provide immobilization and stabilization of spinal segments as an adjunct to fusion of the thoracic, lumbar and/or sacral spine. The System is intended for use with autograft or allograft.

The Lanx Spinal Fixation System is intended for posterior, non-cervical (T1-S2/ilium) pedicle and non-pedicle spinal fixation, to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following instabilities or deformities: degenerative disc disease (DDD, defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e. fracture or dislocation); spinal stenosis; deformities or curvatures (i.e. scoliosis, kyphosis and/or lordosis); tumor; pseudoarthrosis; and failed previous fusion.

The Lanx Spinous Process Fusion Plate (SPFP) is a posterior, non-pedicle supplemental fixation device, intended for use at a single-level in the non-cervical spine (T1-S1). It is intended for

plate fixation/attachment to spinous processes for the purpose of achieving supplemental fusion in the following conditions: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); and/or tumor. The Lanx Spinous Process Fusion Plate is intended for use with bone graft material, not intended for standalone use.

Device Description & Technological Characteristics

The purpose of this Special 510(k) is to add a modified polyaxial pedicle screw configuration in multiple sizes. The modified system has the same intended use and fundamental scientific technology as the previously-cleared system.

The Lanx Spinal Fixation System consists of various screws, hooks, rods, plates, connectors, etc. that are used to build a construct to provide supplemental stabilization of spinal segments to support fusion. The system components can be assembled in a variety of configurations, allowing the surgeon to tailor the construct to the particular needs of the patient.

The Lanx Spinal Fixation System implants are fabricated from medical grade titanium, titanium alloy and/or cobalt chrome alloy per ASTM F67, ASTM F136 and ASTM F1537. Titanium and cobalt chrome components may be used together within the same construct. These components should never be used with stainless steel implant components.

Performance Data

Performance testing was conducted to characterize the modified components of the system. Tests were performed on the modified and predicate systems in accordance with ASTM F1717 (static and dynamic axial compression, static torsion) and ASTM F1798 (static flexion-extension). The Lanx Spinal Fixation System functioned as intended and the observed test results demonstrate substantial equivalence to the predicate device.

Substantial Equivalence

The modified components of the Lanx Spinal Fixation System have the same intended use and indications and similar technological characteristics and principles of operation as the predicate system. The minor differences in the new components raise no new issues of safety or effectiveness. Mechanical testing also demonstrated comparable mechanical properties to the previously cleared Lanx Spinal Fixation System. Thus, the modified Lanx Spinal Fixation System is substantially equivalent to the predicate device.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

Lanx, Inc.
% Mr. Damon Belloni
390 Interlocken Crescent, Suite 890
Broomfield, Colorado 80021

MAR 25 2011

Re: K103237
Trade/Device Name: Lanx Spinal Fixation System
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class III
Product Code: NKB, MNI, MNH, KWP
Dated: March 03, 2011
Received: March 07, 2011

Dear Mr. Belloni:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must


Page 2 – Mr. Damon Belloni

comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Mark N. Melkerson" with a stylized flourish below it.

Mark N. Melkerson
Director
Division of Surgical, Orthopedic
And Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510(k) Number (if known): K103237

Device Name: Lanx Spinal Fixation System

Indications for Use:

The Lanx Spinal Fixation System (SFS) is intended to be used to help provide immobilization and stabilization of spinal segments as an adjunct to fusion of the thoracic, lumbar and/or sacral spine. The System is intended for use with autograft or allograft.

The Lanx Spinal Fixation System is intended for posterior, non-cervical (T1-S2/ilium) pedicle and non-pedicle spinal fixation, to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of the following instabilities or deformities: degenerative disc disease (DDD, defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e. fracture or dislocation); spinal stenosis; deformities or curvatures (i.e. scoliosis, kyphosis and/or lordosis); tumor; pseudoarthrosis; and failed previous fusion.

The Lanx Spinous Process Fusion Plate (SPFP) is a posterior, non-pedicle supplemental fixation device, intended for use at a single-level in the non-cervical spine (T1-S1). It is intended for plate fixation/attachment to spinous processes for the purpose of achieving supplemental fusion in the following conditions: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); and/or tumor. The Lanx Spinous Process Fusion Plate is intended for use with bone graft material, not intended for standalone use.


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

Page ___ of ___

510(k) Number K103237 15