



DEC 1 2010

510(k) Summary

Date Prepared October 1, 2010

Submitter Medtronic, Inc.
Medtronic Perfusion Systems
7611 Northland Drive
Minneapolis, MN 55428
Establish Registration Number: 2184009

Contact Person Jeffrey L. Koll
Senior Regulatory Affairs Specialist
Phone: (763) 514-9842
Fax: (763) 367-8360
Email: jeffrey.l.koll@medtronic.com

Device Name and Classification

Trade Name: Heparin Assay Controls
Common Name: Plasma, Coagulation Control
Regulation Number: 21 CFR 864.5425
Product Code: GGN
Classification: Class II

Predicate Device

Medtronic Heparin Assay Controls (K894317 and K042206) cleared on September 5, 1989 and October 1, 2004, respectively.

Comparison to Predicate Device

A comparison of the modified device and the currently marketed Heparin Assay Controls shows the following similarities:

- Same intended use.
- Same operating principle.
- Same technological characteristics.
- Same performance claims.

Alleviating Pain · Restoring Health · Extending Life

Description of Device Modification

Heparin

- The modification to the current control is to replace old USP heparin with revised USP heparin from the new heparin monograph that became effective October 1, 2009. The source of the heparin remains porcine.

Intended Use

The intended use is unchanged.

Labeling

The current labels are unchanged by this device modification. The current Instructions for Use (IFU) has been updated for style, grammar and readability. Applicable tables have also been updated with information on the construction of the modified controls using revised USP heparin.

- Appendix A contains the current IFU.
- Appendix B contains the draft IFU.
- Appendix C contains the current labels

Conclusion

The modifications to the Medtronic Heparin Assay Controls described in this submission result in a substantially equivalent device because verification testing indicated that the fundamental scientific technology and the intended use are unchanged.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center – WO66-0609
Silver Spring, MD 20993-0002

Medtronic, Inc.
c/o Mr. Jeffrey L. Koll
Senior Regulatory Affairs Specialist
8200 Coral Sea Street NE
Mounds View, MN 55112

DEC 01 2010

Re: k103313

Trade/Device Name: Heparin Assay Controls
Regulation Number: 21 CFR 864.5425
Regulation Name: Multipurpose system for in vitro coagulation studies
Regulatory Class: Class II
Product Code: GGN
Dated: October 1, 2010
Received: November 10, 2010

Dear Mr. Koll:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter


Page 2 – Mr. Jeffrey L. Koll

will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



for

Maria M. Chan, Ph.D

Director

Division of Immunology and Hematology Devices

Office of *In Vitro* Diagnostic Device Evaluation and Safety

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K103313

Device Name: Heparin Assay Controls

DEC 1 2010

Indications for Use:

To verify the performance of the HMS Plus and Heparin Assay Cartridges.

For In Vitro Diagnostic use.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



Division Sign-Off

Office of In Vitro Diagnostic
Device Evaluation and Safety

510(k) K103313