

Venofer Pump Version 1.1  
Special 510(k)

FEB 10 2011

**Section 5: 510(k) Summary**

---

**510(k) Summary**

***Submitter's Information***

<b>Name</b>	Renal Solutions, Inc.
<b>Address</b>	770 Commonwealth Drive Warrendale, Pa 15086 Phone: (724) 720-2840 FAX: (724) 772-6925
<b>Official Contact</b>	David J. Vanella Senior Vice President, Quality Systems
<b>Date Prepared</b>	January 7, 2011

***Device Information***

<b>Name</b>	Venofer Pump, for use on the Fresenius 2008T Hemodialysis Machine
<b>Common/Usual Name</b>	Hemodialysis System
<b>Product Code</b>	KDI
<b>Classification Name</b>	Dialyzer, High Permeability With Or Without Sealed Dialysate System
<b>Regulation Number</b>	876.5860
<b>Proprietary Name</b>	Venofer Pump
<b>Unmodified Device</b>	Venofer Pump (K093964)
<b>Reason for Submission</b>	Modification to existing device

***Substantial Equivalence***

The Venofer Pump 1.1 is substantially equivalent to the unmodified device (Venofer Pump K093964) in terms its intended use, environment of use, operating principles, and technology.

**Venofer Pump Version 1.1  
Special 510(k)**

**Section 5: 510(k) Summary**

***Intended Use***

The intended use of the Venofer Pump 1.1 is identical to the unmodified device.

The Venofer® Pump is an optional accessory for use on Fresenius 2008T Hemodialysis Machines and is intended to administer Venofer®, an iron sucrose, to treat iron deficiency anemia in patients with chronic kidney disease undergoing hemodialysis, where Venofer®, in conjunction with hemodialysis, is prescribed by a physician.

***Venofer Pump 510(k) Verification Testing Non-Clinical Summary***

The verification (non-clinical) testing information consists of performance, safety, and software testing that was performed to verify the Venofer Pump meets its performance specifications and to demonstrate the device is substantially equivalent to the established predicate devices.

The following table summarizes the 510(k) verification testing activities performed. These include performance, safety and software testing, which demonstrates by technical examination that the Venofer Pump meets its performance specifications, the designated (FDA Consensus) standard requirements, and the software design input requirements.

510(k) Verification Testing	510(K) Verification Testing Activities
<p>Venofer Pump Performance Testing</p> <p><i>The performance (verification) testing results support the performance characteristics of the Venofer Pump.</i></p>	<p>Functional testing that demonstrates that the device performs as designed and expected, includes the following:</p> <ul style="list-style-type: none"> <li>• The specific verification tests conducted</li> <li>• A description of all test protocol including:               <ul style="list-style-type: none"> <li>• objective of the tests</li> <li>• test articles used in the tests</li> <li>• test methods and procedures</li> <li>• pre-defined acceptance or pass/fail criteria.</li> </ul> </li> <li>• System-level hazard analysis that confirms that the device does not perform in an unexpected and/or unsafe manner</li> </ul>

Venofor Pump Version 1.1  
Special 510(k)

Section 5: 510(k) Summary

510(k) Verification Testing	510(K) Verification Testing Activities
<p>Venofor Pump Safety Testing</p> <p><i>The verification testing results support the safety characteristics of the Venofor Pump.</i></p> <p><i>Biocompatibility testing was performed on all new materials that are patient-fluid contacting</i></p>	<p>Product safety testing that demonstrates that the device performs per the FDA Consensus Standards, as identified below:</p> <ul style="list-style-type: none"> <li>• Electromagnetic compatibility (EMC) testing (IEC 60601-1-2: 2007 – Medical Electrical Equipment – Part 1-2: General Requirements for Safety – Collateral Standard: Electromagnetic Compatibility – Requirements for Testing</li> <li>• Biocompatibility Testing (AAMI / ANSI / ISO 10993-1:2009 Biological evaluation of medical devices Part 1: Evaluation &amp; testing)</li> </ul>
<p>Venofor Pump Software Testing</p> <p><i>The testing includes the required documentation as described in the guidance titled Guidance for the Content of Premarket Submissions for Software</i></p>	<p>Software testing that demonstrates the device software meets the design input requirements. The device was tested and includes the following</p> <ul style="list-style-type: none"> <li>• Structure chart of flow chart describing software architecture</li> <li>• Summary of software development procedures, including changes made to the software</li> <li>• Software requirements specification with traceability back to the hazard analysis</li> <li>• Verification and validation test plans, including pass/fail criteria and traceability back to the requirements;</li> <li>• System level test results</li> <li>• Current software version number and date of latest revision.</li> </ul>

**Venofer Pump Version 1.1  
Special 510(k)**

**Section 5: 510(k) Summary**

---

The conclusions drawn from this testing demonstrates that the Venofer Pump is as safe, as effective, and performs at least as safely and effectively as the legally marketed devices identified as predicate devices to which it was compared.

***General Safety and Effectiveness***

The Venofer Pump 1.1 is an updated version of the Venofer Pump (K093964). The performance and technological characteristics of the modified device are equivalent to those of the unmodified device and raise no new types of safety or effectiveness questions.

***Device Description***

The Venofer Pump Module is an optional module for use on Fresenius 2008T Hemodialysis Machines. The module is designed to administer Venofer, an iron sucrose supplement, during dialysis treatments. The Venofer Pump Module is to be used in accordance with the approved Venofer Indications for Use and the physician's prescription.

The Venofer Pump Module fits into the module compartment of Fresenius 2008T Machines and consists of a control panel, vial holder, fluid detector, and a peristaltic pump.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Mail Center - WO66-G609  
Silver Spring, MD 20993-0002

Mr. David J. Vanella  
Senior Vice President, Quality Systems  
Renal Solutions<sup>®</sup>, Inc.  
770 Commonwealth Drive, Suite 101  
WARRENDALE PA 15086

FEB 10 2011

Re: K103564  
Trade/Device Name: Venofer Pump  
Regulation Number: 21 CFR §876.5860  
Regulation Name: High permeability hemodialysis system  
Regulatory Class: II  
Product Code: KDI  
Dated: January 7, 2011  
Received: January 11, 2011

Dear Mr. Vanella:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related

adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Herbert P. Lerner, M.D., Director (Acting)  
Division of Reproductive, Gastro-Renal  
and Urological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health.

Enclosure

K103564

Venofer Pump Version 1.1  
Special 510(k)

**Section 5: Indications for Use Statement**

---

**Indications for Use Statement**

510(k) Number (if known): N/A

**Device Name:** Venofer Pump


**Indications for Use:**

Fresenius 2008T is indicated for acute and chronic dialysis therapy.

Prescription Use   X   AND/OR Over-The-Counter Use         
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE  
-CONTINUE ON ANOTHER PAGE IF NEEDED)

---



Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)  
Division of Reproductive, Gastro-Renal, and  
Urological Devices  
510(k) Number   K103564