

K110216

APR - 1 2011

510(K) SUMMARY

Submitter: DePuy Spine, Inc.
325 Paramount Drive
Raynham, MA 02767

Contact Person: Eugene Bang
Regulatory Affairs Associate
Voice: (508) 977-3966
Fax: (508) 828-3797
Email: ebang@its.jnj.com

Date Prepared: January 24, 2011

Trade Name: EXPEDIUM® and VIPER® Systems

Device Class: Class III

Product Code(s): NKB, KWQ, KWP, MNH, MNI

Common Name: Appliance, Fixation, Spinal Interlaminar;
Appliance, Fixation, Spinal Intervertebral Body;
Orthosis, Spondyloisthesis Spinal Fixation;
Orthosis, Spinal Pedicle Fixation;
Orthosis, Spinal Pedicle Fixation, For Degenerative Disc Disease

Classification Name: Spinal interlaminar fixation orthosis
per 21 CFR §888.3050
Spinal intervertebral body fixation orthosis
per 21 CFR §888.3060
Pedicle screw spinal fixation
per 21 CFR §888.3070

Predicate Devices: EXPEDIUM® System – K033901, K051024, K062174, K063156, K063741, K080313, K081898

VIPER® System – K033901, K090648, K101993

Device Description: The proposed cortical fix polyaxial screws for EXPEDIUM and VIPER System are available in various geometries and sizes.

Intended Use: The EXPEDIUM and VIPER Systems are intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of acute and chronic instabilities or deformities of the thoracic, lumbar and sacral spine.

The EXPEDIUM and VIPER Systems are intended for noncervical pedicle fixation and nonpedicle fixation for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; pseudoarthrosis; and failed previous fusion in skeletally mature patients.

When used in a percutaneous approach with MIS Instrumentation, the VIPER Systems are intended for noncervical pedicle fixation and nonpedicle fixation for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; pseudoarthrosis; and failed previous fusion in skeletally mature patients.

Materials: Manufactured from ASTM F138 implant grade stainless steel and ASTM F136 implant grade titanium alloy.

Comparison to

Predicate Device: The substantial equivalence of the subject device to the predicates identified above is based upon the equivalence of intended use, design (fundamental scientific technology), materials, manufacturing methods, performance, sterility, biocompatibility, safety and packaging design.

Non-clinical Test**Summary:**

The following mechanical tests were conducted:

- Static cantilever beam testing in accordance with ASTM F 1798-97 *Standard Guide for Evaluating the Static and Fatigue Properties of Interconnection Mechanisms and Subassemblies Used in Spinal Arthrodesis Implants*. The acceptance criteria was/were met.
- Static axial slip testing in accordance with ASTM F 1798-97 *Standard Guide for Evaluating the Static and Fatigue Properties of Interconnection Mechanisms and Subassemblies Used in Spinal Arthrodesis Implants*. The acceptance criteria was/were met.
- Dynamic cantilever beam testing in accordance with ASTM F 1798-97 *Standard Guide for Evaluating the Static and Fatigue Properties of Interconnection Mechanisms and Subassemblies Used in Spinal Arthrodesis Implants*. The acceptance criteria was/were met.

Clinical Test**Summary:**

No clinical tests were performed.

Conclusion:

Based on the predicate comparison and testing, the subject additions to the EXPEDIUM and VIPER Systems are substantially equivalent to the predicate devices.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room - WO66-G609
Silver Spring, MD 20993-0002

APR - 1 2011

DePuy Spine, Inc.
% Mr. Eugene Bang
Regulatory Affairs Associate
325 Paramount Drive
Raynham, Massachusetts 02767

Re: K110216
Trade/Device Name: EXPEDIUM® and VIPER® Systems
Regulation Number: 21 CFR 888.3070
Regulation Name: Pedicle screw spinal system
Regulatory Class: Class III
Product Code: NKB, MNH, MNI, KWP, KWQ
Dated: January 24, 2011
Received: October 12, 2010

Dear Mr. Bang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act


Page 2 – Mr. Eugene Bang

or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Mark N. Melkerson" with a stylized flourish at the end.

Mark N. Melkerson
Director
Division of Surgical, Orthopedic
And Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

INDICATIONS FOR USE STATEMENT

510(k) Number (if known): K110216

Device Name: EXPEDIUM® and VIPER® Systems

Indications For Use:

The EXPEDIUM and VIPER Systems are intended to provide immobilization and stabilization of spinal segments in skeletally mature patients as an adjunct to fusion in the treatment of acute and chronic instabilities or deformities of the thoracic, lumbar and sacral spine.

The EXPEDIUM and VIPER Systems are intended for noncervical pedicle fixation and nonpedicle fixation for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; pseudoarthrosis; and failed previous fusion in skeletally mature patients.

When used in a percutaneous approach with MIS Instrumentation, the VIPER Systems are intended for noncervical pedicle fixation and nonpedicle fixation for the following indications: degenerative disc disease (defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies); spondylolisthesis; trauma (i.e., fracture or dislocation); spinal stenosis; curvatures (i.e., scoliosis, kyphosis, and/or lordosis); tumor; pseudoarthrosis; and failed previous fusion in skeletally mature patients.

Prescription Use X

AND/OR


Over-The-Counter Use

(Part 21 CFR 801 Subpart D)

(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Signature)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K110216