

K110606

SEP 29 2011

**510 (k) Summary of safety and effectiveness****SUBMITTER INFORMATION**

- A. Company Name: M.V. s.r.l.
- B. Company Address: Via F.lli Cervi, 7  
Gonzaga (MN), Italy 46023  
Manufacturing facility: Via Don G. Dossetti, 5/7  
Gonzaga (MN), Italy 46023  
  
US Agent: Laura Andreatta  
N. Devon Rd., Columbus, OH 43212,  
Phone +1-614-246-1141  
Fax: +1-614-725-0890
- C. Company Phone: +39 0376 536995  
Company Fax: +39 0376 530441  
Company e-mail: info@mvsrl.it
- D. Contact Person: Enrico Bisson  
Studio ingegneria Enrico Bisson
- E. Contact Phone: +39 0498630080  
Contact Fax: +39 0498630080  
Contact E-mail: enrico.bisson@isoplan.org
- F. Date Summary Prepared: February 28, 2011

**DEVICE IDENTIFICATION**

- A. Device name: M.V. Intradermic Needles
- B. Trade/Proprietary Name: M.V. Intradermic Needles
- C. Classification name: Hypodermic Single Lumen Needle (21 CFR §880.5570)
- D. Product Code: FMI

**LEGALLY MARKETED DEVICES (PREDICATE DEVICES)**

- ARTSANA HYPODERMIC NEEDLES, K051783

**DEVICE DESCRIPTION**

The device is comprised of a metal tube. On one end the tip is closed and blunt, while the cannula has an opening laterally at a lower point under the tip. On the other end the device is joined to a female connector (hub) designed to mate with a male connector (nozzle) of a piston syringe.

The MV intradermic needle is marketed alone or combined with a pilot needle. The pilot needle is an hypodermic single lumen needle intended to prepare the site for injection. The tip of the pilot needle is sharpened at one end, while the other end is joined to a hub. The pilot needle can be any traditional legally marketed hypodermic needle.

The MV intradermic needles are single-use devices sold as sterile.

**INTENDED USE**

The MV intradermic needles are intended to inject fluids intradermally.

**DISCUSSION OF NON CLINICAL TESTS**

Following performance tests were performed on the MV Intradermic Needles to test their compliance with Recognized Consensus Standard ISO 7864 Sterile hypodermic needles for single use:

- Cleanness
- Acidity or alkalinity limits
- Limits for extractable metals
- Needle tube – Length requirements
- Needle tube - Absence of defects
- Execution - Hub/tube bond strength
- Execution - Evidence of lumen

None of the data raised any issues of safety and effectiveness. Additionally, a risk analysis was conducted according to Recognized Consensus Standard ISO 14971:2007.

Following biological tests were performed on the sterile final finished device to evaluate biocompatibility:

- Cytotoxicity,
- Sensitization,
- Intracutaneous reactivity,
- Systemic toxicity
- Haemocompatibility.

None of the result of the tests arised any issues of biocompatibility.

**SUBSTANTIAL EQUIVALENCE**

The MV Intradermic Needles are same or similar in design, materials and intended use to the predicate devices. In further support of a substantial equivalence determination, Section 10 provides a comparison chart of the submitted device and the predicate devices. Main comparison element are as follows:

	<b>MV INTRADERMIC NEEDLES</b>	<b>K051783</b>
<b>Intended / Indications For Use</b>	The MN intradermic needles are intended to inject fluids intradermally.	To inject fluids into, or withdraw fluids from, parts of the body below the surface of the skin.
<b>Cannula material</b>	AISI 304 Stainless Steel	AISI 304 Stainless Steel
<b>Hub material</b>	Polypropylene	Polypropylene
<b>Needle diameter</b>	22G, 25G, 27G, 30G; 21G, 23G, 26G, 27G	21G, 22G, 23G, 25G, 26G, 27G, 30G and more
<b>Needle length (mm)</b>	25, 27, 35, 37, 40, 50, 57, 70; 13, 25	13, 25, 40 and more
<b>Tip configuration</b>	closed blunt tip, lateral opening (intradermic needle), triple sharpened, non-coring (pilot needle)	triple sharpened, non-coring
<b>Hub</b>	color coded ISO 6009	color coded ISO 6009
<b>Connection to syringe</b>	Luer taper	Luer taper

Based on the available information, we conclude that the M.V. Intradermic Needles are substantially equivalent to the existing legally marketed devices under Federal Food, Drug and Cosmetic Act. Therefore, the applicant device is determined as safe and effective.

#### **SAMPLE**

Sample of the MV Intradermic Needles is included in this submission.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room - WO66-G609  
Silver Spring, MD 20993-0002

M.V. S.R.L.  
C/O Mr. Enrico Bisson  
President  
Studio di Ingegneria Enrico Bisson  
Via Marzia 9  
Abano Terme, Padova Italy 35031

SEP 29 2011

Re: K110606  
Trade/Device Name: MV Intradermic Needles  
Regulation Number: 21 CFR 880.5570  
Regulation Name: Hypodermic Single Lumen Needle  
Regulatory Class: II  
Product Code: FMI  
Dated: September 9, 2011  
Received: September 16, 2011

Dear Mr. Bisson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.

Director

Division of Anesthesiology, General Hospital

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

**INDICATIONS FOR USE**510(k) Number (if known): K110606

Device Name: MV intradermic needles

## Indications for Use:


*The MV intradermic needles are intended to inject fluids intradermally.*Prescription Use   X    
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

  
(Division Sign-Off)Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices510(k) Number: K110606