

**510(k) Summary**  
Page 1 of 2  
15-Mar-11

JUN - 8 2011

EasyLap Ltd.  
30 Ha'Ella St.  
Kfar Truman, 73150  
Israel

Tel - 011-972-528565644  
Fax – 011-972-776201003

**Official Contact:** Einat Duvdevany - CEO

**Proprietary or Trade Name:** iMESH Tacker

**Common/Usual Name:** Implantable staple

**Classification Name/Code:** GDW – Implantable staple  
CFR 878.4750

**Device:** iMESH Tacker

**Predicate Devices:** K090470 – Covidien – Protack  
K071061 – Sorbx (Covidien) – Absorbatack

**Device Description:**

The iMESH Tacker is a sterile, single use device for the fixation of prosthetic material, such as a hernia mesh, onto soft tissue. The applicator features an articulating tip. The absorbable tack is made of synthetic polyester derived from a lactic acid and glycolic acid copolymer. The Applicator is pre-loaded with 30 tacks.

**Indications for Use:**

The iMESH Tacker is indicated for fixation of prosthetic material to soft tissue in various minimally invasive and open general surgical procedures, such as hernia repair.

**Patient population:**

Individuals undergoing procedures where prosthetic mesh is being used.

**Environment of Use:**

Hospitals, sub-acute care institutions, and surgery centers.

**510(k) Summary**

Page 2 of 3

15-Mar-11

**Comparison to Predicates**

	iMESH Tacker	Covidien Absorbatack K071061	Covidien Protack K090470
Product Classification	GDW	GDW	GDW
CFR	874.4750	874.4750	874.4750
Indications for Use	For fixation of prosthetic material to soft tissue in various minimally invasive and open general surgical procedures, such as hernia repair	For fixation of prosthetic material to soft tissue in various minimally invasive and open general surgical procedures, such as hernia repair	For use in affixing prosthetic material or approximately tissue. May be used both endoscopically and in open procedures
Environments of Use	Hospitals, sub-acute care institutions, and surgery centers.	Hospitals, sub-acute care institutions, and surgery centers.	Hospitals, sub-acute care institutions, and surgery centers.
Patient Population	Individuals undergoing procedures where prosthetic mesh is being used.	Individuals undergoing procedures where prosthetic mesh is being used.	Individuals undergoing procedures where prosthetic mesh is being used.
Delivery Device Design	Handles with triggers	Handles with triggers	Handles with triggers
Tip design	Articulating tip	Non-articulating tip	Non-articulating tip
Tacks pre-loaded in tip	Yes	Yes	Yes
Material of tip	Stainless steel	Stainless steel	Stainless steel
# of Tack pre-loaded	30	10 and 20	30
Tack Design	Helical	Screw	Helical
Tack shape	Bio-Absorbable	Bio-Absorbable	Non-bio-Absorbable
Tack Length	6.3 mm	5 mm	3.8 mm
Material of Tack	PLA/PGA	PLA/PGA	Titanium
Packaging	Sterile	Sterile	Sterile
Single patient use, disposable	Yes	Yes	Yes
<b>Performance Testing</b>			
Fixation Force	On implantation 17.07 N	On implantation 16.46 N	Not tested
	After 14 days 18.82 N	After 14 days 8.35 N	Not Tested
Fixation Force Mid-section average, upper and lower sections average	17.06  17.08	11.50  22.113	Not tested

---

**510(k) Summary**

Page 3 of 3

15-Mar-11

**Summary of substantial equivalence:**

We have demonstrated that the iMESH Tacker is equivalent to the predicates in design and performance characteristics:

The iMESH Tacker is viewed as substantially equivalent to the predicate devices because:

**Indications –**

- Identical indications for use - indicated for fixation of prosthetic material to soft tissue in various minimally invasive and open general surgical procedures, such as hernia repair to predicate – K071061 – Absorbatack

**Technology –**

- Similar technology which incorporates, an in-line handle, a tip and pre-loaded tacks – K071061 – Absorbatack
- Similar tack design – helical – K090470 - Protack

**Materials –**

- Identical materials to predicate – K071061 – Absorbatack

**Environment of Use –**

- Identical to predicate – K071061 – Absorbatack

**Performance specifications –**

- Degradation and consistency of fixation force were tested and found to be similar to predicate – K071061 – Absorbatack

**Differences –**

The only difference between the proposed iMESH Tacker and the predicates is:

- Helical, bio-absorbable tack and
- Deployment device, the iMESGH Tacker, has an articulating tip to help facilitate tack placement.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room -WO66-G609  
Silver Spring, MD 20993-0002

EasyLap Ltd.  
% PreMedic, Inc.  
Mr. Paul Dryden  
24301 Woodsage Drive  
Bonita Springs, Florida 34134

JUN - 8 2011

Re: K110728  
Trade/Device Name: iMESH Tacker  
Regulation Number: 21 CFR 878.4750  
Regulation Name: Implantable staple  
Regulatory Class: II  
Product Code: GDW  
Dated: May 29, 2011  
Received: June 1, 2011

Dear Mr. Dryden:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21


Page 2 - Mr. Paul Dryden

CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink, appearing to read 'Mark N. Melkerson', with a date '6/8/11' written below it.

Mark N. Melkerson  
Director  
Division of Surgical, Orthopedic  
and Restorative Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**Indications for Use Statement**

Page 1 of 1

510(k) Number: K110728 (To be assigned)

Device Name: iMESH Tacker

**Indications for Use:**

The iMESH Tacker is indicated for fixation of prosthetic material to soft tissue in various minimally invasive and open general surgical procedures, such as hernia repair.

Prescription Use **XX**  
(Part 21 CFR 801 Subpart D)

or

Over-the-counter use      
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

Division of Surgical, Orthopedic,  
and Restorative Devices

510(k) Number K110728