



K111447(1/2)

OCT - 5 2011

P.O. Box 708
Warsaw, IN 46581-0708
574 267-6131

510(k) Summary

Sponsor: Zimmer, Inc.
P.O. Box 708
Warsaw, IN 46581-0708

Contact Person: Stephen H. McKelvey
Senior Project Manager, Trauma Regulatory Affairs
Telephone: (574) 372-4944
Fax: (574) 371-8760

Date: August 22, 2011

Trade Name: *Zimmer*[®] Periarticular Plating System - Screws

Common Name: Temporary Internal Fixation Devices

Classification Names and References: Smooth or threaded metallic bone fixation fastener - 21 CFR 888.3040

Predicate Devices: *Zimmer* Periarticular Locking Plate System

Device Description: Temporary internal fixation devices are designed to stabilize fractures during the normal healing process.

Intended Use: Temporary internal fixation devices are designed to stabilize fractures during the normal healing process.

Comparison to Predicate Device: The *Zimmer*[®] Periarticular Plating System - Screws are similar in intended use, type of materials, and performance characteristics to the predicate devices (*Zimmer* Periarticular Locking Plate System - K042598, cleared 10/29/2004). The proposed screws are provided sterile vs. non-sterile.

Performance Data (Nonclinical and/or Clinical): Non-Clinical Performance and Conclusions:

- **Sterilization Validation** - To demonstrate that at a minimum gamma dose of 20kGy the devices can be terminally sterilized to a SAL greater than or equal to 10^{-6} .

- **Shelf Life** - Accelerated aging showed that the product has a shelf life of 10 years.
- **Sterile Packaging** - To withstand normal distribution and storage conditions and maintain the sterile barrier properties throughout the specified product shelf life.
- **Biocompatibility** - Biocompatibility testing on the screw material was conducted per ISO 10993-1 and Good Laboratory Practices (21 CFR § 58). All testing passed.

Providing these screws pre-sterilized did not change the intended use or the fundamental scientific technology of any of the devices. Each sterile device uses the same operating principle and incorporates the same basic labeling.

The results of either engineering evaluations and/or non-clinical (lab) performance testing demonstrate that the devices are safe and effective and substantially equivalent to the predicate devices.

Engineering evaluations included any differences between screw diameters, partial thread vs. full threads, starting load, bending or fatigue failure, material strength and the elimination of locking threads on the drive heads. Screw testing/analysis performed included: cross-sectional analysis, fatigue failure, insertion torque and torque to failure.

In summary, the sterile devices described in this submission are substantially equivalent to their predicates.

Clinical Performance and Conclusions:

Clinical data and conclusions were not needed for this device.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -W066-G609
Silver Spring, MD 20993-0002

Zimmer, Inc.
% Mr. Stephen McKelvey
Senior Project Manager, Trauma Regulatory Affairs
P.O. Box 708
Warsaw, Indiana 46581-0708

OCT - 5 2011

Re: K111447

Trade/Device Name: Zimmer® Periarticular Plating System - Screws
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: HWC
Dated: August 22, 2011
Received: August 23, 2011

Dear Mr. McKelvey:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,


Mark N. Melkersen
Director
Division of Surgical, Orthopedic,
and Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K111447

Device Name:

Zimmer® Periarticular Plating System - Screws

Indications for Use:

Temporary internal fixation devices are designed to stabilize fractures during the normal healing process.


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(Please do not write below this line – Continue on another page if needed)

Concurrence of CDRH, Office of Device Evaluation (ODE)


(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K111447