



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room - WO66-G609
Silver Spring, MD 20993-0002

Kunshan Jiehong NonWoven Product Company, Limited
C/O Mr. Glen Feye
President
Accurate Consultants Incorporated
1340 West Pennsylvania Avenue
San Diego, California 92103

SEP 29 2011

Re: K112136
Trade/Device Name: Surgical and Equipment Drape-Sterile and Non Sterile
Regulation Number: 21 CFR 878.4370
Regulation Name: Surgical Drape and Drape Accessories
Regulatory Class: II
Product Code: KXX
Dated: July 20, 2011
Received: July 26, 2011

Dear Mr. Feye:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 – Mr. Feye

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Section 5: Indications for Use Statements

Indications for Use

510(k) Number (if known): K112136

Device Name: **Surgical and Equipment Drape-Sterile and Non Sterile**

Indications for Use:

Disposable sterile drapes are intended for use for :

- patient protective covering used to isolate incision sites and protect against contamination during surgical procedures.
- cover a variety of surgical and non-surgical equipment in various settings throughout the clinical setting. These drapes are used protect surgical and non-surgical equipment from contamination during surgical procedures

Drapes provided as sterile and non-sterile. Non-sterile surgical drapes are to be sold to OEMs for EtO sterilization according to ISO 11135. Sterile Surgical Drapes are to be sold directly to users after EtO sterilization validation to ISO 11135.

The following disposable surgical drape types and model numbers are included in this submission

MATERIAL	DRAPE STYLE	STERILE	DRAPE TYPE	MODEL #
SMS	Non Reinforced	Sterile	Top	JHSAD 150-240E
			Bottom	JHSAD 180-175E
		Non Sterile	Top	JHSAD 150-240NS
			Bottom	JHSAD 180-175NS

Prescription Use _____
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use X
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Kunshan Jiehong NonWoven Product Corp. Ltd.
Section 510(k) Notification
Surgical and Equipment Drapes

Elizabeth P. O'Brien-Walsh
(Division Sign-Off)
Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

Page 7 of 24

510(k) Number: K112136