

510(k) Summary of Safety and Effectiveness

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 CFR 807.92.

The assigned 510(k) number is: __K112159__

SEP 21 2011

Submitter

Advanced Instrumentations, Inc.
6800 N.W. 77th Court
Miami, FL 33166
Telephone: 305-477-6331
Fax: 305-477-5351

Registration # 1066270

Official correspondent :

Jorge Millan, PhD
Email: jmillan@hiatec.org
601 West 20 St
Hialeah, FL 33010
Phone : (305) 925-1260

Date Prepared:

September 16, 2011

Device name and classification:

- **Device Name:** CMS-1000FM Central Monitoring System
- **Classification Name:** 884.2740 Perinatal Monitoring System and accessories
- **Product code:** HGM
- **Regulatory Class:** Class II

Predicate Device:

Central Monitoring System (model MFM-CNS) K100358 Manufacturer: EDAN Instruments

Device Description:

CMS-1000FM Central Monitoring System is a software product that runs on a PC station running the Microsoft Windows XP operating system. The CMS-1000FM Central Monitoring System connects one central station with some bedside fetal / maternal monitors, and carries out centralized monitoring management for many beds. It can monitor a pregnant woman during the whole parturition process, and all the monitoring information can be recorded, saved and printed, and alarm when the parameter exceeds the user defined limit and poor signal quality. At the same time, the old records can be searched conveniently and quickly.

Device features:

- Connect maximum **32** bedside fetal / maternal monitors with Ethernet.
- Display FHR, **UA**, Maternal HR, PR, **SpO2**, NIBP, RR and TEMP numeric on the screen.
- The screen displays all the monitors simultaneously, or displays one monitor in full screen.
- 24-hour **CTG**, 1440-group Maternal Vital Sign data, 200-group NIBP data review.
- Print CTG report, Maternal Vital Sign list, NIBP list on the paper
- Audible & visible alarm when FHR or maternal vital sign exceeds the user defined limit or poor signal quality.
- Patient information, **CTG**, Maternal Vital Sign list and NIBP list can be saved, and burned into CDs for backup.
- Support user accessing control

Intended Use:

The CMS-1000FM Central Monitoring System (hereinafter called "CMS-1000FM") is a clinical data managing software application and is indicated for antepartum and intrapartum monitoring of pregnant women in a healthcare setting.

The CMS-1000FM is intended to manage perinatal monitoring data acquired from bedside monitors or manual input for viewing at the central nursing station. The system also produces an electronic medical record.

The CMS-1000FM has display fields for the following obstetric data:

- patient demographics

- provider notes
- FHR
- uterine activity (via tocodynamometry or IUP)
- maternal heart rate
- SpO₂
- NIBP
- respiratory rate
- temperature
- pulse rate

Comparison to the predicate device:

The subject device has similar technology characteristics and has the same intended use as the predicate device.

Substantially Equivalent Determination:

This premarket notification submission demonstrates that CMS-1000FM Central Monitoring System is substantially equivalent to the predicate device.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center - WO66-G605
Silver Spring, MD 20993-0002

Advanced Instrumentations, Inc.
% Jorge Millan, Ph.D.
Executive Director
Hiialeah Technology Center, Inc.
601 West 20 Street
HIALEAH FL 33010

SEP 21 2011

Re: K112159
Trade Name: CMS-1000FM Central Monitoring System
Regulation Number: 21 CFR § 884.2740
Regulation Name: Perinatal monitoring system and accessories
Regulatory Class: II
Product Code: HGM
Dated: September 12, 2011
Received: September 14, 2011

Dear Dr. Millan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related

adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOoffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Herbert P. Lerner, M.D., Director (Acting)
Division of Reproductive, Gastro-Renal
and Urological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K112159

Device Name:

CMS-1000FM Central Monitoring System

Indications for Use:

The CMS-1000FM Central Monitoring System (hereinafter called "CMS-1000FM") is a clinical data managing software application and is indicated for antepartum and intrapartum monitoring of pregnant women in a healthcare setting.

The CMS-1000FM is intended to manage perinatal monitoring data acquired from bedside monitors or manual input for viewing at the central nursing station. The system also produces an electronic medical record.

The CMS-1000FM has display fields for the following obstetric data:

- patient demographics
- provider notes
- FHR
- uterine activity (via tocodynamometry or IUP)
- maternal heart rate
- SpO 2
- NIBP
- respiratory rate
- temperature
- pulse rate

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

[Signature]
 (Division Sign-Off)
 Division of Reproductive, Gastro-Renal, and
 Urological Devices
 510(k) Number K112159