

SEP 28 2011

**510(K) SUMMARY****1. 510(k) Owner:**

Covidien
15 Hampshire Street
Mansfield, MA 02048
Telephone: (508) 261 - 6596
Fax: (508) 261 - 8149

Contact: Mr. Wing Ng
Title: Manager, Regulatory Affairs
Date Prepared: August 23, 2011

2. Device:

Trade Names: Palindrome™ H Chronic Catheter
Palindrome™ SI Chronic Catheter
Palindrome™ HSI Chronic Catheter
Common Name: Catheter
Classification Name: Implanted Hemodialysis Catheter
Coated Implanted Hemodialysis Catheter
Regulation Number: 21 CFR 876.5540
Product Code(s): MSD, NYU
Classification: Class III

3. Predicate Devices:

Palindrome™ H Chronic Catheter (K060509)
Palindrome™ SI Chronic Catheter (K060972)
Palindrome™ HSI Chronic Catheter (K062671)
Modified Chronic Hemodialysis Catheters (K111372)

4. Device Description:

The Palindrome™ H Chronic Catheter with Heparin coating has a radiopaque polyurethane shaft with two large inner lumens designed in a "double D" configuration. The distal end of the catheter extends to a symmetrical tip. The proximal end of the catheter shaft contains a polyurethane hub assembly and silicone extension sets. The catheter contains a heparin coating on its surface from the tip of the catheter to the cuff on the external surface and throughout the entire length on the internal surface (tip to luer adapters). The heparin coating serves to reduce platelet adhesion.

The Palindrome™ SI Chronic Catheter with Silver Impregnated sleeve has a radiopaque polyurethane shaft with two large inner lumens designed in a "double D" configuration.

The distal end of the catheter extends to a symmetrical tip. The proximal end of the catheter shaft contains a polyurethane hub assembly and silicone extension sets. The catheter contains a silver impregnated sleeve permanently bonded to the outer surface of the device from the hub to the cuff. The silver impregnated sleeve serves to reduce microbial colonization on the external surface of the sleeve which is placed within the subcutaneous tunnel tract.

The Palindrome™ HSI Chronic Catheter with Heparin coating and Silver Impregnated sleeve has a radiopaque polyurethane shaft with two large inner lumens designed in a "double D" configuration. The distal end of the catheter extends to a symmetrical tip. The proximal end of the catheter shaft contains a polyurethane hub assembly and silicone extension sets. The catheter contains a heparin coating on its surface from the tip of the catheter to the cuff on the external surface and throughout the entire length on the internal surface (tip to luer adapters). The heparin coating serves to reduce platelet adhesion. The catheter also contains a silver impregnated sleeve permanently bonded to the outer surface of the device from the hub to the cuff. The silver impregnated sleeve serves to reduce microbial colonization on the external surface of the sleeve which is placed within the subcutaneous tunnel tract.

5. Intended Use:

The Palindrome™ H Chronic Catheter is intended for acute and chronic hemodialysis, apheresis, and infusion. It may be inserted either percutaneously or by cutdown. The performance of the heparin coating on this catheter in reducing platelet adhesion on the catheter surface for up to 720 hours of dialysis treatment is supported by bench and animal testing.

The Palindrome™ SI Chronic Catheter is intended for acute and chronic hemodialysis, apheresis, and infusion. It may be inserted either percutaneously or by cutdown. Catheters greater than 40 cm implant length are indicated for femoral insertion. The performance of the silver impregnated sleeve in reducing colonization on the catheter surface for up to 30 days is supported by bench and animal testing.

The Palindrome™ HSI Chronic Catheter is intended for acute and chronic hemodialysis, apheresis, and infusion. It may be inserted either percutaneously or by cutdown. The performance of the heparin coating on this catheter in reducing platelet adhesion on the catheter surface for up to 720 hours of dialysis treatment is supported by bench and animal testing. The performance of the silver impregnated sleeve in reducing colonization on the catheter surface for up to 30 days is supported by bench and animal testing.

6. Technological Characteristics:

The modified devices have the same technological characteristics as compared to their respective predicate devices.

7. Performance Data:

Bench top functional testing was completed to support substantial equivalence between the modified device and the current device. The test regimen evaluated the devices'

resistance to kink, leak, burst, catheter collapse, fatigue, and the tensile strength at various points of the catheter. Functional testing also evaluated heparin activity, heparin concentration, coating durability, and sleeve adhesion. The results of the performance testing show that the modified devices continue to meet the relevant product specifications.

Biocompatibility testing per ISO 10993: Biological Evaluation of Medical Devices was completed to support biocompatibility between the modified device and the current device. Material characterization testing was included to show material equivalence where applicable. The results of the biocompatibility testing show that the modified devices continue to be biocompatible for its intended use.

The results of functional testing, biocompatibility testing, and material analytical testing support the determination of substantial equivalence.

8. Conclusion:

Based on non-clinical testing results, Covidien has demonstrated that the modified coated catheters are substantially equivalent to their respective existing coated catheters.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room –WO66-G609
Silver Spring, MD 20993-0002

Mr. Wing Ng
Manager, Regulatory Affairs
Covidien
Vascular Therapies
15 Hampshire Street
MANSFIELD MA 02048

SEP 28 2011

Re: K112477

Trade/Device Name: Palindrome™ H Chronic Catheter
Palindrome™ SI Chronic Catheter
Palindrome™ HSI Chronic Catheter

Regulation Number: 21 CFR §876.5540

Regulation Name: Blood access device and accessories

Regulatory Class: III

Product Code: NYU, MSD

Dated: August 23, 2011

Received: August 29, 2011

Dear Mr. Ng:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. However, you are responsible to determine that the medical devices you use as components in the kit have either been determined as substantially equivalent under the premarket notification process (Section 510(k) of the act), or were legally on the market prior to May 28, 1976, the enactment date of the Medical Device Amendments. *Please note:* If you purchase your device components in bulk (i.e., unfinished) and further process (e.g., sterilize) you must submit a new 510(k) before including these components in your kit. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, and labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Herbert P. Lerner, M.D., Director (Acting)
Division of Reproductive, Gastro-Renal,
and Urological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Section 4
Indications for Use Statement

510(k) Number (if known): K112477

Device Name: Palindrome™ H Chronic Catheter
Palindrome™ SI Chronic Catheter
Palindrome™ HSI Chronic Catheter

Indications for Use:


The Palindrome™ H Chronic Catheter is indicated for acute and chronic hemodialysis, apheresis, and infusion. It may be inserted either percutaneously or by cutdown. The performance of the heparin coating on this catheter in reducing platelet adhesion on the catheter surface for up to 720 hours of dialysis treatment is supported by bench and animal testing.

The Palindrome™ SI Chronic Catheter is indicated for acute and chronic hemodialysis, apheresis, and infusion. It may be inserted either percutaneously or by cutdown. Catheters greater than 40 cm implant length are indicated for femoral insertion. The performance of the silver impregnated sleeve in reducing colonization on the catheter surface for up to 30 days is supported by bench and animal testing.

The Palindrome™ HSI Chronic Catheter is indicated for acute and chronic hemodialysis, apheresis, and infusion. It may be inserted either percutaneously or by cutdown. The performance of the heparin coating on this catheter in reducing platelet adhesion on the catheter surface for up to 720 hours of dialysis treatment is supported by bench and animal testing. The performance of the silver impregnated sleeve in reducing colonization on the catheter surface for up to 30 days is supported by bench and animal testing.

Prescription Use X AND/OR Over-The-Counter Use
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)


(Division Sign-Off)
Division of Reproductive, Gastro-Renal, and
Urological Devices
510(k) Number K112477