

FEB 16 2012

510(k) SUMMARY K 112831

A summary of 510(k) safety and effectiveness information in accordance with 21 CFR 807.92.

SUBMITTER INFORMATION	
Name	CareFusion
Address	1500 Waukegan Road MPWM, McGaw Park, IL 60085 USA
Phone number	(847) 473-7404
Fax number	(847) 473-7990
Establishment Registration Number	1423507
Name of contact person	Joy Greidanus
Date prepared	February 10, 2012
NAME OF DEVICE	
Trade or proprietary name	Pleurx Pleural Catheter System
Common or usual name	Pleural Drainage Catheter
Classification name	Patient Care Suction Apparatus
Classification panel	Anesthesiology
Regulation	Class II per 21CFR §870.5050, Procode DWM
Product Code(s)	Multiple
Legally marketed device(s) to which equivalence is claimed	CareFusion Pleurx Catheter Kit and the Denver Pleurx Drainage Kit/Vacuum Bottle : K051084 & K052436 Bard Aspira Pleural Drainage System: K110409
Reason for 510(k) submission	Change to materials, expanding the indications for use, and updates to the instructions for use.
Device description	The Pleurx Pleural Catheter System provides patients with a convenient method to relieve pleural effusion symptoms at home. The primary components of the Pleurx Catheter System are the Pleurx Pleural Catheter and the Pleurx Drainage Kits.

Intended use of the device	<p>The Pleurx Pleural Catheter Kits are indicated for intermittent, long term drainage of symptomatic, recurrent, pleural effusion, including malignant pleural effusion and other recurrent effusions that do not respond to medical management of the underlying disease. The devices are indicated for 1) the palliation of dyspnea due to pleural effusion and 2) providing pleurodesis (resolution of the pleural effusion).</p> <p>The Pleurx Drainage Kits and Drainage Line Set are indicated for use only with the Pleurx Catheter for intermittent drainage. The Drainage Line Kit is used to drain fluid using standard wall suction, water seal drainage system, vacuum bottle or other appropriate method.</p> <p>The Pleurx Dressing Kits are indicated for dressing of a catheter and exit site.</p>
----------------------------	---

SUMMARY OF THE TECHNOLOGICAL CHARACTERISTICS OF THE DEVICE COMPARED TO THE PREDICATE DEVICE

Characteristic	New Device	Predicates: CareFusion Pleurx Pleural Catheter System (K051084, K052436) Bard Aspira Pleural Catheter System (K110409)
Catheter Description	Internal: fenestrations, radiopaque markings & cuff External: valve	Same
Method	Percutaneously tunneled - indwelling	Same
Means of Drainage	Wall suction, water seal drainage system, portable suction, vacuum bottles or other appropriate method	Wall suction, water seal drainage system, vacuum bottles, syringe, drainage bag or other appropriate method

PERFORMANCE DATA

SUMMARY OF NON-CLINICAL TESTS CONDUCTED FOR DETERMINATION OF SUBSTANTIAL EQUIVALENCE

Performance Test Summary

Characteristic	Standard/Test/FDA Guidance
Biocompatibility	ISO 10993-1:2009 Biological evaluation of Medical Devices Part 1: Evaluation and Testing
Residuals	ISO 10993-7:2008 Biological evaluation of Medical Devices Part 7: Ethylene Oxide Sterilization Residuals
Performance	EN 1617:1997 Sterile Drainage Catheters and Accessory Devices for Single Use
Performance	EN 1618:1997 Catheters Other Than Intravascular Catheters – Test Methods for Common Properties

Performance	ANSI/AAMI/ISO 11607-1,2:2006 Packaging for Terminally Sterilized Medical Devices
Performance	ISO 11138-1,2:2006 Sterilization of Healthcare products - Biological Indicators
Performance	ISO 11737-1,2:2006 Sterilization of Medical Devices – Microbiological Methods Part 1 & 2
Performance	ISO 11135:2007 Medical Device, Validation and Routine Control of Ethylene Oxide Sterilization
Specification	ASTM F560-08 (excluding part 7) Standard Specification for Unalloyed Tantalum for Surgical Implants
SUMMARY OF CLINICAL TESTS CONDUCTED FOR DETERMINATION OF SUBSTANTIAL EQUIVALENCE AND/OR OF CLINICAL INFORMATION	
N/A – No clinical tests were conducted for this submission	
CONCLUSIONS DRAWN FROM NON-CLINICAL AND CLINICAL DATA	
The results of the non-clinical tests show that the CareFusion Pleurx Pleural Catheter System meets or exceed all performance requirements, and is substantially equivalent to the predicate devices.	



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

Ms. Joy Greidanus
Manager, Regulatory Affairs
CareFusion
1500 Waukegan Road
McGaw Park, Illinois 60085

FEB 16 2012

Re: K112831
Trade/Device Name: Pleurx Pleural Catheter System
Regulation Number: 21 CFR 870.5050
Regulation Name: Patient Care Suction Apparatus
Regulatory Class: II
Product Code: DWM
Dated: February 13, 2012
Received: February 14, 2012

Dear Ms. Griedanus:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure



1500 Waukegan Road
McGaw Park, Illinois 60085-6787
PH: 847.473.7404
FAX: 847.473.7790

Indication for Use

510(k) Number (if known):

K112831

Device Name:

Pleurx Pleural Catheter System

Indications For Use:

The Pleurx Pleural Catheter Kits are indicated for intermittent, long term drainage of symptomatic, recurrent, pleural effusion, including malignant pleural effusion and other recurrent effusions that do not respond to medical management of the underlying disease. The devices are indicated for 1) the palliation of dyspnea due to pleural effusion and 2) providing pleurodesis (resolution of the pleural effusion).

The Pleurx Drainage Kits and Drainage Line Set are indicated for use only with the Pleurx Catheter for intermittent drainage. The Drainage Line Kit is used to drain fluid using standard wall suction, water seal drainage system, vacuum bottle or other appropriate method.

The Pleurx Dressing Kits are indicated for dressing of a catheter and exit site.

Prescription Use X (Per 21 CFR 801.109) or Over-The Counter Use _____

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE)

Concurrence of CDRH, Office of Device Evaluation (ODE)

(Division Sign-Off)

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K 112 831