

OCT 28 2011

### Section III 510(k) Summary

This 510(k) Summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of SMDA 1990 and 21 CFR 807.92.

The assigned 510(k) Number: K113091

1. Date of Submission: 27 AUG 2011.
2. Sponsor  
Jiangyin Caina Technology Co., Ltd.  
No.2, Taifu Road, Huashi Town, Jiangyin  
Jiangsu, 214421, China  
  
Contact Person: Mr. Jun Lu  
Position: General Manager  
Tel: +86-510-86205183  
Fax: +86-510-86215183  
Email: jun.lu@cainamedical.com
3. Submission Correspondent  
Ms. Diana Hong & Mr. Lee Fu  
Mid-Link Consulting Co., Ltd  
P.O. Box 237-023, Shanghai, 200237, China  
Tel: +86-21-22815850  
Fax: 240-238-7587  
Email: info@mid-link.net

4. Proposed Device Identification

Proposed Device Name: Syringes with or without needles  
Classification: Class II  
Product Code: FMF  
Regulation Number: 21 CFR 880.5860  
Review Panel: General Hospital  
Intended Use Statement:  
Syringes with or without needle are intended to inject fluids into or withdraw fluids from the body.

Proposed Device Name: Needles

Classification: Class II

Product Code: FMI

Regulation Number: 21 CFR 880.5570

Review Panel: General Hospital

Intended Use Statement:

Needles are intended to be used with a legally marketed syringe intend to inject fluids into or withdraw fluids from the body.

#### 5. Predicate Device Identification

510(k) Number: K072739

Product Name: Sterile Hypodermic Syringe for single use, with/without needle and Sterile Hypodermic Needle for single use

Manufacturer: Shandong Weigao Group Medical Polymer Products Co., Ltd

#### 6. Device Description

The proposed devices include: a standard piston syringes consisting of the proposed syringes consist of three components: (1) barrel, (2) piston and (3) plunger rod, and a needle consists of (1) needle tube, (2) needle hub and (3) needle sheath. They would be available in three different configurations respectively as follows:

##### Configuration 1 Syringes with Needle

Proposed syringes will be provided along with a proposed needle in one single package. The combinations of sizes are various upon the request of the users.

##### Configuration 2 Syringes without Needle

Proposed syringes will be provided without a needle. The combination of sizes are various upon the request of the users. a) It could be used together with any other U.S. legally marketed hypodermic needle with luer slip female connector complied with ISO594-1:1988 or luer lock female connector complied with ISO 594-2:1996.

##### Configuration 3 Needles

Proposed needles will be provided alone, which could be used together with any other U.S. legally marketed syringes with luer slip male connector complied with ISO594-1:1988 or luer lock male connector complied with ISO 594-2:1996 to complete its intended use.

7. Non-Clinical Test Conclusion

Bench tests were conducted to verify that the proposed device met all design specifications as was Substantially Equivalent (SE) to the predicate device per the following standards:

- ISO 7886-1:1993 Sterile hypodermic syringes for single use - Part 1: Syringes for manual use;
- ISO 7864:1993 Sterile hypodermic needles for single use.
- ISO 9626:1991, AMENDMENT 1 2001 Stainless steel needle tubing for the manufacture of medical devices.
- ISO 594-1:1986 Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment - Part 1: General requirements;
- ISO 594-2:1998 Conical fittings with a 6% (Luer) taper for syringes, needles and certain other medical equipment - Part 2: Lock fittings.

8. Substantially Equivalent Conclusion

Compared with the predicate devices, the proposed devices are mainly different in size specifications and sterilization method from the predicate devices. But the sterilization validation was performed to demonstrate that the SAL of the proposed device were the same to that of the predicate devices, and the performance of both proposed and predicate device complied with same performance standards.

The proposed devices, are determined to be Substantially Equivalent (SE) to the predicate devices, in respect of safety and effectiveness.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room --WO66-G609  
Silver Spring, MD 20993-0002

Jiangyin Caina Technology Company, Limited  
C/O Mr. Marc M. Mouser  
Responsible Third Party Official  
Underwriters Laboratories, Incorporated  
2600 NW Lake Road  
Camas, Washington 98607-9526

OCT 28 2011

Re: K113091  
Trade/Device Name: Syringes, With or without Needles  
Regulation Number: 21 CFR 880.5860  
Regulation Name: Piston Syringe  
Regulatory Class: II  
Product Code: FMF, FMI  
Dated: September 28, 2011  
Received: October 19, 2011

Dear Mr. Mouser:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.

Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

## Section II Indications for Use

510(k) Number: K113091

Device Name: Syringes, with or without needles

Indications for Use:

Syringes with or without needle are intended to inject fluids into or withdraw fluids from the body.

PRESCRIPTION USE  
(Part 21 CFR 801 Subpart D)

OVER-THE-COUNTER USE  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page 1 of 2

Robert C. Chynoweth 10/28/11

Division Sign-Off)

Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

510(k) Number: K113091

Indication for Use

510(k) Number: K113091

Device Name: Needles

Indications for Use:

Needles are intended to be used with a legally marketed syringe intend to inject fluids into or withdraw fluids from the body.

PRESCRIPTION USE  
(Part 21 CFR 801 Subpart D)

OVER-THE-COUNTER USE  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page 2 of 2

*Richard C. Olyn* 10/28/11

(Division Sign-Off)

Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

510(k) Number: K113091