

K113400

DEC 16 2011

510(k) Summary

As Required by 21 CFR 807.92

Submitter: Anulex Technologies, Inc.
5600 Rowland Road, Suite 280
Minnetonka, MN 55343

Contact Person: Rachel Kennedy
Director, Regulatory Affairs and Quality Systems
Anulex Technologies, Inc.
Direct: 952.224.4034
Fax: 952.224.5417
E-Mail: rkennedy@anulex.com

Date Prepared: November 15, 2011

Trade Name: fiXate Tissue Band

Classification: II

Product Code: GZB
21 CFR 882.5880

Predicate Device(s): The subject device is substantially equivalent to the following predicate device:

- Anulex Technologies fiXate Tissue Band (K111462 cleared September 8, 2011 and K112849 cleared October 27, 2011)

Device Description: The fiXate Tissue Band consists of an adjustable loop of non-absorbable 2-0 suture with two (2) attached anchors. The construct is provided sterile and preloaded on a disposable delivery instrument. The instrument's needle facilitates placement of the suture by positioning the T-anchors in the sub-layer of tissue.

Indications for Use: The fiXate Tissue Band is intended to be an accessory to the leads component of Spinal Cord Stimulator systems functioning to secure the lead to the fascia or inter-spinous/supra-spinous ligament.

Functional and Safety Testing: Biocompatibility and bench testing were completed and support the safety and effectiveness of the fiXate Tissue Band.

Conclusion: The fiXate Tissue Band is similar in materials, design, and performance characteristics and has the same intended use as the original fiXate Tissue Band (K111462). Substantial equivalence is demonstrated through the detailed device description, performance testing and conformance with voluntary performance standards.



Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center - WO66-G609
Silver Spring, MD 20993-0002

Anulex Technologies, Inc.
c/o Ms. Rachel Kennedy
Director of Regulatory and Quality Systems
5600 Rowland Road, Suite 280
Minnetonka, Minnesota 55343

DEC 16 2011

Re: K113400

Trade/Device Name: fiXate Tissue Band
Regulation Number: 21 CFR 882.5880
Regulation Name: Implanted Spinal Cord Stimulator for Pain Relief
Regulatory Class: Class II
Product Code: GZB, GAT
Dated: November 16, 2011
Received: November 17, 2011

Dear Ms. Kennedy:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Malvina B. Eydelman, M.D.

Director

Division of Ophthalmic, Neurological,
and Ear, Nose and Throat Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use Statement

Device Name: fiXate Tissue Band

Indications for Use:

The fiXate Tissue Band is intended to be an accessory to the leads component of Spinal Cord Stimulator systems functioning to secure the lead to the fascia or inter-spinous/supra-spinous ligament.

Prescription Use Use _____

AND/OR

Over-The-Counter

(Part 21 CFR 801 Subpart D
C)

(21 CFR 801 Subpart

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

KRISTEN BOWSHER
(Division Sign-Off)
Division of Ophthalmic, Neurological and Ear,
Nose and Throat Devices

510(k) Number 113400