

MAR 13 2012

**510(k) Summary**  
**MegaBeam Reusable Fiber Optic Delivery System**

K113688

**Submitter's Name, Address, Telephone Number, Contact Person  
and Date Prepared**

Biolitec Medical Devices, Inc.  
515 Shaker Road  
East Longmeadow, Massachusetts 01028  
Phone: (413) 525-0600  
Facsimile: (413) 525-0611

Contact Person: Harry Hayes, Ph.D. – Regulatory Consultant  
Date prepared: December 14, 2011

**Name of Device and Name/Address of Sponsor**

MegaBeam Reusable Fiber Optic Delivery System  
Biolitec Medical Devices, Inc.  
515 Shaker Road  
East Longmeadow, Massachusetts 01028

**Classification Name**

Surgical laser accessory

**Predicate Devices**

MegaBeam Reusable Fiber Optic Delivery System, (K980838)

**Intended Use/Indication for Use**

The MegaBeam Reusable Fiber Optic Delivery System is intended for use as an accessory to any surgical laser system to vaporize, coagulate, incise and excise tissue for any application for which the lasers are cleared.

**Technological Characteristics**

The MegaBeam Reusable Fiber Optic Delivery System for Biolitec Medical Devices, Inc. contains the identical same components and design as the device cleared under K980838 for Biolitec Inc. There are no differences in technology and as such does not raise any new questions on safety or efficacy.

**Performance Data**

Since the performance of the MegaBeam Reusable Fiber Optic Delivery System is well established and documented on soft tissue no performance testing is being specifically included in this submission.

**Substantial Equivalence**

The MegaBeam Reusable Fiber Optic Delivery System is as safe and effective for these Indications for Use as the Biolitec Inc MegaBeam Reusable Fiber Optic Delivery System. Thus, the MegaBeam Reusable Fiber Optic Delivery System for Biolitec Medical Devices, Inc. is substantially equivalent to its predicate device.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room -WO66-G609  
Silver Spring, MD 20993-0002

Biolitec Medical Devices, Inc.  
% Mr. Harry Hayes  
515 Shaker Road  
East Longmeadow, Massachusetts 01028

MAR 13 2012

Re: K113688

Trade/Device Name: MegaBeam Reusable Fiber Optic Delivery System

Regulation Number: 21 CFR 878.4810

Regulation Name: Laser surgical instrument for use in general and plastic surgery and  
in dermatology

Regulatory Class: II

Product Code: GEX

Dated: December 14, 2011

Received: December 15, 2011

Dear Mr. Hayes:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices; good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21


Page 2 - Mr. Harry Hayes

CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Mark N. Melkerson

Director

Division of Surgical, Orthopedic  
and Restorative Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use Statement

510(k) Number (if known): K 113688

Device Name: **MegaBeam Reusable Fiber Optic Delivery System**

Indications for Use: The MegaBeam Reusable Fiber Optic Delivery System is intended for use as an accessory to any surgical laser system to vaporize, coagulate, incise and excise tissue for any application for which the lasers are cleared.

(PLEASE DO NOT WRITE BELOW THIS LINE -- CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Neil A.P. Ogden for mxm

(Division Sign-Off)  
Division of Surgical, Orthopedic,  
and Restorative Devices

510(k) Number K 113688

Prescription Use   
(Per 21 C.F.R. 801.109)

OR

Over-The-Counter Use   
(Optional Format 1-2-96)