



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room – WO66-G609
Silver Spring, MD 20993-0002

Ivoclar Vivadent AG
C/O Ms. Donna Marie Hartnett
Director of Quality Assurance / Regulatory Affairs
Ivoclar Vivadent, Incorporated
175 Pineview Drive
Amherst, New York 14228

OCT 18 2012

Re: K120053
Trade/Device Name: IPS e.max[®] Press – Abutment Solutions
Regulation Number: 21 CFR 872.3630
Regulation Name: Endosseous Dental Implant Abutment
Regulatory Class: II
Product Code: NHA, EIH
Dated: September 28, 2012
Received: October 2, 2012

Dear Ms. Hartnett:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to

<http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

K120053

Indications for Use

510(k) Number (if known): K120053

Device Name: IPS e.max® Press – Abutment Solutions

Indications For Use:

IPS e.max® Press Abutment Solutions is intended for use in partially or fully edentulous mandibles and maxillae in support of single or multiple-unit cement retained restorations.

IPS e.max Press Abutment Solutions is recommended for the fabrication of:

- Hybrid abutments for single-tooth restorations
- Hybrid abutment crowns for restorations

The following Ti bases are intended to be used with IPS e.max Press Abutment Solutions.

Implant manufacture	Implant system, diameter	Compatible Ti base (abutment), dimensions diameter = D gingiva height = GH height = HTi
Straumann®	Bone Level RC Ø 4.1 mm or Ø 4.8 mm 510K K062129	RC Cementable abutment D 5.0–6.5 mm GH 1.0–3.0 mm HTi 4.0–5.5 mm 510K K072071

Prescription Use AND/OR
(Part 21 CFR 801 Subpart D)

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)



 (Division Sign-Off)
 Division of Anesthesiology, General Hospital
 Infection Control, Dental Devices

510(k) Number: K120053