

Unimed Medical Supplies Inc.

No. 37, Yanshan Road, Shekou, Shenzhen, China 518067
Tel: 86 755 26695165 Fax: 86 755 26697984
Website: www.unimed.cn Email: info@unimed.cn



Section 3

510(K) Summary

Summary prepared Date: January 12, 2012

Submitter Information:

Unimed Medical supplies Inc.

No.37, Yanshan Road, Shekou, Shenzhen, China 518067

Contact Person:

Xinmei Tan, QA manager

No.37, Yanshan Road, Shekou, Shenzhen, China 518067

Tel: (86) 755 26695165

Fax: (86) 755 26697984

Email: tanxinmei@unimed.cn; caijiye@unimed.cn

Device Name

Trade Name: Unimed Disposable Blood Pressure Cuff

Common Name: Non-invasive Blood pressure cuff

Product Code/Classification: DXQ/21 CFR870.1120

Review Panel: Cardiovascular

Predicate Device

Soft-Cuff (K974080)

JOHNSON & JOHNSON MEDICAL, INC. (GE Healthcare)

Intended Use

The Unimed disposable blood pressure cuff is an accessory used in conjunction with

Unimed Medical Supplies Inc.

No. 37, Yanshan Road, Shekou, Shenzhen, China 518067
 Tel: 86 755 26695165 Fax: 86 755 26697984
 Website: www.unimed.cn Email: info@unimed.cn



noninvasive blood pressure measurement systems. The cuff is non-sterile and for single-patient use. It is available in neonatal, pediatric and adult sizes. The cuff is not designed, sold, or intended for use except as indicated.

Device Description

The device comprises tubing attached to an inelastic sleeve with an integrated inflatable bladder that is wrapped around the patient's limb and secured by hook and loop closure. The device tubing is connected to a non-invasive blood pressure measurement system.

Comparison to the Predicate Device

Features	Unimed Medical	JOHNSON & JOHNSON MEDICAL (GE Healthcare)
Intended use	Indirect measurement of blood pressure	Indirect measurement of blood pressure
Patient Populations	Adults/Pediatrics	Adults/Pediatrics
Tube Configuration	One or two tube	One or two tube
Size (Range in cm)	Conform to AHA bladder sizes recommendations Neonate 1 (3-6) Neonate 2 (4-8) Neonate 3 (6-11) Neonate 4 (7-13) Neonate 5 (8-15) Infant (9-14.8) Child (13.8-21.5) Small Adult (20.5-28.5) Adult (27.5-36.5) Adult Long(27.5-36.5) Large Adult (35.5-46cm) Large Adult Long(35.5-46cm) Thigh (45-56cm)	Conform to AHA bladder sizes recommendations Neonatal 1 (3-6) Neonatal 2 (4-8) Neonatal 3 (6-11) Neonatal 4 (7-13) Neonatal 5 (8-15) Infant (8-13) Child (12-19) Small Adult (17-25) Adult (23-33) Large Adult (31-40) Thigh (38-50)

Unimed Medical Supplies Inc.

No. 37, Yanshan Road, Shekou, Shenzhen, China 518067
 Tel: 86 755 26695165 Fax: 86 755 26697984
 Website: www.unimed.cn Email: info@unimed.cn



Sterility	Not supplied sterile	Not supplied sterile
Pressure limits	0-300mmHg	0-300mmHg
Sterility	Non-sterile	Non-sterile
Biocompatibility	Comply with ISO 10993 biocompatibility evaluation	Comply with ISO 10993 biocompatibility evaluation

The Unimed Disposable Blood Pressure Cuff has the same Intended Use, basic construction, and technology specification as the predicated device. Both devices are wrapped the patient's arm or leg and secured by a hook and loop fastener commonly called Velcro. Both devices are available in the same size and range and are intended for the same patient populations. The materials of both devices are all conformed to ISO 10993. We extend the length of the cuffs in order to accommodate special groups, such as overweight subjects. Based on the performance testing in this submission, the slight difference on the range of these blood pressure cuffs does not raise any safety or effectiveness issue.

Performance Summary

The Unimed Blood Pressure Cuff has been tested according to the following standards:

- ◆ ANSI/AAMI SP10, Manual, electronic or automated sphygmomanometers, 2002+A1:2003+A2:2006+(R)2008
- ◆ ISO 10993-1, Biological evaluation of medical devices -- Part 1: Evaluation and testing within a risk management process, 2009
- ◆ ISO 10993-5, Biological evaluation of medical devices -- Part 5: Tests for In Vitro cytotoxicity, 2009
- ◆ ISO 10993-10, Biological evaluation of medical devices -- Part 10: Tests for irritation and delayed-type hypersensitivity, 2002+A1:2006

Unimed Medical Supplies Inc.

No. 37, Yanshan Road, Shekou, Shenzhen, China 518067

Tel: 86 755 26695165 Fax: 86 755 26697984

Website: www.unimed.cn Email: info@unimed.cn



Conclusion

The subject device Unimed Blood Pressure Cuff has all features of the predicate device. The few differences do not affect the safety and effectiveness of the subject devices.

Thus, the subject device is substantially equivalent to the predicate device.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

MAY - 4 2012

Unimed Medical Supplies, Inc.
Xinmei Tan
Quality Assurance Manager
No.37, Yanshan Road, Shekou, Shenshen 518067
China

Re: K120364

Trade/Device Name: Unimed Blood Pressure Cuffs: U1710S, U1720S, U1730S, U1740S, U1750S, U1760S, U1770S, U1790S, U1710D, U1720D, U1730D, U1740D, U1750D, U1760D, U1770D, U1790D, U1781S, U1782S, U1783S, U1784S, U1785S, U1781D, U1782D, U1783D, U1784D, U1785D

Regulation Number: 21 CFR 870.1120

Regulation Name: Blood Pressure Cuff

Regulatory Class: Class II (two)

Product Code: DXQ

Dated: February 2, 2012

Received: February 6, 2012

Dear Mr. Tan:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Unimed Medical Supplies Inc.

No. 37, Yanshan Road, Shekou, Shenzhen, China 518067

Tel: 86 755 26695165 Fax: 86 755 26697984

Website: www.unimed.cn Email: info@unimed.cn



Section 4 Indications for use Statement

510(k) Number: _____

Device Name: Unimed Disposable Blood Pressure Cuff

Indications for Use

The Unimed disposable blood pressure cuff is an accessory used in conjunction with noninvasive blood pressure measurement systems. The cuff is non-sterile and for single-patient use. It is available in neonatal, pediatric and adult sizes. The cuff is not designed, sold, or intended for use except as indicated.

Prescription Use X AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Sub part D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

A large, stylized handwritten signature in black ink, written over a horizontal line.

(Division Sign-Off)
Division of Cardiovascular Devices

Page 1 of 1

510(k) Number K120364