

K122182

AUG 16 2012

Special 510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of 21 CFR Part 807.92.

Date: June 29, 2012

1. Company sponsoring this submission:

Name –Rayence Co., Ltd.

Address – (Seogu-dong, 2F/4F) 14, Samsung 1ro 1-gil, Hwaseong-si, Gyeonggi-do,
445-170, Korea

Telephone – +82-31-8015-6459

Fax – +82-31-8015-6598

Contact – Kee Dock Kim / Manager

Internet – <http://www.rayence.com>

2. Official correspondent (U.S. Designated agent)

Mtech Group

12946 Kimberley Ln Houston, TX 77079

Tel: +713-467-2607

Fax: +713-464-8880

Contact person: Mr. Dave Kim

Email: davekim@mtech-inc.net

3. Device :

Trade/proprietary name : 1717SGC
Common Name : Digital Flat Panel X-ray Detector
Classification Name : Solid State X-ray Imaging Device

4. Predicate Device :

Manufacturer : Rayence Co.,Ltd.
Device : Xmaru1717
510(k) Number : K091090 (Decision Date - September 9, 2010)

5. Classifications Names & Citations :

21CFR 892.1650, MQB, Solid State X-ray Imaging Device, Class2

6. Description :

6.1 General

1717SGC is a digital solid state X-ray detector that is based on flat-panel technology. This radiographic image detector and processing unit consists of a scintillator coupled to an a-Si TFT sensor. This device needs to be integrated with a radiographic imaging system. It can be utilized to capture and digitalize X-ray images for radiographic diagnosis. The RAW files can be further processed as DICOM compatible image files by separate console SW (not part of this 510k submission) for a radiographic diagnosis and analysis.

7. Indication for use :

1717SGC Digital Flat Panel X-Ray Detector is indicated for digital imaging solution designed for general radiographic system for human anatomy. It is intended to replace film or screen based radiographic systems in all general purpose diagnostic procedures. Not to be used for mammography.

8. Comparison with predicate device :

Rayence Co., Ltd. believes that 1717SGC is substantially equivalent in comparison with Xmaru1717 of Rayence Co., Ltd.

9. Safety, EMC and Performance Data :

Electrical, mechanical, environmental safety and performance testing according to standard IEC 60601-1:2005(Medical electrical equipment – Part 1: General requirements for basic safety and essential performance) + CORR.1(2006) + CORR.2 (2007) / EN 60601-1:2006 was performed, and EMC testing were conducted in accordance with standard IEC 60601-1-2:2007(Medical electrical equipment – Part 1-2: General Requirements for safety – Collateral Standard : Electromagnetic Compatibility Requirements and tests) / EN 60601-1-2:2007. All test results were satisfactory.

10. Conclusions :

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807 and based on the information provided in this premarket notification Rayence Co., Ltd. concludes that 1717SGC is safe and effective and substantially equivalent in comparison with the predicate device as described herein.

11. Rayence Co., Ltd. will update and include in this summary any other information deemed seasonably necessary by FDA.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room – WO66-G609
Silver Spring, MD 20993-0002

Rayence Co., Ltd.
% Mr. Dave Kim, MBA
Medical Device Regulatory Affairs
Mtech Group
12946 Kimberly Lane
HOUSTON TX 77079

AUG 16 2012

Re: K122182
Trade/Device Name: Digital Flat Panel X-Ray Detector/1717SGC
Regulation Number: 21 CFR 892.1650
Regulation Name: Image-intensified fluoroscopic x-ray system
Regulatory Class: II
Product Code: MQB
Dated: June 29, 2012
Received: July 23, 2012

Dear Mr. Kim:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of

medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely Yours,



Janine M. Morris
Acting Director
Division of Radiological Devices
Office of In Vitro Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indications for Use

AUG 16 2012

510(K) Number (if known): K122182

Device Name: Digital Flat Panel X-Ray Detector /1717SGC

Indications for Use:

1717SGC Digital Flat Panel X-Ray Detector is indicated for digital imaging solution designed for general radiographic system for human anatomy. It is intended to replace film or screen based radiographic systems in all general purpose diagnostic procedures. Not to be used for mammography.

Prescription Use AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (Part 21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

[Signature]
(Division Sign-Off)
Division of Radiological Devices
510k K122182