

510(k) SUMMARY

FEB 19 2013

Applicant: Ansell Healthcare Products, LLC
1635 Industrial Road
Dothan, AL 36303
USA
Phone: (732) 345-2174
Fax: (732) 603-7960

Contact Person: Robert Mahler, Regulatory Affairs Director, Americas

Date Prepared: January 19, 2013

510K: K122477

Proprietary Name: LifeStyles® Luxe™ Premium Personal Lubricant

Common Name: Personal Lubricant

Classification Name: Lubricant, patient, vaginal, latex compatible (Class II, 21 CFR 884.5300, Product Code NUC)

Predicate Device: Vielle™ Lubricant (K051288)

Device Description:

LifeStyles® Luxe™ Premium Personal Lubricant is a non-sterile, silicone-based personal lubricant designed to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. The formula is a biocompatible gel-like liquid that lasts longer than hydrous lubricants, without drying out or becoming sticky, and is compatible with natural rubber latex, polyurethane, and polyisoprene condoms. The product is provided in a plastic pump dispenser.

Indications for Use:

LifeStyles® Luxe™ Premium Personal Lubricant is a personal lubricant, for penile and/or vaginal application, intended to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is compatible with natural rubber latex, polyurethane, and polyisoprene condoms.

Technological Characteristics:

LifeStyles® Luxe™ Premium Personal Lubricant contains a similar blend of silicone oils as the predicate device. Testing per ASTM D7661 indicated that the lubricant formulation is compatible with natural rubber latex, polyisoprene, and polyurethane condoms. As with the predicate, testing for cytotoxicity, vaginal irritation, sensitization, and systemic toxicity in accordance with ISO 10993 indicated device biocompatibility. Bench testing indicated that the lubricant lasts longer than hydrous lubricants, without drying out or becoming sticky, and that it has an appropriate viscosity, specific gravity, appearance, color and odor for substantial equivalence to the predicate. Microbial limits testing was conducted per USP <61> and USP <62> indicated microbial quality. Real-time and accelerated aging tests indicate a 3 year shelf-life for the lubricant.

Summary:

LifeStyles® Luxe™ Premium Personal Lubricant has the same intended use and basic technological characteristics as the predicate device. This lubricant is substantially equivalent to its proposed predicate device.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

February 19, 2013

Ansell Healthcare Products, LLC
% Ms. Donna Di Gangi
Principal Consultant
DiGangi Consulting
4 Los Verdes Drive
SAN LUIS OBISPO CA 93401

Re: K122477

Trade/Device Name: LifeStyles® Luxe™ Premium Personal Lubricant
Regulation Number: 21 CFR§ 884.5300
Regulation Name: Condom
Regulatory Class: II
Product Code: NUC
Dated: January 29, 2013
Received: February 4, 2013

Dear Ms. Di Gangi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Benjamin R. Fisher -S

Benjamin R. Fisher, Ph.D.
Director
Division of Reproductive, Gastro-Renal,
and Urological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K122477

Device Name: LifeStyles® Luxe™ Premium Personal Lubricant

Indications for Use: LifeStyles® Luxe™ Premium Personal Lubricant is a personal lubricant, for penile and/or vaginal application, intended to moisturize and lubricate, to enhance the ease and comfort of intimate sexual activity and supplement the body's natural lubrication. This product is compatible with natural rubber latex, polyurethane, and polyisoprene condoms.

Prescription Use _____ AND/OR Over-The-Counter Use X
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Benjamin R. Fisher -S
2013.02.19 18:17:55 -05'00'

(Division Sign-Off)
Division of Reproductive, Gastro-Renal, and
Urological Devices
510(k) Number K122477