

510(k) Summary of Safety and Effectiveness**MAR 19 2013****Date Prepared:** March 12, 2013**Applicant:** Medtronic, Inc.
Medtronic Perfusion Systems
7611 Northland Drive
Minneapolis, MN 55428
Establish Registration Number: 2184009**Contact Person:** Julia A. Nelson
Principal Regulatory Affairs Specialist
Phone: (763) 514-9844
Fax: (763) 367-8360
E-mail: julia.a.nelson@medtronic.com**Trade Name:** Affinity® AF100 Arterial Filter with Carmeda® BioActive Surface
Common Name: Arterial Filter
Classification Name: Cardiopulmonary bypass arterial line blood filter
Classification: Class II, 21 CFR 870.4260**Product Code:** DTM**Name of Predicate Device:** Affinity Arterial Filter with Carmeda® BioActive Surface (20µm)
Model CB353 (K001138)**Device Description:**

The AF100 is designed to filter from the circuit microemboli larger than the specified micron size for periods up to six hours during cardiopulmonary bypass surgery.

The AF100 with Carmeda BioActive Surface (CB851) is coated with a nonleaching bioactive surface (heparin) to enhance blood compatibility and provide thromboresistant blood-contacting surfaces. The device is single-use, nontoxic, nonpyrogenic, supplied STERILE in individual packaging. The AF100 is sterilized by ethylene oxide.

Intended Use:

The AF100 is indicated for use in cardiopulmonary bypass procedures up to 6 hours in duration for the removal of particulate and gaseous microemboli.

Contraindications:

Do not use this device for any purpose other than indicated.

Comparison to Predicate Devices:

A comparison of Affinity AF100[®] Arterial Filter with Carmeda[®] BioActive Surface to the predicate device indicates the following similarities:

- Same intended use
- Same technological characteristics
- Same operating principle
- Same design features
- Similar materials with the exception of the housing material of the AF100 device. The AF100 housing is made of a Bisphenol A-free (BPA-free) copolyester material, which differs from the polycarbonate material used in the predicate device.
- Same shelf life

Summary of Performance Data

Pre-clinical bench testing was used to verify the performance characteristics of this device. Clinical testing was not required to establish substantial equivalence with the predicate devices.

The following performance tests were conducted:

- Blood Damage Testing
- Pressure Drop
- Structural Integrity
- Air Handling Capabilities
- Filtration Efficiency
- Burst Pressure
- Coating Integrity
- Priming Volume
- Particulate Shedding

Conclusion:

Pursuant to the statutory requirements under Section 513(i)(1)(A) of the Federal Food, Drug, and Cosmetic Act (the Act), this new device, the Affinity[®] AF100 Arterial Filter with Carmeda[®] BioActive Surface is substantially equivalent to the legally marketed predicate device, Affinity Arterial Filter with Carmeda[®] BioActive Surface (20µm) Model CB353 (K001138).



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

March 19, 2013

Medtronic CardioVascular
c/o Julia A. Nelson, MS, RAC
Principal Regulatory Affairs Specialist
8200 Coral Street NE
Mailstop MVS83
Mounds View, MN 55112

Re: K123351

Trade Name: Affinity® AF100 Arterial Filter with Carmeda® Bioactive Surface
Regulation Number: 21 CFR 870.4260
Regulation Name: Cardiopulmonary Bypass Arterial Line Blood Filter
Regulatory Class: Class II
Product Code: DTM
Dated: February 28, 2013
Received: March 1, 2013

Dear Ms. Nelson:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21

CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Matthew G. Hillebrenner

for Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use Statement

510(k) Number (if known): K123351

Device Name:

Affinity® AF100 with Carmeda® BioActive Surface

Indications for Use:

The AF100 is indicated for use in cardiopulmonary bypass procedures up to 6 hours in duration for the removal of particulate and gaseous microemboli.

Prescription Use X AND/OR Over-The-Counter Use
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Matthew G. Hillebrenner