



Focus Diagnostics
c/o Sharon Young, Senior Regulatory Affairs Specialist
11331 Valley View Street
Cypress, CA 90630

June 9, 2025

Re: DEN130049 (K133621)
Trade/Device Name: Simplexa™ HSV 1 & 2 Direct
Regulation Number: 21 CFR 866.3307
Regulation Name: Herpes simplex virus nucleic acid-based assay for central nervous system infections
Regulatory Class: Class II
Product Code: PGH
Dated: November 19, 2013
Received: December 4, 2013

Dear Sharon Young:

This letter corrects our classification order, originally dated March 21, 2014, to align special controls with existing design controls requirements.

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has completed its review of your De Novo request for classification of the Simplexa™ HSV 1 & 2 Direct, a prescription device with the following indications for use:

The Focus Diagnostics Simplexa™ HSV 1 & 2 Direct is intended for use on the 3M Integrated Cycler instrument for the qualitative detection and differentiation of HSV-1 and HSV-2 DNA in cerebrospinal fluid (CSF) samples from patients suspected of Herpes Simplex Virus (HSV) infections of the central nervous system (CNS). This test is intended as an aid in the diagnosis of HSV-1 and HSV-2 infections of the CNS.

Negative results do not preclude HSV-1 or HSV-2 infection and should not be used as the sole basis for treatment or other patient management decisions.

The assay is not intended for use as a donor screening test. The assay is for professional use only.

The Positive Control is intended to be used as a control with the Simplexa™ HSV 1 & 2 Direct. This control is not intended for use with other assays or systems.

FDA concludes that this device should be classified into Class II. This order, therefore, classifies the Simplexa™ HSV 1 & 2 Direct, and substantially equivalent devices of this generic type, into Class II under the generic name herpes simplex virus nucleic acid-based assay for central nervous system infections.

FDA identifies this generic type of device as:

Herpes simplex virus nucleic acid-based assay for central nervous system infections. A herpes simplex virus nucleic acid-based assay for central nervous system infections is a qualitative in vitro diagnostic device intended for the detection and differentiation of HSV-1 and HSV-2 in cerebrospinal fluid (CSF) samples from patients suspected of Herpes Simplex Virus (HSV) infections of the central nervous system (CNS). This test is intended as an aid in the diagnosis of HSV-1 and HSV-2 infections of the CNS. Negative results do not preclude HSV-1 or HSV-2 infection and should not be used as the sole basis for treatment or other patient management decisions.

Section 513(f)(2) of the Food, Drug and Cosmetic Act (the FD&C Act) was amended by section 607 of the Food and Drug Administration Safety and Innovation Act (FDASIA) on July 9, 2012. This law provides two options for De Novo classification. First, any person who receives a "not substantially equivalent" (NSE) determination in response to a 510(k) for a device that has not been previously classified under the Act may request FDA to make a risk-based classification of the device under section 513(a)(1) of the Act. On December 13, 2016, the 21st Century Cures Act removed a requirement that a De Novo request be submitted within 30 days of receiving an NSE determination. Alternatively, any person who determines that there is no legally marketed device upon which to base a determination of substantial equivalence may request FDA to make a risk-based classification of the device under section 513(a)(1) of the Act without first submitting a 510(k). FDA shall, within 120 days of receiving such a request, classify the device. This classification shall be the initial classification of the device. Within 30 days after the issuance of an order classifying the device, FDA must publish a notice in the Federal Register announcing the classification.

On December 4, 2013, FDA received your De Novo requesting classification of the Simplexa™ HSV 1 & 2 Direct. The request was submitted under section 513(f)(2) of the FD&C Act. In order to classify the Simplexa™ HSV 1 & 2 Direct into class I or II, it is necessary that the proposed class have sufficient regulatory controls to provide reasonable assurance of the safety and effectiveness of the device for its intended use. After review of the information submitted in the De Novo request, FDA has determined that, for the previously stated indications for use, the Simplexa™ HSV 1 & 2 Direct can be classified in class II with the establishment of special controls for class II. FDA believes that class II (special) controls provide reasonable assurance of the safety and effectiveness of the device type. The identified risks and mitigation measures associated with the device type are summarized in the following table:

Identified Risks to Health	Mitigation Measures
Risk of false results	Special controls (1)(i), (1)(ii), and (1)(iii)
Failure to correctly interpret test results	Special control (2)
Failure to correctly operate the instrument	Special controls (1)(iv) and (1)(v)

In combination with the general controls of the FD&C Act, the herpes simplex virus nucleic acid-based assay for central nervous system infections is subject to the following special controls:

- (1) Design verification and validation must include:
 - (i) Detailed documentation for the device description, including the device components, ancillary reagents required but not provided and a detailed explanation of the methodology including primer design and selection.

- (ii) Detailed documentation from the following analytical and clinical performance studies: Analytical sensitivity (Limit of Detection), reactivity, inclusivity, precision, reproducibility, interference, cross reactivity, carry-over, and cross contamination. Documentation must also include reagent and sample stability recommendations.
 - (iii) Detailed documentation from a clinical study. The study, performed on a study population consistent with the intended use population, must compare the device performance to the results of two polymerase chain reaction (PCR) methods followed by bidirectional sequencing.
 - (iv) Documentation of an appropriate end user device training program that will be offered as part of efforts to mitigate the risk of failure to correctly operate the instrument.
 - (v) Quality assurance protocols and a detailed documentation for device software, including standalone software applications and hardware-based devices that incorporate software.
- (2) The labeling required under § 809.10(b) of this chapter must include:
- (i) A detailed explanation of the interpretation of results and acceptance criteria.
 - (ii) A limiting statement that negative results do not preclude HSV-1 or HSV-2 infection and should not be used as the sole basis for treatment or other patient management decisions.

Although this letter refers to your product as a device, please be aware that some granted products may instead be combination products. If you have questions on whether your product is a combination product, contact CDRHProductJurisdiction@fda.hhs.gov.

Section 510(m) of the FD&C Act provides that FDA may exempt a class II device from the premarket notification requirements under section 510(k) of the FD&C Act, if FDA determines that premarket notification is not necessary to provide reasonable assurance of the safety and effectiveness of the device type. FDA has determined premarket notification is necessary to provide reasonable assurance of the safety and effectiveness of the device type and, therefore, the device is not exempt from the premarket notification requirements of the FD&C Act. Thus, persons who intend to market this device type must submit a premarket notification containing information on the herpes simplex virus nucleic acid-based assay for central nervous system infections they intend to market prior to marketing the device.

Please be advised that FDA's decision to grant this De Novo request does not mean that FDA has made a determination that your device complies with other requirements of the FD&C Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the FD&C Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and if applicable, the electronic product radiation control provisions (Sections 531-542 of the FD&C Act; 21 CFR 1000-1050).

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System Rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its

label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

A notice announcing this classification order will be published in the Federal Register. A copy of this order and supporting documentation are on file in the Dockets Management Branch (HFA-305), Food and Drug Administration, 5630 Fishers Lane, Room 1061, Rockville, MD 20852 and are available for inspection between 9 a.m. and 4 p.m., Monday through Friday.

As a result of this order, you may immediately market your device as described in the De Novo request, subject to the general control provisions of the FD&C Act and the special controls identified in this order.

For comprehensive regulatory information about medical devices and radiation-emitting products, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

If you have any questions concerning the contents of the letter, please contact Maria (Ines) Garcia at 301-796-7017.

Sincerely,

Uwe Scherf, M.Sc., Ph.D.
Director
Division of Microbiology Devices
OHT7: Office of In Vitro Diagnostics
Office of Product Evaluation and Quality
Center for Devices and Radiological Health