

K130552  
Page 1 of 2

**510(k) Summary  
CT Pulmonary Analysis  
Qi Imaging, LLC**

JAN 13 2014

This 510(k) summary is prepared in accordance with 21 CFR 807.92.

**General Information**

Trade Name            CT PULMONARY ANALYSIS

Common Name        Picture Archiving and Communications System (PACS)

Classification Name   System, Image Processing, Radiological (21 CFR § 892.2050 - LLZ)

Applicant:            **Qi Imaging, LLC**  
1235 Radio Road, Suite 110  
Redwood City, CA 94065  
Tel 650-413-1300  
Fax 650-596-7319

Contact                Richard Ball  
Director, Regulatory and Quality Affairs

**Intended Use**

Ziostation is an image processing workstation software package designed to run on standard PC hardware. It provides for the viewing, quantification, manipulation, communication, printing, and management of medical images. It is intended for use by trained medical professionals to aid in their reading and review of such data. In addition, Ziostation has the following indications:

The **CT PULMONARY ANALYSIS** software option is an independent image analysis software tool providing additional image process capabilities to the Ziostation system. The **CT PULMONARY ANALYSIS** option is intended to assist the physician in assessing possible airway obstruction.

**Predicate Devices**

Table 2: Predicate devices to CT Pulmonary Analysis Software Tool

Qi Imaging tool	Manufacturer of Predicate Device	Device Name	510(k) Number
CT PULMONARY ANALYSIS	VIDA	VIDA Pulmonary Workstation 2.0	K083227
	Intrasense	Myrian v1.11.2 XP-Lung	K113620

### Device Description

CT PULMONARY ANALYSIS is an optional software package designed to be used with the basic Ziostation DICOM image management system to further aid clinicians in their analysis of anatomy, physiology and pathology. Universal functions such as data retrieval, storage, management, querying and listing, and output are handled by the basic Ziostation software. The added capabilities provided by this additional software option for use with CT DICOM compliant images are:

- Segmentation of lung and bronchial airway
- Contour detection of airway walls
- Evaluation of bronchial dimensions
- Observation of cross-sectional image and variance of diameter/thickness/area of airway wall along airway
- Integrated 3D visualization with MPR or A/C/S slice/slab viewing
- Low Attenuation Analysis by Lung Sections

### Materials

This software tool consists entirely of software. No materials are contained in this product.

### Testing Summary

The CT PULMONARY ANALYSIS software package successfully completed integration testing/verification testing prior to Beta validation. Software Beta testing/validation was successfully completed prior to final testing and release. In addition, potential hazards have been addressed by the Qi Imaging Risk Management process.

### Summary of Substantial Equivalence

CT PULMONARY ANALYSIS is substantially equivalent in intended use and function to the predicate devices identified above.



January 13, 2014

Qi Imaging, LLC  
% Mr. Richard Ball  
Director, Regulatory & Quality Affairs  
1235 Radio Road, Suite 110  
REDWOOD CITY CA 94065

Re: K130552  
Trade/Device Name: CT Pulmonary Analysis  
Regulation Number: 21 CFR 892.1750  
Regulation Name: Computed tomography x-ray system  
Regulatory Class: II  
Product Code: JAK  
Dated: January 7, 2014  
Received: January 8, 2014

Dear Mr. Ball:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



for

Janine M. Morris  
Director, Division of Radiological Health  
Office of In Vitro Diagnostics  
and Radiological Health  
Center for Devices and Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)  
K130552

Device Name  
CT PULMONARY ANALYSIS Software Tool for Ziostation

Indications for Use (Describe)

Ziostation is an image processing workstation software package designed to run on standard PC hardware. It provides for the viewing, quantification, manipulation, communication, printing, and management of medical images. It is intended for use by trained medical professionals to aid in their reading and review of such data. In addition, Ziostation has the following indication:  
The CT PULMONARY ANALYSIS software option is an independent image analysis software tool providing additional image process capabilities to the Ziostation system. The CT PULMONARY ANALYSIS option is intended to assist the physician in assessing possible airway obstruction.

Type of Use (Select one or both, as applicable)


Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE -- CONTINUE ON A SEPARATE PAGE IF NEEDED.**

**FOR FDA USE ONLY**

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)



---

This section applies only to requirements of the Paperwork Reduction Act of 1995.

**\*DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.\***

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services  
Food and Drug Administration  
Office of Chief Information Officer  
Paperwork Reduction Act (PRA) Staff  
[PRASStaff@fda.hhs.gov](mailto:PRASStaff@fda.hhs.gov)

*"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."*