



November 2, 2016

Stryker Corporation
Ms. Estela Celi
Regulatory Affairs Specialist
325 Corporate Drive
Mahwah, New Jersey 07430

Re: K130895

Trade/Device Name: ReUnion RSA Shoulder System

Regulation Number: 21 CFR 888.3660

Regulation Name: Shoulder joint metal/polymer semi-constrained cemented prosthesis

Regulatory Class: Class II

Product Code: PHX, KWS

Dated: November 15, 2013

Received: November 18, 2013

Dear Ms. Celi:

This letter corrects our substantially equivalent letter of December 27, 2013.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins -S

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K130895

Device Name: **ReUnion RSA Shoulder System**

Indications for Use:

The ReUnion RSA Shoulder System is intended for primary, fracture, or revision total shoulder replacement. The patient's joint must have gross rotator cuff deficiency, a functional deltoid muscle and be anatomically and structurally suited to receive the selected implant(s).

- Painful, disabling joint disease of the shoulder resulting from: degenerative arthritis or rheumatoid arthritis
- Proximal humeral fracture
- Revision of previously failed shoulder joint replacement
- Glenoid Baseplate components are intended for cementless use with the addition of screw fixation. The humeral stem components are intended for both cemented and cementless use.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Casey 
Division of  Devices

K130895

510(k) Summary

Proprietary Name: ReUnion RSA Shoulder System

Common Name: Shoulder Prosthesis

Classification Name and Reference: Shoulder joint metal/polymer semi-constrained cemented prosthesis 21 CFR §888.3660

Regulatory Class: Class II

Product Codes: KWS

For Information contact: Estela Celi, Regulatory Affairs Specialist
Stryker Trauma AG
325 Corporate Drive
Mahwah, NJ 07430
Phone: (201) 831-6461 Fax: (201) 831-3461
estela.celi@stryker.com

Alternate Contact: Karen Ariemma
Howmedica Osteonics Corp.
325 Corporate Drive
Mahwah, NJ 07430
Phone: (201) 831-5718 Fax: (201) 831-4718
karen.ariemma@stryker.com

Date Prepared: March 28, 2013

Legally Marketed Device to Which Substantial Equivalence is Claimed:

- K080642 & K072804-Comprehensive Reverse Shoulder (Biomet)
- K112069- Reverse Shoulder System (DJO Encore)
- K081171-Tritanium Acetabular Shell (Howmedica Osteonics)
- K062250-Delta Xtend Reverse Shoulder System (DePuy)
- K021478-Delta Shoulder (DePuy)

Description:

This Traditional 510(k) submission is being supplied to the U.S. FDA to provide authorization to market the ReUnion RSA Shoulder System. The ReUnion RSA Shoulder System is a system of components intended for total shoulder replacement in a reverse shoulder configuration. The system is comprised of a Humeral Cup, Humeral Insert, Glenoid Baseplate, Center Screw, Peripheral Screws and Glenosphere. The Humeral Cup with the Humeral Insert are attached to

K130895

the humeral side of the joint via the ReUnion TSA Humeral Stem while the Glensphere is implanted with the Glenoid Baseplate onto the glenoid side of the joint fixated with locking Center and Peripheral Screws. The ReUnion RSA Shoulder System components are indicated for primary reverse shoulder or revision reverse shoulder replacement procedures having gross rotator cuff deficiency.

Intended Use:

The ReUnion RSA Shoulder System is intended for primary, fracture, or revision total shoulder replacement. The patient's joint must have gross rotator cuff deficiency, a functional deltoid muscle and be anatomically and structurally suited to receive the selected implant(s).

Indications:

The ReUnion RSA Shoulder System is intended for primary, fracture, or revision total shoulder replacement. The patient's joint must have gross rotator cuff deficiency, a functional deltoid muscle and be anatomically and structurally suited to receive the selected implant(s).

- Painful, disabling joint disease of the shoulder resulting from: degenerative arthritis or rheumatoid arthritis
- Proximal humeral fracture
- Revision of previously failed shoulder joint replacement
- Glenoid Baseplate components are intended for cementless use with the addition of screw fixation. The humeral stem components are intended for both cemented and cementless use.

Summary of Technologies:

Device comparison showed that the proposed device is substantially equivalent in intended use, materials and performance characteristics to the predicate devices. The following differences exist between the subject and predicate devices.

- Both the Reverse Shoulder System and the ReUnion RSA Shoulder System subject device contain two taper junctions. The Comprehensive Reverse Shoulder contains an additional taper present in the taper adaptor between the Glensphere and Glenoid Baseplate.
- The Comprehensive Reverse Shoulder contains a post as well as Center Screw to fix the Glenoid Baseplate to the glenoid bone. Both the ReUnion RSA Shoulder System and the Reverse Shoulder System contain only a Center Screw.

K130895

Non-Clinical Testing:

Non-clinical laboratory testing was performed for the ReUnion RSA Shoulder System components to determine substantial equivalence. Testing and evaluations demonstrated that the subject device is substantially equivalent to devices currently cleared for marketing. The testing/analysis included the following:

- Fatigue Strength Testing
- Micromotion Testing
- Engineering, Literature, and MAUDE Analysis of Fretting in the Shoulder Joint
- Glenoid Baseplate Substrate and Coating Characterization
- Static Testing of the humeral Insert/humeral cup junction
- Disassembly testing of the glenosphere/glenoid baseplate connection.
- Evaluation of Range of Motion

Clinical Testing: Clinical testing was not required for this submission.

Conclusion: The ReUnion RSA Shoulder System is substantially equivalent to the predicate devices identified in this premarket notification.