

**510(k) Summary**  
**PLATEAU® Spacer System**

**Submitted By:** Life Spine, Inc.  
2401 W. Hassell Road, Suite 1535  
Hoffman Estates, IL 60169  
Telephone: 847-884-6117  
Fax: 847-884-6118

**JUL 17 2013**

**510(k) Contact:** Randy Lewis  
Life Spine  
2401 W. Hassell Road, Suite 1535  
Hoffman Estates, IL 60169  
Telephone: 847-884-6117  
Fax: 847-884-6118

**Date Prepared:** April 14<sup>th</sup>, 2013

**Trade Name:** Life Spine Plateau Spacer System

**Common Name:** Intervertebral Body Fusion Device

**Classification:** MAX, 21 CFR 888.3080, Class II

**Predicate Device:** Plateau Spacer System (K121116) and The HONOUR Spacer System (K120345)

**Device Description:**

The Plateau Spacer System is intended to serve as an intervertebral body fusion device. The implant is available in a range of sizes and footprints to suit the individual pathology and anatomical conditions of the patient. It is fabricated and manufactured from Polyetheretherketone conforming to ASTM F2026 with tantalum markers. The implant is hollow to permit packing with autogenous bone graft to help promote intervertebral body fusion. The superior and inferior surfaces have teeth to assist in the interface with the vertebral endplates to prevent rotation and/or migration.

All implants are intended for single use only and should not be reused under any circumstances. **Do not use any of the Plateau Spacer System components with components from any other system or manufacturer. The Plateau Spacer System components should never be reused under any circumstances.**

**Intended Use of the Device:**

The Plateau Spacer System is intended for spinal fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels (L2-S1). DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. DDD patients may also have up to Grade I spondylolisthesis at the involved level(s). It is to be used in patients who have had at least six months of non-operative treatment. This device is intended to be used with autogenous bone graft and a supplemental internal spinal fixation system (e.g., pedicle screw or anterolateral plating system) that is cleared for use in the lumbosacral spine.

**Technological Characteristics:**

The Plateau Spacer System is substantially equivalent to the predicate systems in terms of design, materials, indications for use and sizing.

**Performance Data:**

Static and dynamic testing in compression and torsion in accordance with ASTM F2077 was presented to demonstrate the substantial equivalency of the Plateau Spacer System.

**Substantial Equivalence:**

The Plateau Spacer System was shown to be substantially equivalent to previously cleared devices in indications for use, design, function, and materials used.



July 17, 2013

Life Spine, Incorporated  
% Mr. Randy Lewis  
Director, RA/QA  
2401 West Hassell Road, Suite 1535  
Hoffman Estates, Illinois 60169

Re: K131077

Trade/Device Name: PLATEAU<sup>®</sup> Spacer System and PLATEAU<sup>®</sup>-C Spacer System  
Regulation Number: 21 CFR 888.3080  
Regulation Name: Intervertebral body fusion device  
Regulatory Class: Class II  
Product Code: MAX, ODP  
Dated: April 26, 2013  
Received: April 29, 2013

Dear Mr. Lewis:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Mr. Randy Lewis

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Erin I. Keith**

For

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) number (if known): K131077

### Device Name: PLATEAU® Spacer System

The Plateau Spacer System is intended for spinal fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one or two contiguous levels (L2-S1). DDD is defined as back pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies. DDD patients may also have up to Grade 1 spondylolisthesis at the involved level(s). It is to be used in patients who have had at least six months of non-operative treatment. This device is intended to be used with autogenous bone graft and a supplemental internal spinal fixation system (e.g., pedicle screw or anterolateral plating system) that is cleared for use in the lumbosacral spine.

Prescription Use  x   
(Part 21 CFR 801 Subpart D)

And/Or

Over-the-Counter Use        
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER  
PAGE IF NEEDED)

---

Anton E. Dmitriev, PhD  
Division of Orthopedic Devices

### Indications for Use

510(k) number (if known): K131077

Device Name: PLATEAU<sup>®</sup>-C Spacer System

The Plateau-C Spacer System is intended for spinal fusion procedures in skeletally mature patients with degenerative disc disease (DDD) at one disc level (C2-T1). DDD is defined as discogenic pain with degeneration of the disc confirmed by history and radiographic studies. It is to be used in patients who have had at least six weeks of non-operative treatment. This device is intended to be used with autogenous bone graft and a supplemental internal spinal fixation system.

Prescription Use  x  And/Or Over-the-Counter Use    
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Anton E. Dmitriev, PhD  
Division of Orthopedic Devices