

510(k) Summary
S-Scan
Esaote S.p.A.

K131996
DEC 13 2013

510(k) Summary

The following 510(k) summary has been prepared pursuant to requirements specified in 21CFR 807.92(a).

807.92(a)(1)

Submitter Information

Donatella Ragghianti
Esaote, S.p.A.
Via Siffredi 58
Genova 16153 Italy

Contact Person: Allison Scott, RAC
P: 317.569.9500 x106
F: 317.569.9520
ascott@ansongroup.com

Date: June 26, 2013

807.92(a)(2)

Devices

Trade Name: S-scan
Common Name: System, Nuclear Magnetic Resonance Imaging
Classification Name(s): Magnetic Resonance Diagnostic Device
Classification Number: 90LNH

807.92(a)(3)

Predicate Device(s)

Trade Name	Common name	Class	Product code	Manufacturer	K number
S-scan	System, nuclear magnetic resonance imaging	II	LNH	ESAOTE S.P.A.	K121650

807.92(a)(4)

Device Description

The changes performed on the modified S-scan device, with respect to the cleared version – S-scan K121650 – are due to the improvement of the system performance. These modifications, which do not affect the intended use or alter the fundamental scientific technology of the device, are the following:

1. Upgrading of the electronics
2. Patient bed unlocking function
3. A new software version

807.92(a)(5)

Intended Use(s)

S-scan is a Magnetic Resonance (MR) system that produces transversal, sagittal and coronal and oblique cross-section images of the limbs, joints and spinal column. It is intended for imaging portions of the upper limb, including the hand, wrist, forearm, elbow, arm and shoulder, imaging portions of the lower limb, including the foot, ankle, calf, knee, thigh and hip, imaging of the temporomandibular joint and imaging the cervical spine and the lumbar spine sections as portions of the spinal column.

S-scan images correspond to the spatial distribution of protons (hydrogen nuclei) that determine magnetic resonance properties and are dependent on the MR parameters, including spin-lattice relaxation time (T1), spin-spin relaxation time (T2), nuclei density, flow velocity and "chemical shift". When interpreted by a medical expert trained in the use of MR equipment, the images can provide diagnostically useful information.

807.92(a)(6)

Technological Characteristics

The changes to the S-scan system, reflected in this Special 510(k), do not alter the fundamental scientific technology of the S-scan system, the predicate device, cleared via K121650.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Esaote, S.p.A.
% Allison Scott, RAC
Regulatory Consultant
Anson Group, LLC
9001 Wesleyan Road, Suite 200
INDIANAPOLIS IN 46268

December 13, 2013

Re: K131996
Trade/Device Name: S-scan
Regulation Number: 21 CFR 892.1000
Regulation Name: Magnetic resonance diagnostic device
Regulatory Class: II
Product Code: LNH
Dated: November 14, 2013
Received: November 15, 2013

Dear Ms. Scott:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink that reads "Michael D. O'Hara" followed by a small "for" written below the end of the signature.

Janine M. Morris
Director, Division of Radiological Health
Office of In Vitro Diagnostics
and Radiological Health
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K131996

Device Name
S-scan MR System

Indications for Use (Describe)

S-scan is a Magnetic Resonance (MR) system that produces transversal, sagittal and coronal and oblique cross-section images of the limbs, joints and spinal column. It is intended for imaging portions of the upper limb, including the hand, wrist, forearm, elbow, arm and shoulder, imaging portions of the lower limb, including the foot, ankle, calf, knee, thigh and hip, imaging of the temporomandibular joint and imaging the cervical spine and the lumbar spine sections as portions of the spinal column.

S-scan images correspond to the spatial distribution of protons (hydrogen nuclei) that determine magnetic resonance properties and are dependent on the MR parameters, including spin-lattice relaxation time (T1), spin-spin relaxation time (T2), nuclei density, flow velocity and "chemical shift". When interpreted by a medical expert trained in the use of MR equipment, the images can provide diagnostically useful information.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

