



November 25, 2013

Karex Industries Sdn. Bhd.
Mary Goh Yin
Quality Assurance Director
PTD 7906 & 7907, Taman Pontian Jaya, Batu 34, Jalan Johor
Pontian Johor Darul Takzim 82000
Malaysia

Re: K132597
Trade/Device Name: Black Colored Condoms
Regulation Number: 21 CFR§ 884.5300
Regulation Name: Condom
Regulatory Class: II
Product Code: HIS
Dated: September 5, 2013
Received: October 1, 2013

Dear Mary Goh Yin,

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Herbert P. Lerner -S

for

Benjamin R. Fisher, Ph.D.
Director
Division of Reproductive, Gastro-Renal,
and Urological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K132597

Device Name
Black Colored Condoms

Indications for Use (Describe)

Black Colored Condoms [Parallel Sided or Contoured or Flared, Textured or Non-textured, Lubricated or Non-lubricated]

- a) Parallel sided smooth, 53mm Black colored condom
- b) Parallel sided dotted, 53mm Black colored condom
- c) Parallel sided ribbed, 53mm Black colored condom
- d) Parallel sided smooth, 56mm Black colored condom
- e) Contoured studded, 53mm Black colored condom
- f) Contoured ribbed & studded, 53mm Black colored condom
- g) Contoured ribbed, 52mm Black colored condom
- h) Flared smooth, 56mm Black colored condom

The Karex condom is used for contraceptive and for prophylactic purposes (to help prevent pregnancy and the transmission of sexually transmitted diseases).

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE -- CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Herbert P. Lerner, S
2013.11.25 15:03:45 -05'00'