

**510(k) Summary
K132820**

JUN 23 2014

Sponsor: Vilex in Tennessee, Inc.
111 Moffitt St.
McMinnville, TN 37110
931-474-7550

Contact: Abraham Lavi

Date Prepared: June 16, 2014

Trade Name: Ultima HA Coated Half Pins & Wires

Common Name: Bone fixation fastener

Classification: 21 CFR 888.3040 – Smooth or threaded memtallic bone fastener

Product Code: JDW/Orthopedics, Class II

Predicate Devices: Smith & Nephew Jet-X Half Pins (HA Coated) (K023921, K033289)
Orthofix External Fixation Screw (Pin) with Hydroxyapatite Coating (K974186) and Pins used with Mini Rail System (K955848)
Stryker Apex Pins (K061493)
Wright Medical Sidekick Rail Half Pins (K080071)
SBI Mini Rail Half Pins by SBI (K093550)
Treu Kirschner Wires by Treu-Instrumente GMBH (K083912)
Medical Facets Kirschner Wires by Medical Facets (K112727)

Description of Device:

This submission includes medical grade stainless steel half pins and wires coated with Hydroxyapatite (HA). They are intended to be used with external fixation systems.

Indications for Use:

The Vilex X-Fix is intended for external fixation with the following indications:

1. Stabilization of Fractures & Osteotomy
2. Rear & Mid-foot Foot Arthrodesis
3. Adult and Pediatric Leg Lengthening
4. Correction of Bone Deformity in Upper & Lower Extremities

Technological Characteristics:

The technological characteristics for the Ultima HA Coated Half Pins and Wires are the same as the characteristics of the predicate devices. All of the sizes included in the Vilex Ultima HA Coated Half Pin and Wire system are within the range of offerings of the predicate devices and the designs of the Vilex devices are similar to the predicate devices. The materials used to manufacture the Vilex Ultima devices are the same as those used to manufacture the predicate devices.

Substantial Equivalence

The design features of the Ultima HA Coated Half Pins & Wires are substantially equivalent to the design features of other predicate devices previously cleared for market. The methods used to establish equivalence are indications for use, material of construction, sizes, and shapes. The safety and effectiveness of the Ultima HA Coated Half Pins & Wires are adequately supported by

the substantial equivalence information, material information and analysis data provided within this Premarket Notification. Therefore, it is concluded that the Ultima HA Coated Half Pins and Wires are substantially equivalent to the noted predicate devices.

Conclusions

While the Ultima HA Coated Half Pins and Wires are not identical to the predicate devices, any differences that may exist do not significantly affect device safety and effectiveness. In addition, the differences do not add new or increased risks and complications. Therefore, it is concluded that the Ultima HA Coated Half Pins and Wires are substantially equivalent to the predicate devices as outlined previously and should not render the subject device NSE.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

July 23, 2014

Vilex, Incorporated
Dr. Abraham Lavi
President
7214 Lake Forest Glen
Lakewood Ranch, Florida 34202

Re: K132820
Trade/Device Name: Ultima HA Coated Half Pins and Wires
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: JDW
Dated: June 19, 2014
Received: June 23, 2014

Dear Dr. Lavi:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you; however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Lori A. Wiggins

for

Mark N. Melkerson

Director

Division of Orthopedic Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure



Manufacturer of Vilex™ bone implants,
Power equipment & surgical instruments

Phone: (931) 474-7550
Fax: (931) 474-7551
www.vilex.com

111 Moffitt Street
McMinnville, TN 37110 USA
E-mail: info@vilex.com

INDICATIONS FOR USE

510(k) NUMBER: K132820

DEVICE NAME: Ultima HA Coated Half Pins and Wires

INDICATIONS FOR USE:

The Vilex X-Fix is intended for external fixation with the following indications:

1. Stabilization of Fractures & Osteotomy
2. Rear & Mid-foot Foot Arthrodesis
3. Adult and Pediatric Leg Lengthening
4. Correction of Bone Deformity in Upper & Lower Extremities

Prescription Use _____ x _____
(Per 21 CFR 801.109)

Concurrence of CDRH, Office of Device Evaluation (~~DOE~~) (ODE)

Page 1 of 1

Casey L. Hanley, Ph.D.

Division of Orthopedic Devices