



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

August 27, 2014

Steritec Products Manufacturing, Company Incorporated.  
Jonathan Rutigliano  
74 Inverness Drive East  
Englewood, CO 80112 US

Re: K133549  
Trade/Device Name: Steritec biological test pack with instant readout integrator, Model  
No.: BI 111  
Regulation Number: 21 CFR 880.2800  
Regulation Name: Biological Test Pack  
Regulatory Class: II  
Product Code: FRC  
Dated: July 31, 2014  
Received: August 1, 2014

Dear Jonathan Rutigliano:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Erin I. Keith -S**

Erin I. Keith, M.S.  
Director  
Division of Anesthesiology, General Hospital,  
Respiratory, Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known): K133549

Device Name: SteriTec Biological Test Pack with Instant Readout Integrator, ModelNo.: BI 111

### Indications For Use:

SteriTec Biological Test Pack with Instant Readout Integrator is a single-use device designed specifically for testing of 3 or 4 minutes at 132°C (270°F) and 3 minutes at 135°C (275°F) in pre-vacuum steam sterilizers. The test pack consists of a self-contained biological indicator containing *Geobacillus stearothermophilus* inside a small package of porous and nonporous materials, simulating the biological indicator 16 towel test packs defined by AAMI ST79. When used as directed, the Chemical Indicator Record Card gives visible indication that sterilizing conditions were met.

Prescription Use  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use    X  
(21 CFR 801 Subpart C)

**(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER  
PAGE  
IF NEEDED.)**

---

Concurrence of CDRH, Office of Device Evaluation (ODE)